

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032422\
 Data File : BN019143.D
 Acq On : 25 Mar 2022 01:20
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 25 04:30:22 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.494	0.505	-2.2	97	0.00
3	n-Nitrosodimethylamine	0.577	0.450	22.0	82	0.00
4 S	2-Fluorophenol	0.896	0.820	8.5	92	0.00
5 S	Phenol-d6	0.940	0.807	14.1	92	0.00
6	bis(2-Chloroethyl)ether	1.118	1.038	7.2	94	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
8 S	Nitrobenzene-d5	0.289	0.234	19.0	85	-0.01
9	Naphthalene	1.138	1.133	0.4	95	-0.01
10	Hexachlorobutadiene	0.278	0.264	5.0	89	-0.01
11 SURR	2-Methylnaphthalene-d10	0.664	0.613	7.7	88	0.00
12	2-Methylnaphthalene	0.774	0.711	8.1	88	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	84	-0.01
14 S	2,4,6-Tribromophenol	0.193	0.133	31.1#	65	-0.01
15 S	2-Fluorobiphenyl	1.603	1.579	1.5	86	-0.01
16	Acenaphthylene	1.856	1.620	12.7	80	0.00
17	Acenaphthene	1.302	1.236	5.1	83	-0.01
18	Fluorene	1.641	1.549	5.6	81	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	80	-0.01
20	4,6-Dinitro-2-methylphenol	0.049	0.037	24.5	67	0.01
21	4-Bromophenyl-phenylether	0.252	0.241	4.4	80	0.00
22	Hexachlorobenzene	0.325	0.334	-2.8	83	0.00
23	Atrazine	0.188	0.142	24.5	68	-0.01
24	Pentachlorophenol	0.100	0.053	47.0#	49#	0.00
25	Phenanthrene	1.251	1.219	2.6	82	-0.01
26	Anthracene	1.057	0.907	14.2	75	-0.01
27 SURR	Fluoranthene-d10	1.187	1.156	2.6	80	0.00
28	Fluoranthene	1.468	1.410	4.0	79	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	79	0.00
30	Pyrene	1.993	2.025	-1.6	81	0.00
31 S	Terphenyl-d14	0.982	0.892	9.2	74	0.00
32	Benzo(a)anthracene	1.233	1.008	18.2	72	0.00
33	Chrysene	1.662	1.718	-3.4	81	0.00
34	Bis(2-ethylhexyl)phthalate	0.817	0.662	19.0	70	0.00
35 I	Perylene-d12	1.000	1.000	0.0	89	-0.01
36	Indeno(1,2,3-cd)pyrene	1.443	1.553	-7.6	100	-0.02
37	Benzo(b)fluoranthene	1.459	1.354	7.2	90	-0.01
38	Benzo(k)fluoranthene	2.000	1.992	0.4	88	-0.01
39 C	Benzo(a)pyrene	1.328	1.321	0.5	91	-0.01
40	Dibenzo(a,h)anthracene	1.124	1.126	-0.2	93	-0.02
41	Benzo(g,h,i)perylene	1.417	1.609	-13.5	111	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4