

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032422\
 Data File : BN019145.D
 Acq On : 25 Mar 2022 03:43
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 25 04:30:28 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	99	0.00
2	1,4-Dioxane	0.400	0.414	-3.5	97	0.00
3	n-Nitrosodimethylamine	0.400	0.314	21.5	82	0.00
4 S	2-Fluorophenol	0.400	0.374	6.5	93	0.00
5 S	Phenol-d6	0.400	0.351	12.3	93	0.00
6	bis(2-Chloroethyl)ether	0.400	0.372	7.0	93	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	94	0.00
8 S	Nitrobenzene-d5	0.400	0.320	20.0	84	-0.01
9	Naphthalene	0.400	0.394	1.5	95	-0.01
10	Hexachlorobutadiene	0.400	0.378	5.5	90	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.363	9.3	87	0.00
12	2-Methylnaphthalene	0.400	0.369	7.8	89	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	84	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.281	29.8#	66	-0.01
15 S	2-Fluorobiphenyl	0.400	0.393	1.8	86	-0.01
16	Acenaphthylene	0.400	0.352	12.0	81	0.00
17	Acenaphthene	0.400	0.381	4.8	83	-0.01
18	Fluorene	0.400	0.380	5.0	81	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	82	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.299	25.3#	65	0.01
21	4-Bromophenyl-phenylether	0.400	0.381	4.8	81	0.00
22	Hexachlorobenzene	0.400	0.405	-1.3	83	-0.01
23	Atrazine	0.400	0.301	24.8	69	-0.01
24	Pentachlorophenol	0.400	0.199	50.2#	47	0.00
25	Phenanthrene	0.400	0.384	4.0	82	-0.01
26	Anthracene	0.400	0.341	14.8	76	-0.01
27 SURR	Fluoranthene-d10	0.400	0.391	2.3	81	0.00
28	Fluoranthene	0.400	0.392	2.0	82	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	81	0.00
30	Pyrene	0.400	0.397	0.8	82	0.00
31 S	Terphenyl-d14	0.400	0.364	9.0	76	0.00
32	Benzo(a)anthracene	0.400	0.367	8.3	83	0.00
33	Chrysene	0.400	0.398	0.5	80	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.329	17.8	73	0.00
35 I	Perylene-d12	0.400	0.400	0.0	92	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.434	-8.5	105	-0.01
37	Benzo(b)fluoranthene	0.400	0.368	8.0	94	-0.01
38	Benzo(k)fluoranthene	0.400	0.382	4.5	88	0.00
39 C	Benzo(a)pyrene	0.400	0.401	-0.3	95	0.00
40	Dibenzo(a,h)anthracene	0.400	0.424	-6.0	102	-0.02
41	Benzo(g,h,i)perylene	0.400	0.453	-13.2	116	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4