

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032521\
 Data File : BN014106.D
 Acq On : 25 Mar 2021 15:00
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Mar 25 15:35:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 23 11:06:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.692	152	5519	0.40	ng	0.00
7) Naphthalene-d8	10.467	136	21244	0.40	ng	-0.01
13) Acenaphthene-d10	14.321	164	12658	0.40	ng	0.00
19) Phenanthrene-d10	17.055	188	26854	0.40	ng	-0.01
29) Chrysene-d12	21.229	240	26812	0.40	ng	-0.01
36) Perylene-d12	23.438	264	27674	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.292	112	6366	0.50	ng	0.00
5) Phenol-d6	6.863	99	8129	0.50	ng	0.00
8) Nitrobenzene-d5	8.832	82	5851	0.42	ng	-0.01
11) 2-Methylnaphthalene-d10	12.057	152	14818	0.45	ng	0.00
14) 2,4,6-Tribromophenol	15.805	330	1695	0.41	ng	-0.01
15) 2-Fluorobiphenyl	12.938	172	19794	0.42	ng	-0.01
27) Fluoranthene-d10	19.084	212	33641	0.45	ng	0.00
31) Terphenyl-d14	19.685	244	26896	0.46	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.239	88	3273	0.45	ng	94
3) n-Nitrosodimethylamine	3.577	42	1887	0.39	ng	# 98
6) bis(2-Chloroethyl)ether	7.129	93	6447	0.45	ng	95
9) Naphthalene	10.518	128	23506	0.45	ng	100
10) Hexachlorobutadiene	10.809	225	4190	0.42	ng	# 99
12) 2-Methylnaphthalene	12.133	142	15955	0.45	ng	100
16) Acenaphthylene	14.029	152	19952	0.41	ng	99
17) Acenaphthene	14.384	154	15063	0.42	ng	99
18) Fluorene	15.361	166	18713	0.42	ng	98
20) 4,6-Dinitro-2-methylph...	15.450	198	1179	0.39	ng	# 83
21) 4-Bromophenyl-phenylether	16.252	248	5940	0.43	ng	# 91
22) Hexachlorobenzene	16.374	284	6670	0.42	ng	98
23) Atrazine	16.520	200	3856	0.41	ng	# 90
24) Pentachlorophenol	16.715	266	1575	0.42	ng	97
25) Phenanthrene	17.092	178	32267	0.44	ng	100
26) Anthracene	17.189	178	26155	0.43	ng	99
28) Fluoranthene	19.114	202	36158	0.44	ng	99
30) Pyrene	19.476	202	36728	0.45	ng	100
32) Benzo(a)anthracene	21.218	228	33384	0.44	ng	100
33) Chrysene	21.271	228	38373	0.46	ng	97
34) Bis(2-ethylhexyl)phtha...	21.154	149	9961	0.35	ng	98
35) Indeno(1,2,3-cd)pyrene	25.646	276	45047	0.50	ng	99
37) Benzo(b)fluoranthene	22.777	252	41548	0.41	ng	99
38) Benzo(k)fluoranthene	22.777	252	41548	0.42	ng	100
39) Benzo(a)pyrene	23.342	252	38146	0.44	ng	99
40) Dibenzo(a,h)anthracene	25.656	278	38045	0.41	ng	98
41) Benzo(g,h,i)perylene	26.313	276	39777	0.41	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032521\
Data File : BN014106.D
Acq On : 25 Mar 2021 15:00
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Mar 25 15:35:45 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 23 11:06:27 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.692 min Scan# 610
 Delta R.T. -0.008 min
 Lab File: BN014106.D
 Acq: 25 Mar 2021 15:00




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:	152	Resp:	5519
Ion	Ratio	Lower	Upper
152	100		
150	158.0	121.8	182.8
115	61.6	46.6	69.8



1,4-Dioxane
 Concen: 0.45 ng
 RT: 3.239 min Scan# 57
 Delta R.T. -0.008 min
 Lab File: BN014106.D
 Acq: 25 Mar 2021 15:00

Tgt Ion:	88	Resp:	3273
Ion	Ratio	Lower	Upper
88	100		
58	74.7	64.6	96.8
43	30.4	26.6	40.0









n-Nitrosodimethylamine
Concen: 0.39 ng
RT: 3.577 min Scan# 99
Delta R.T. -0.016 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 1887
Ion Ratio Lower Upper
42 100
74 120.8 95.0 142.6
44 4.2 5.8 8.8#



2-Fluorophenol
Concen: 0.50 ng
RT: 5.292 min Scan# 312
Delta R.T. 0.000 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Tgt Ion: 112 Resp: 6366
Ion Ratio Lower Upper
112 100
64 60.1 44.3 66.5
63 31.7 23.0 34.6









Phenol-d6
 Concen: 0.50 ng
 RT: 6.863 min Scan# 507
 Delta R.T. 0.000 min
 Lab File: BN014106.D
 Acq: 25 Mar 2021 15:00




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion:	99	Resp:	8129
Ion	Ratio	Lower	Upper
99	100		
42	22.3	15.2	22.8
71	33.2	26.1	39.1




bis(2-Chloroethyl)ether
 Concen: 0.45 ng
 RT: 7.129 min Scan# 540
 Delta R.T. 0.000 min
 Lab File: BN014106.D
 Acq: 25 Mar 2021 15:00

Tgt Ion:	93	Resp:	6447
Ion	Ratio	Lower	Upper
93	100		
63	81.5	60.8	91.2
95	33.3	26.0	39.0





Abundance Scan 858 (10.480 min): BN013994.D\data.ms (-857#7-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.467 min Scan# 857
Delta R.T. -0.012 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	21244
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.0	13.6
54	7.6	6.4	9.6
68	5.6	4.5	6.7



Abundance Scan 857 (10.467 min): BN014106.D\data.ms (-842) (-)





Abundance Scan 729 (8.845 min): BN013994.D\data.ms (-726#8)




Nitrobenzene-d5
Concen: 0.42 ng
RT: 8.832 min Scan# 728
Delta R.T. -0.012 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00


Tgt Ion:	82	Resp:	5851
Ion	Ratio	Lower	Upper
82	100		
128	34.6	39.2	58.8#
54	56.9	40.1	60.1



Abundance Scan 728 (8.832 min): BN014106.D\data.ms (-713) (-)






Abundance Scan 862 (10.530 min): BN013994.D\data.ms (-855) (-)




Naphthalene
Concen: 0.45 ng
RT: 10.518 min Scan# 861
Delta R.T. 0.000 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	128	Resp:	23506
Ion	Ratio	Lower	Upper
128	100		
129	11.0	9.0	13.4
127	12.8	10.4	15.6






Abundance Scan 885 (10.822 min): BN013994.D\data.ms (-880) (-)



Hexachlorobutadiene
Concen: 0.42 ng
RT: 10.809 min Scan# 884
Delta R.T. 0.000 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00


Tgt Ion:	225	Resp:	4190
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.6	51.2	76.8






Abundance Scan 1066 (12.071 min): BN013994.D\data.ms (-1035) (-)

2-Methylnaphthalene-d10
Concen: 0.45 ng
RT: 12.057 min Scan# 1063
Delta R.T. -0.008 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:152 Resp: 14818
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.2






Abundance Scan 1082 (12.142 min): BN013994.D\data.ms (-1070) (-)

2-Methylnaphthalene
Concen: 0.45 ng
RT: 12.133 min Scan# 1080
Delta R.T. -0.004 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00




Tgt Ion:142 Resp: 15955
Ion Ratio Lower Upper
142 100
141 88.3 70.5 105.7
115 34.1 27.4 41.0





Abundance Scan 1308 (14.334 min): BN013994.D\data.ms (-1308) (-)

Acenaphthene-d10
Concen: 0.40 ng
RT: 14.321 min Scan# 1307
Delta R.T. 0.000 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:	164	Resp:	12658
Ion Ratio	Lower	Upper	
164	100		
162	98.1	79.6	119.4
160	44.4	36.0	54.0






Abundance Scan 1307 (14.321 min): BN014106.D\data.ms (-1291) (-)




Abundance Scan 1425 (15.818 min): BN013994.D\data.ms (-1419) (-)

2,4,6-Tribromophenol
Concen: 0.41 ng
RT: 15.805 min Scan# 1424
Delta R.T. -0.012 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00


Tgt Ion:	330	Resp:	1695
Ion Ratio	Lower	Upper	
330	100		
332	0.0	0.0	0.0
141	36.6	26.8	40.2



Abundance Scan 1424 (15.805 min): BN014106.D\data.ms (-1417) (-)





Abundance Scan 1199 (12.951 min): BN013994.D\data.ms (-1191) (-)




2-Fluorobiphenyl
Concen: 0.42 ng
RT: 12.938 min Scan# 1198
Delta R.T. -0.012 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	172	Resp:	19794
Ion	Ratio	Lower	Upper
172	100		
171	34.7	28.6	42.8
170	23.4	18.9	28.3



Abundance Scan 1198 (12.938 min): BN014106.D\data.ms (-1191) (-)





Abundance Scan 1285 (14.042 min): BN013994.D\data.ms (-1278) (-)




Acenaphthylene
Concen: 0.41 ng
RT: 14.029 min Scan# 1284
Delta R.T. -0.012 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Tgt Ion:	152	Resp:	19952
Ion	Ratio	Lower	Upper
152	100		
151	19.3	15.7	23.5
153	12.8	10.5	15.7




Abundance Scan 1284 (14.029 min): BN014106.D\data.ms (-1269) (-)






Abundance Scan 1312 (14.384 min): BN013994.D\data.ms (-1304) (-)

Acenaphthene
Concen: 0.42 ng
RT: 14.384 min Scan# 1312
Delta R.T. 0.000 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:	154	Resp:	15063
Ion	Ratio	Lower	Upper
154	100		
153	111.2	88.4	132.6
152	54.0	43.0	64.4







Abundance Scan 1390 (15.374 min): BN013994.D\data.ms (-1304) (-)

Fluorene
Concen: 0.42 ng
RT: 15.361 min Scan# 1389
Delta R.T. -0.012 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00



Tgt Ion:	166	Resp:	18713
Ion	Ratio	Lower	Upper
166	100		
165	99.1	77.8	116.8
167	13.6	10.6	16.0









#119 (-)
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.055 min Scan# 1526
Delta R.T. -0.012 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	188	Resp:	26854
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	10.0	9.0	13.6




#120 (-)
4,6-Dinitro-2-methylphenol
Concen: 0.39 ng
RT: 15.450 min Scan# 1396
Delta R.T. 0.000 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Tgt Ion:	198	Resp:	1179
Ion Ratio	Lower	Upper	
198	100		
51	93.4	61.7	92.5#
105	33.5	33.0	49.6





Abundance Scan 1461 (16.264 min): BN013994.D\data.ms (-1453) (-)




4-Bromophenyl-phenylether
Concen: 0.43 ng
RT: 16.252 min Scan# 1460
Delta R.T. -0.012 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	248	Resp:	5940
Ion	Ratio	Lower	Upper
248	100		
250	91.8	75.5	113.3
141	81.5	53.8	80.8#



Abundance Scan 1460 (16.252 min): BN014106.D\data.ms (-1453) (-)





Abundance Scan 1471 (16.386 min): BN013994.D\data.ms (-1462) (-)





Hexachlorobenzene
Concen: 0.42 ng
RT: 16.374 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Tgt Ion:	284	Resp:	6670
Ion	Ratio	Lower	Upper
284	100		
142	36.2	27.9	41.9
249	31.0	24.2	36.2



Abundance Scan 1470 (16.374 min): BN014106.D\data.ms (-1462) (-)









#23 (-)
Atrazine
Concen: 0.41 ng
RT: 16.520 min Scan# 1482
Delta R.T. -0.012 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	200	Resp:	3856
Ion Ratio	Lower	Upper	
200	100		
173	31.0	18.6	28.0#
215	46.6	41.0	61.6




#24 (-)
Pentachlorophenol
Concen: 0.42 ng
RT: 16.715 min Scan# 1498
Delta R.T. -0.012 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Tgt Ion:	266	Resp:	1575
Ion Ratio	Lower	Upper	
266	100		
264	62.9	51.6	77.4
268	64.6	54.6	82.0





Abundance Scan 1530 (17.104 min): BN013994.D\data.ms (-1525) (-)




Phenanthrene
Concen: 0.44 ng
RT: 17.092 min Scan# 1529
Delta R.T. -0.012 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	178	Resp:	32267
Ion Ratio	Lower	Upper	
178	100		
176	18.7	15.0	22.4
179	15.6	12.3	18.5



Abundance Scan 1529 (17.092 min): BN014106.D\data.ms (-1505) (-)





Abundance Scan 1538 (17.201 min): BN013994.D\data.ms (-1526) (-)




Anthracene
Concen: 0.43 ng
RT: 17.189 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00


Tgt Ion:	178	Resp:	26155
Ion Ratio	Lower	Upper	
178	100		
176	18.3	14.6	21.8
179	15.6	12.2	18.2



Abundance Scan 1537 (17.189 min): BN014106.D\data.ms (-1529) (-)






Abundance Scan 1823 (19.099 min): BN013994.D\data.ms (-1823) (-)




Fluoranthene-d10
Concen: 0.45 ng
RT: 19.084 min Scan# 1820
Delta R.T. -0.005 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	212	Resp:	33641
Ion Ratio	Lower	Upper	
212	100		
106	12.6	9.7	14.5
104	7.1	5.5	8.3






Abundance Scan 1829 (19.129 min): BN013994.D\data.ms (-1829) (-)




Fluoranthene
Concen: 0.44 ng
RT: 19.114 min Scan# 1826
Delta R.T. -0.004 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Tgt Ion:	202	Resp:	36158
Ion Ratio	Lower	Upper	
202	100		
101	11.0	8.1	12.1
203	17.2	13.9	20.9






Abundance Scan 2123 (21.250 min): BN013994.D\data.ms (-2123) (-)




Chrysene-d12
Concen: 0.40 ng
RT: 21.229 min Scan# 2121
Delta R.T. -0.010 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	240	Resp:	26812
Ion	Ratio	Lower	Upper
240	100		
120	14.3	10.2	15.4
236	27.0	21.3	31.9






Abundance Scan 1902 (19.491 min): BN013994.D\data.ms (-1899) (-)




Pyrene
Concen: 0.45 ng
RT: 19.476 min Scan# 1899
Delta R.T. -0.009 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Tgt Ion:	202	Resp:	36728
Ion	Ratio	Lower	Upper
202	100		
200	20.4	16.3	24.5
203	17.5	14.0	21.0





Abundance Scan 1944 (19.700 min): BN013994.D\data.ms (-1935) (-)




Terphenyl-d14
Concen: 0.46 ng
RT: 19.685 min Scan# 1941
Delta R.T. -0.004 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.6	6.1	9.1
122	11.9	9.8	14.8



Abundance Scan 1941 (19.685 min): BN014106.D\data.ms (-1912) (-)





Abundance Scan 2121 (21.229 min): BN013994.D\data.ms (-2102) (-)




Benzo(a)anthracene
Concen: 0.44 ng
RT: 21.218 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00


Tgt Ion	Ratio	Lower	Upper
228	100		
226	25.8	20.5	30.7
229	20.0	16.2	24.4



Abundance Scan 2120 (21.218 min): BN014106.D\data.ms (-2101) (-)






Abundance Scan 2126 (21.282 min): BN013994.D\data.ms (-2126) (-)




Chrysene
Concen: 0.46 ng
RT: 21.271 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	228	Resp:	38373
Ion	Ratio	Lower	Upper
228	100		
226	28.7	24.3	36.5
229	19.9	15.1	22.7






Abundance Scan 2115 (21.165 min): BN013994.D\data.ms (-2115) (-)




Bis(2-ethylhexyl)phthalate
Concen: 0.35 ng
RT: 21.154 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Tgt Ion:	149	Resp:	9961
Ion	Ratio	Lower	Upper
149	100		
167	26.7	22.1	33.1
279	4.2	3.5	5.3




Abundance Scan 3186 (25.673 min): BN013994.D\data.ms (-3186) (-)




Indeno(1,2,3-cd)pyrene
Concen: 0.50 ng
RT: 25.646 min Scan# 3178
Delta R.T. -0.006 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion	Ratio	Lower	Upper
276	100		
138	23.8	19.8	29.6
277	24.7	19.9	29.9




Abundance Scan 3178 (25.646 min): BN014106.D\data.ms (-3151) (-)




Abundance Scan 2539 (23.458 min): BN013994.D\data.ms (-2539) (-)




Perylene-d12
Concen: 0.40 ng
RT: 23.438 min Scan# 2533
Delta R.T. -0.010 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00



Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.4	19.3	28.9
265	32.5	66.2	99.4#




Abundance Scan 2533 (23.438 min): BN014106.D\data.ms (-2492) (-)






Abundance Scan 2345 (22.794 min): BN013994.D\data.ms (-2337) (-)

Benzo(b)fluoranthene
Concen: 0.41 ng
RT: 22.777 min Scan# 2340
Delta R.T. -0.006 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:	252	Resp:	41548
Ion Ratio	Lower	Upper	
252	100		
253	25.7	20.4	30.6
125	15.3	11.6	17.4



Abundance Scan 2358 (22.839 min): BN013994.D\data.ms (-2338) (-)

Benzo(k)fluoranthene
Concen: 0.42 ng
RT: 22.777 min Scan# 2340
Delta R.T. -0.051 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00


Tgt Ion:	252	Resp:	41548
Ion Ratio	Lower	Upper	
252	100		
253	25.7	20.5	30.7
125	15.3	12.3	18.5





Abundance Scan 2511 (23.363 min): BN013994.D\data.ms (-2449) (-)

Benzo(a)pyrene
Concen: 0.44 ng
RT: 23.342 min Scan# 2505
Delta R.T. -0.010 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:	252	Resp:	38146
Ion Ratio	Lower	Upper	
252	100		
253	26.9	21.8	32.6
125	17.2	14.6	21.8




Abundance Scan 2505 (23.342 min): BN014106.D\data.ms (-2449) (-)





Abundance Scan 3190 (25.687 min): BN013994.D\data.ms (-3174) (-)


Dibenzo(a,h)anthracene
Concen: 0.41 ng
RT: 25.656 min Scan# 3181
Delta R.T. -0.013 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00



Tgt Ion:	278	Resp:	38045
Ion Ratio	Lower	Upper	
278	100		
139	16.3	13.1	19.7
279	25.4	21.4	32.0



Abundance Scan 3181 (25.656 min): BN014106.D\data.ms (-3161) (-)




Abundance Scan 3384 (26.351 min): BN013994.D\data.ms (-3359) (-)




Benzo(g,h,i)perylene
Concen: 0.41 ng
RT: 26.313 min Scan# 3373
Delta R.T. -0.017 min
Lab File: BN014106.D
Acq: 25 Mar 2021 15:00

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Abundance Scan 3373 (26.313 min): BN014106.D\data.ms



Tgt Ion:	276	Resp:	39777
Ion	Ratio	Lower	Upper
276	100		
277	24.0	20.0	30.0
138	20.9	17.8	26.8



Abundance Scan 3373 (26.313 min): BN014106.D\data.ms (-3319) (-)

