

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN032521\  
 Data File : BN014110.D  
 Acq On : 25 Mar 2021 17:25  
 Operator : CG/JU  
 Sample : M1788-05  
 Misc :  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 RE105D1-20210322

Quant Time: Mar 25 18:10:45 2021  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN032021.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Mar 23 11:06:27 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc | Units | Dev(Min) |
|-------------------------------|--------|------|----------|------|-------|----------|
| -----                         |        |      |          |      |       |          |
| Internal Standards            |        |      |          |      |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.692  | 152  | 4941     | 0.40 | ng    | 0.00     |
| 7) Naphthalene-d8             | 10.467 | 136  | 19101    | 0.40 | ng    | -0.01    |
| 13) Acenaphthene-d10          | 14.321 | 164  | 11100    | 0.40 | ng    | 0.00     |
| 19) Phenanthrene-d10          | 17.055 | 188  | 23478    | 0.40 | ng    | -0.01    |
| 29) Chrysene-d12              | 21.229 | 240  | 23339    | 0.40 | ng    | -0.01    |
| 36) Perylene-d12              | 23.438 | 264  | 23803    | 0.40 | ng    | # 0.00   |
| System Monitoring Compounds   |        |      |          |      |       |          |
| 4) 2-Fluorophenol             | 5.292  | 112  | 1876     | 0.17 | ng    | 0.00     |
| 5) Phenol-d6                  | 6.863  | 99   | 1361     | 0.09 | ng    | 0.00     |
| 8) Nitrobenzene-d5            | 8.832  | 82   | 4095     | 0.32 | ng    | -0.01    |
| 11) 2-Methylnaphthalene-d10   | 12.057 | 152  | 9847     | 0.33 | ng    | 0.00     |
| 14) 2,4,6-Tribromophenol      | 15.805 | 330  | 1402     | 0.39 | ng    | -0.01    |
| 15) 2-Fluorobiphenyl          | 12.938 | 172  | 14178    | 0.34 | ng    | -0.01    |
| 27) Fluoranthene-d10          | 19.084 | 212  | 28815    | 0.44 | ng    | 0.00     |
| 31) Terphenyl-d14             | 19.685 | 244  | 32179    | 0.63 | ng    | 0.00     |
| Target Compounds              |        |      |          |      |       |          |
| 2) 1,4-Dioxane                | 3.231  | 88   | 19596    | 3.01 | ng    | # 96     |
| 6) bis(2-Chloroethyl)ether    | 7.120  | 93   | 335      | 0.03 | ng    | 96       |
| 34) Bis(2-ethylhexyl)phtha... | 21.154 | 149  | 1120     | 0.05 | ng    | # 92     |
| -----                         |        |      |          |      |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN032521\  
Data File : BN014110.D  
Acq On : 25 Mar 2021 17:25  
Operator : CG/JU  
Sample : M1788-05  
Misc :  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
RE105D1-20210322

Quant Time: Mar 25 18:10:45 2021  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN032021.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Mar 23 11:06:27 2021  
Response via : Initial Calibration

