

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032521\
 Data File : BN014108.D
 Acq On : 25 Mar 2021 16:12
 Operator : CG/JU
 Sample : M1787-04MS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT180D-20210318MS

Manual Integrations
 APPROVED

mohammad
 3/26/2021 1:30:21 PM

Quant Time: Mar 25 16:44:42 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 23 11:06:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.692	152	5795	0.40	ng	0.00
7) Naphthalene-d8	10.467	136	22548	0.40	ng	-0.01
13) Acenaphthene-d10	14.321	164	12990	0.40	ng	0.00
19) Phenanthrene-d10	17.055	188	28023	0.40	ng	-0.01
29) Chrysene-d12	21.229	240	28175	0.40	ng	-0.01
36) Perylene-d12	23.438	264	26994	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.292	112	2058	0.15	ng	0.00
5) Phenol-d6	6.863	99	1824	0.11	ng	0.00
8) Nitrobenzene-d5	8.832	82	4628	0.31	ng	-0.01
11) 2-Methylnaphthalene-d10	12.057	152	12043	0.34	ng	0.00
14) 2,4,6-Tribromophenol	15.805	330	1476	0.35	ng	-0.01
15) 2-Fluorobiphenyl	12.938	172	15393	0.32	ng	-0.01
27) Fluoranthene-d10	19.084	212	29733	0.38	ng	0.00
31) Terphenyl-d14	19.685	244	25357	0.41	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.239	88	4690	0.61	ng	# 75
3) n-Nitrosodimethylamine	3.593	42	601	0.12	ng	# 71
6) bis(2-Chloroethyl)ether	7.128	93	5406	0.36	ng	95
9) Naphthalene	10.518	128	18694	0.33	ng	100
10) Hexachlorobutadiene	10.809	225	2547	0.24	ng	# 99
12) 2-Methylnaphthalene	12.133	142	12545	0.33	ng	99
16) Acenaphthylene	14.029	152	17098	0.34	ng	100
17) Acenaphthene	14.384	154	12979	0.35	ng	98
18) Fluorene	15.361	166	16566	0.36	ng	100
20) 4,6-Dinitro-2-methylph...	15.437	198	1005	0.32	ng	# 49
21) 4-Bromophenyl-phenylether	16.252	248	5183	0.36	ng	95
22) Hexachlorobenzene	16.374	284	5393	0.33	ng	98
23) Atrazine	16.520	200	2453	0.25	ng	# 91
24) Pentachlorophenol	16.715	266	2930	0.75	ng	97
25) Phenanthrene	17.092	178	37585	0.49	ng	100
26) Anthracene	17.189	178	26658	0.42	ng	100
28) Fluoranthene	19.114	202	42041	0.49	ng	99
30) Pyrene	19.476	202	41146	0.48	ng	99
32) Benzo(a)anthracene	21.218	228	35760	0.44	ng	99
33) Chrysene	21.271	228	37116	0.43	ng	97
34) Bis(2-ethylhexyl)phtha...	21.154	149	16268	0.54	ng	98
35) Indeno(1,2,3-cd)pyrene	25.642	276	39729	0.42	ng	99
37) Benzo(b)fluoranthene	22.774	252	38413	0.39	ng	98
38) Benzo(k)fluoranthene	22.818	252	35127m	0.36	ng	
39) Benzo(a)pyrene	23.339	252	26234	0.31	ng	95
40) Dibenzo(a,h)anthracene	25.656	278	32676	0.36	ng	99
41) Benzo(g,h,i)perylene	26.316	276	34037	0.36	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032521\
Data File : BN014108.D
Acq On : 25 Mar 2021 16:12
Operator : CG/JU
Sample : M1787-04MS
Misc :
ALS Vial : 4 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS

Manual Integrations
APPROVED


mohammad
3/26/2021 1:30:21 PM

Quant Time: Mar 25 16:44:42 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 23 11:06:27 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.692 min Scan# 610
 Delta R.T. -0.008 min
 Lab File: BN014108.D
 Acq: 25 Mar 2021 16:12





Tgt Ion: 152 Resp: 5795
 Ion Ratio Lower Upper
 152 100
 150 158.2 121.8 182.8
 115 61.6 46.6 69.8



#2
 1,4-Dioxane
 Concen: 0.61 ng
 RT: 3.239 min Scan# 57
 Delta R.T. -0.008 min
 Lab File: BN014108.D
 Acq: 25 Mar 2021 16:12




Tgt Ion: 88 Resp: 4690
 Ion Ratio Lower Upper
 88 100
 58 89.4 64.6 96.8
 43 68.6 26.6 40.0#



Instrument :
 BNA_N
 ClientSampleId :
 TT180D-20210318MS

Manual Integrations
 APPROVED

mohammad
 3/26/2021 1:30:21 PM







#3
n-Nitrosodimethylamine
Concen: 0.12 ng
RT: 3.593 min Scan# 101
Delta R.T. 0.000 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12

Tgt Ion: 42 Resp: 601
Ion Ratio Lower Upper
42 100
74 150.7 95.0 142.6#
44 15.8 5.8 8.8#

Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS




Manual Integrations
APPROVED


mohammad
3/26/2021 1:30:21 PM



#4
2-Fluorophenol
Concen: 0.15 ng
RT: 5.292 min Scan# 312
Delta R.T. 0.000 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12

Tgt Ion: 112 Resp: 2058
Ion Ratio Lower Upper
112 100
64 61.3 44.3 66.5
63 31.8 23.0 34.6









#5
Phenol-d6
Concen: 0.11 ng
RT: 6.863 min Scan# 507
Delta R.T. 0.000 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12

Tgt Ion: 99 Resp: 1824
Ion Ratio Lower Upper
99 100
42 40.6 15.2 22.8#
71 35.9 26.1 39.1

Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS




Manual Integrations
APPROVED


mohammad
3/26/2021 1:30:21 PM



#6
bis(2-Chloroethyl)ether
Concen: 0.36 ng
RT: 7.128 min Scan# 540
Delta R.T. 0.000 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12


Tgt Ion: 93 Resp: 5406
Ion Ratio Lower Upper
93 100
63 81.3 60.8 91.2
95 33.1 26.0 39.0





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.467 min Scan# 857
Delta R.T. -0.012 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12



Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS




Tgt Ion:	136	Resp:	22548
Ion Ratio	Lower	Upper	
136	100		
137	11.2	9.0	13.6
54	8.0	6.4	9.6
68	6.0	4.5	6.7

Manual Integrations
APPROVED



mohammad
3/26/2021 1:30:21 PM




#8
Nitrobenzene-d5
Concen: 0.31 ng
RT: 8.832 min Scan# 728
Delta R.T. -0.012 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12



Tgt Ion:	82	Resp:	4628
Ion Ratio	Lower	Upper	
82	100		
128	35.1	39.2	58.8#
54	59.1	40.1	60.1









#9
Naphthalene
Concen: 0.33 ng
RT: 10.518 min Scan# 861
Delta R.T. 0.000 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12

Tgt Ion:	128	Resp:	18694
Ion Ratio	Lower	Upper	
128	100		
129	11.2	9.0	13.4
127	13.0	10.4	15.6

Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS




Manual Integrations
APPROVED


mohammad
3/26/2021 1:30:21 PM



#10
Hexachlorobutadiene
Concen: 0.24 ng
RT: 10.809 min Scan# 884
Delta R.T. 0.000 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12

Tgt Ion:	225	Resp:	2547
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.1	51.2	76.8









#11
2-Methylnaphthalene-d10
Concen: 0.34 ng
RT: 12.057 min Scan# 1063
Delta R.T. -0.008 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12

Tgt Ion:152 Resp: 12043
Ion Ratio Lower Upper
152 100
151 18.5 16.2 24.2

Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS




Manual Integrations
APPROVED


mohammad
3/26/2021 1:30:21 PM




#12
2-Methylnaphthalene
Concen: 0.33 ng
RT: 12.133 min Scan# 1080
Delta R.T. -0.004 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12

Tgt Ion:142 Resp: 12545
Ion Ratio Lower Upper
142 100
141 87.9 70.5 105.7
115 35.2 27.4 41.0









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.321 min Scan# 1307
Delta R.T. 0.000 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12





Tgt Ion:164 Resp: 12990
Ion Ratio Lower Upper
164 100
162 99.0 79.6 119.4
160 43.9 36.0 54.0



#14
2,4,6-Tribromophenol
Concen: 0.35 ng
RT: 15.805 min Scan# 1424
Delta R.T. -0.012 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12




Tgt Ion:330 Resp: 1476
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 38.8 26.8 40.2



Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS

Manual Integrations
APPROVED

mohammad
3/26/2021 1:30:21 PM







#15
2-Fluorobiphenyl
Concen: 0.32 ng
RT: 12.938 min Scan# 1198
Delta R.T. -0.012 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12

Tgt Ion:	172	Resp:	15393
Ion Ratio	Lower	Upper	
172	100		
171	34.8	28.6	42.8
170	23.5	18.9	28.3

Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS




Manual Integrations
APPROVED


mohammad
3/26/2021 1:30:21 PM



#16
Acenaphthylene
Concen: 0.34 ng
RT: 14.029 min Scan# 1284
Delta R.T. -0.012 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12


Tgt Ion:	152	Resp:	17098
Ion Ratio	Lower	Upper	
152	100		
151	19.4	15.7	23.5
153	12.9	10.5	15.7





#17
Acenaphthene
Concen: 0.35 ng
RT: 14.384 min Scan# 1312
Delta R.T. 0.000 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12




Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS



Tgt Ion	Ratio	Lower	Upper
154	100		
153	108.7	88.4	132.6
152	52.4	43.0	64.4




Manual Integrations
APPROVED


mohammad
3/26/2021 1:30:21 PM



#18
Fluorene
Concen: 0.36 ng
RT: 15.361 min Scan# 1389
Delta R.T. -0.012 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12


Tgt Ion	Ratio	Lower	Upper
166	100		
165	97.6	77.8	116.8
167	13.7	10.6	16.0





#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.055 min Scan# 1526
Delta R.T. -0.012 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12




Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS



Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	9.7	9.0	13.6




Manual Integrations
APPROVED


mohammad
3/26/2021 1:30:21 PM



#20
4,6-Dinitro-2-methylphenol
Concen: 0.32 ng
RT: 15.437 min Scan# 1395
Delta R.T. -0.012 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12


Tgt Ion	Ratio	Lower	Upper
198	100		
51	123.1	61.7	92.5#
105	69.9	33.0	49.6#





#21
4-Bromophenyl-phenylether
Concen: 0.36 ng
RT: 16.252 min Scan# 1460
Delta R.T. -0.012 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12




Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS




Tgt Ion:248 Resp: 5183
Ion Ratio Lower Upper
248 100
250 90.4 75.5 113.3
141 72.6 53.8 80.8

Manual Integrations
APPROVED



mohammad
3/26/2021 1:30:21 PM




#22
Hexachlorobenzene
Concen: 0.33 ng
RT: 16.374 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12




Tgt Ion:284 Resp: 5393
Ion Ratio Lower Upper
284 100
142 36.0 27.9 41.9
249 31.0 24.2 36.2









#23
Atrazine
Concen: 0.25 ng
RT: 16.520 min Scan# 1482
Delta R.T. -0.012 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12





Tgt Ion:200 Resp: 2453
Ion Ratio Lower Upper
200 100
173 33.1 18.6 28.0#
215 49.0 41.0 61.6



#24
Pentachlorophenol
Concen: 0.75 ng
RT: 16.715 min Scan# 1498
Delta R.T. -0.012 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12




Tgt Ion:266 Resp: 2930
Ion Ratio Lower Upper
266 100
264 62.5 51.6 77.4
268 65.3 54.6 82.0




Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS

Manual Integrations
APPROVED




mohammad
3/26/2021 1:30:21 PM




#25
Phenanthrene
Concen: 0.49 ng
RT: 17.092 min Scan# 1529
Delta R.T. -0.012 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12





Tgt Ion:178 Resp: 37585
Ion Ratio Lower Upper
178 100
176 18.7 15.0 22.4
179 15.6 12.3 18.5



#26
Anthracene
Concen: 0.42 ng
RT: 17.189 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12




Tgt Ion:178 Resp: 26658
Ion Ratio Lower Upper
178 100
176 18.1 14.6 21.8
179 15.1 12.2 18.2



Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS

Manual Integrations
APPROVED

mohammad
3/26/2021 1:30:21 PM







#27
Fluoranthene-d10
Concen: 0.38 ng
RT: 19.084 min Scan# 1820
Delta R.T. -0.005 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12

Tgt Ion:212 Resp: 29733
Ion Ratio Lower Upper
212 100
106 12.7 9.7 14.5
104 7.4 5.5 8.3

Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS




Manual Integrations
APPROVED


mohammad
3/26/2021 1:30:21 PM



#28
Fluoranthene
Concen: 0.49 ng
RT: 19.114 min Scan# 1826
Delta R.T. -0.005 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12

Tgt Ion:202 Resp: 42041
Ion Ratio Lower Upper
202 100
101 10.9 8.1 12.1
203 17.1 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.229 min Scan# 2121
Delta R.T. -0.010 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12

Tgt Ion:	240	Resp:	28175
Ion Ratio	Lower	Upper	
240	100		
120	12.3	10.2	15.4
236	26.3	21.3	31.9

Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS




Manual Integrations
APPROVED


mohammad
3/26/2021 1:30:21 PM




#30
Pyrene
Concen: 0.48 ng
RT: 19.476 min Scan# 1899
Delta R.T. -0.010 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12

Tgt Ion:	202	Resp:	41146
Ion Ratio	Lower	Upper	
202	100		
200	20.4	16.3	24.5
203	18.5	14.0	21.0









#31
Terphenyl-d14
Concen: 0.41 ng
RT: 19.685 min Scan# 1941
Delta R.T. -0.005 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12





Tgt Ion:244 Resp: 25357
Ion Ratio Lower Upper
244 100
212 7.6 6.1 9.1
122 12.0 9.8 14.8



#32
Benzo(a)anthracene
Concen: 0.44 ng
RT: 21.218 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12




Tgt Ion:228 Resp: 35760
Ion Ratio Lower Upper
228 100
226 26.4 20.5 30.7
229 19.9 16.2 24.4



Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS

Manual Integrations
APPROVED

mohammad
3/26/2021 1:30:21 PM







#33
Chrysene
Concen: 0.43 ng
RT: 21.271 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12

Tgt Ion:	228	Resp:	37116
Ion Ratio	Lower	Upper	
228	100		
226	28.4	24.3	36.5
229	20.2	15.1	22.7

Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS




Manual Integrations
APPROVED


mohammad
3/26/2021 1:30:21 PM




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.54 ng
RT: 21.154 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12

Tgt Ion:	149	Resp:	16268
Ion Ratio	Lower	Upper	
149	100		
167	26.5	22.1	33.1
279	4.1	3.5	5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.42 ng
RT: 25.642 min Scan# 3177
Delta R.T. -0.010 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12





Tgt Ion: 276 Resp: 39729
Ion Ratio Lower Upper
276 100
138 24.0 19.8 29.6
277 24.5 19.9 29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.438 min Scan# 2533
Delta R.T. -0.010 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12




Tgt Ion: 264 Resp: 26994
Ion Ratio Lower Upper
264 100
260 24.5 19.3 28.9
265 32.3 66.2 99.4#



Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS

Manual Integrations
APPROVED

mohammad
3/26/2021 1:30:21 PM







#37
Benzo(b)fluoranthene
Concen: 0.39 ng
RT: 22.774 min Scan# 2339
Delta R.T. -0.010 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12

Tgt Ion:252 Resp: 38413
Ion Ratio Lower Upper
252 100
253 26.2 20.4 30.6
125 15.8 11.6 17.4

Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS




Manual Integrations
APPROVED


mohammad
3/26/2021 1:30:21 PM




#38
Benzo(k)fluoranthene
Concen: 0.36 ng m
RT: 22.818 min Scan# 2352
Delta R.T. -0.010 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12

Tgt Ion:252 Resp: 35127
Ion Ratio Lower Upper
252 100
253 26.0 20.5 30.7
125 11.6 12.3 18.5#









#39
Benzo(a)pyrene
Concen: 0.31 ng
RT: 23.339 min Scan# 2504
Delta R.T. -0.013 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12





Tgt Ion:252 Resp: 26234
Ion Ratio Lower Upper
252 100
253 29.5 21.8 32.6
125 21.1 14.6 21.8



#40
Dibenzo(a,h)anthracene
Concen: 0.36 ng
RT: 25.656 min Scan# 3181
Delta R.T. -0.013 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12




Tgt Ion:278 Resp: 32676
Ion Ratio Lower Upper
278 100
139 16.5 13.1 19.7
279 26.0 21.4 32.0




Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS

Manual Integrations
APPROVED



mohammad
3/26/2021 1:30:21 PM



#41
Benzo(g,h,i)perylene
Concen: 0.36 ng
RT: 26.316 min Scan# 3374
Delta R.T. -0.013 min
Lab File: BN014108.D
Acq: 25 Mar 2021 16:12



Tgt Ion: 276 Resp: 34037
Ion Ratio Lower Upper
276 100
277 23.9 20.0 30.0
138 21.4 17.8 26.8



Instrument :
BNA_N
ClientSampleId :
TT180D-20210318MS

Manual Integrations
APPROVED

mohammad
3/26/2021 1:30:21 PM