

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032522\
 Data File : BN019169.D
 Acq On : 25 Mar 2022 22:57
 Operator : CG/JU
 Sample : N2090-15
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-59

Quant Time: Mar 26 01:48:57 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 25 19:03:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.847	152	4596	0.40	ng	0.00
7) Naphthalene-d8	10.648	136	13722	0.40	ng	# 0.00
13) Acenaphthene-d10	14.484	164	11086	0.40	ng	0.00
19) Phenanthrene-d10	17.225	188	17481	0.40	ng	0.00
29) Chrysene-d12	21.413	240	14072	0.40	ng	# 0.00
35) Perylene-d12	23.773	264	12157	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.399	112	1622	0.15	ng	0.00
5) Phenol-d6	7.009	99	1005	0.08	ng	0.00
8) Nitrobenzene-d5	8.993	82	3023	0.28	ng	0.00
11) 2-Methylnaphthalene-d10	12.238	152	7223	0.33	ng	0.00
14) 2,4,6-Tribromophenol	15.974	330	1776	0.33	ng	0.00
15) 2-Fluorobiphenyl	13.105	172	10792	0.23	ng	0.00
27) Fluoranthene-d10	19.253	212	20634	0.40	ng	0.00
31) Terphenyl-d14	19.851	244	12660	0.40	ng	0.00
Target Compounds						
9) Naphthalene	10.701	128	2029	0.05	ng	# 1
17) Acenaphthene	14.538	154	7962	0.22	ng	99
28) Fluoranthene	19.286	202	39735	0.63	ng	99
30) Pyrene	19.649	202	77788	1.17	ng	99
34) Bis(2-ethylhexyl)phtha...	21.324	149	4446	0.16	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032522\
Data File : BN019169.D
Acq On : 25 Mar 2022 22:57
Operator : CG/JU
Sample : N2090-15
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-59

Quant Time: Mar 26 01:48:57 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 25 19:03:12 2022
Response via : Initial Calibration

