

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032522\
 Data File : BN019175.D
 Acq On : 26 Mar 2022 03:45
 Operator : CG/JU
 Sample : PB143615BL
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB143615BL

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 03/28/2022
 Supervised By : Jagrut Upadhyay 03/29/2022

Quant Time: Mar 26 05:29:15 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 25 19:03:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.847	152	4265	0.400	ng	0.00
7) Naphthalene-d8	10.648	136	10754	0.400	ng	# 0.00
13) Acenaphthene-d10	14.484	164	6355	0.400	ng	0.00
19) Phenanthrene-d10	17.225	188	12599	0.400	ng	0.00
29) Chrysene-d12	21.422	240	9544	0.400	ng	# 0.00
35) Perylene-d12	23.785	264	8286	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.399	112	3658	0.363	ng	0.00
5) Phenol-d6	7.009	99	3186m	0.276	ng	0.00
8) Nitrobenzene-d5	9.014	82	2140m	0.257	ng	0.02
11) 2-Methylnaphthalene-d10	12.246	152	5804	0.337	ng	0.00
14) 2,4,6-Tribromophenol	15.984	330	777	0.252	ng	0.01
15) 2-Fluorobiphenyl	13.116	172	7789	0.292	ng	0.01
27) Fluoranthene-d10	19.261	212	13576	0.368	ng	0.00
31) Terphenyl-d14	19.855	244	7501	0.352	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032522\
Data File : BN019175.D
Acq On : 26 Mar 2022 03:45
Operator : CG/JU
Sample : PB143615BL
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB143615BL

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 03/28/2022
Supervised By :Jagrut Upadhyay 03/29/2022

Quant Time: Mar 26 05:29:15 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 25 19:03:12 2022
Response via : Initial Calibration

