

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032522\
 Data File : BN019163.D
 Acq On : 25 Mar 2022 18:44
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICBN032522

Quant Time: Mar 25 23:58:29 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 25 19:03:12 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	91	0.00
2	1,4-Dioxane	0.400	0.353	11.8	86	0.00
3	n-Nitrosodimethylamine	0.400	0.346	13.5	85	0.00
4 S	2-Fluorophenol	0.400	0.352	12.0	84	0.00
5 S	Phenol-d6	0.400	0.329	17.8	84	0.00
6	bis(2-Chloroethyl)ether	0.400	0.359	10.3	87	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	91	0.00
8 S	Nitrobenzene-d5	0.400	0.326	18.5	88	0.01
9	Naphthalene	0.400	0.367	8.3	92	0.01
10	Hexachlorobutadiene	0.400	0.381	4.8	94	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.359	10.3	91	0.00
12	2-Methylnaphthalene	0.400	0.363	9.3	91	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	91	0.00
14 S	2,4,6-Tribromophenol	0.400	0.319	20.3	82	0.00
15 S	2-Fluorobiphenyl	0.400	0.364	9.0	91	0.00
16	Acenaphthylene	0.400	0.346	13.5	89	0.01
17	Acenaphthene	0.400	0.363	9.3	89	0.00
18	Fluorene	0.400	0.360	10.0	89	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	89	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.313	21.8	81	0.01
21	4-Bromophenyl-phenylether	0.400	0.364	9.0	89	0.00
22	Hexachlorobenzene	0.400	0.384	4.0	91	0.00
23	Atrazine	0.400	0.334	16.5	83	0.00
24	Pentachlorophenol	0.400	0.288	28.0#	75	0.00
25	Phenanthrene	0.400	0.365	8.8	89	0.00
26	Anthracene	0.400	0.338	15.5	86	0.00
27 SURR	Fluoranthene-d10	0.400	0.363	9.3	89	0.00
28	Fluoranthene	0.400	0.361	9.8	88	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	88	0.00
30	Pyrene	0.400	0.391	2.3	89	0.00
31 S	Terphenyl-d14	0.400	0.377	5.8	89	0.00
32	Benzo(a)anthracene	0.400	0.318	20.5	82	0.00
33	Chrysene	0.400	0.391	2.3	90	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.352	12.0	85	0.00
35 I	Perylene-d12	0.400	0.400	0.0	89	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.369	7.8	93	0.00
37	Benzo(b)fluoranthene	0.400	0.343	14.2	89	0.00
38	Benzo(k)fluoranthene	0.400	0.391	2.3	94	0.00
39 C	Benzo(a)pyrene	0.400	0.387	3.3	98	0.00
40	Dibenzo(a,h)anthracene	0.400	0.343	14.2	93	0.00
41	Benzo(g,h,i)perylene	0.400	0.366	8.5	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
ClientSampleId :
ICVBN032522