

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032522\
 Data File : BN019172.D
 Acq On : 26 Mar 2022 00:45
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 26 01:49:11 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 25 19:03:12 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	1,4-Dioxane	0.533	0.509	4.5	88	0.00
3	n-Nitrosodimethylamine	0.595	0.518	12.9	80	0.00
4 S	2-Fluorophenol	0.945	0.939	0.6	90	0.00
5 S	Phenol-d6	1.082	1.005	7.1	90	0.00
6	bis(2-Chloroethyl)ether	1.168	1.062	9.1	83	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	88	0.00
8 S	Nitrobenzene-d5	0.310	0.261	15.8	88	0.01
9	Naphthalene	1.126	1.020	9.4	88	0.01
10	Hexachlorobutadiene	0.253	0.240	5.1	90	0.00
11 SURR	2-Methylnaphthalene-d10	0.640	0.575	10.2	88	0.00
12	2-Methylnaphthalene	0.740	0.679	8.2	89	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
14 S	2,4,6-Tribromophenol	0.194	0.183	5.7	97	0.00
15 S	2-Fluorobiphenyl	1.681	1.485	11.7	88	0.00
16	Acenaphthylene	1.826	1.652	9.5	93	0.01
17	Acenaphthene	1.294	1.171	9.5	89	0.00
18	Fluorene	1.646	1.479	10.1	89	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
20	4,6-Dinitro-2-methylphenol	0.065	0.052	20.0	82	0.01
21	4-Bromophenyl-phenylether	0.260	0.233	10.4	88	0.00
22	Hexachlorobenzene	0.307	0.291	5.2	90	0.00
23	Atrazine	0.180	0.162	10.0	90	0.00
24	Pentachlorophenol	0.110	0.102	7.3	96	0.00
25	Phenanthrene	1.283	1.160	9.6	89	0.00
26	Anthracene	1.064	0.939	11.7	89	0.00
27 SURR	Fluoranthene-d10	1.170	1.082	7.5	91	0.00
28	Fluoranthene	1.450	1.352	6.8	91	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
30	Pyrene	1.890	1.789	5.3	93	0.00
31 S	Terphenyl-d14	0.893	0.802	10.2	91	0.00
32	Benzo(a)anthracene	1.285	1.204	6.3	104	0.00
33	Chrysene	1.647	1.484	9.9	90	0.00
34	Bis(2-ethylhexyl)phthalate	0.783	0.867	-10.7	115	0.00
35 I	Perylene-d12	1.000	1.000	0.0	103	0.00
36	Indeno(1,2,3-cd)pyrene	1.599	1.426	10.8	104	0.00
37	Benzo(b)fluoranthene	1.519	1.255	17.4	99	0.00
38	Benzo(k)fluoranthene	1.928	1.657	14.1	95	0.00
39 C	Benzo(a)pyrene	1.302	1.189	8.7	106	0.00
40	Dibenzo(a,h)anthracene	1.204	1.039	13.7	108	0.00
41	Benzo(g,h,i)perylene	1.646	1.412	14.2	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4