

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032522\
 Data File : BN019172.D
 Acq On : 26 Mar 2022 00:45
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 26 01:49:11 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 25 19:03:12 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	86	0.00
2	1,4-Dioxane	0.400	0.382	4.5	88	0.00
3	n-Nitrosodimethylamine	0.400	0.348	13.0	80	0.00
4 S	2-Fluorophenol	0.400	0.398	0.5	90	0.00
5 S	Phenol-d6	0.400	0.372	7.0	90	0.00
6	bis(2-Chloroethyl)ether	0.400	0.364	9.0	83	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	88	0.00
8 S	Nitrobenzene-d5	0.400	0.337	15.8	88	0.01
9	Naphthalene	0.400	0.362	9.5	88	0.01
10	Hexachlorobutadiene	0.400	0.380	5.0	90	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.359	10.3	88	0.00
12	2-Methylnaphthalene	0.400	0.367	8.3	89	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	91	0.00
14 S	2,4,6-Tribromophenol	0.400	0.377	5.8	97	0.00
15 S	2-Fluorobiphenyl	0.400	0.353	11.8	88	0.00
16	Acenaphthylene	0.400	0.362	9.5	93	0.01
17	Acenaphthene	0.400	0.362	9.5	89	0.00
18	Fluorene	0.400	0.359	10.3	89	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	89	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.317	20.8	82	0.01
21	4-Bromophenyl-phenylether	0.400	0.358	10.5	88	0.00
22	Hexachlorobenzene	0.400	0.379	5.3	90	0.00
23	Atrazine	0.400	0.359	10.3	90	0.00
24	Pentachlorophenol	0.400	0.370	7.5	96	0.00
25	Phenanthrene	0.400	0.362	9.5	89	0.00
26	Anthracene	0.400	0.353	11.8	89	0.00
27 SURR	Fluoranthene-d10	0.400	0.370	7.5	91	0.00
28	Fluoranthene	0.400	0.373	6.8	91	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	94	0.00
30	Pyrene	0.400	0.379	5.3	93	0.00
31 S	Terphenyl-d14	0.400	0.359	10.3	91	0.00
32	Benzo(a)anthracene	0.400	0.375	6.3	104	0.00
33	Chrysene	0.400	0.360	10.0	90	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.443	-10.7	115	0.00
35 I	Perylene-d12	0.400	0.400	0.0	103	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.357	10.8	104	0.00
37	Benzo(b)fluoranthene	0.400	0.330	17.5	99	0.00
38	Benzo(k)fluoranthene	0.400	0.344	14.0	95	0.00
39 C	Benzo(a)pyrene	0.400	0.365	8.8	106	0.00
40	Dibenzo(a,h)anthracene	0.400	0.345	13.8	108	0.00
41	Benzo(g,h,i)perylene	0.400	0.343	14.2	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4