

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032522\
 Data File : BN019174.D
 Acq On : 26 Mar 2022 03:09
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 26 04:26:27 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 25 19:03:12 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.533	0.480	9.9	92	0.00
3	n-Nitrosodimethylamine	0.595	0.508	14.6	88	0.00
4 S	2-Fluorophenol	0.945	0.895	5.3	95	0.00
5 S	Phenol-d6	1.082	0.948	12.4	94	0.00
6	bis(2-Chloroethyl)ether	1.168	1.024	12.3	90	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
8 S	Nitrobenzene-d5	0.310	0.254	18.1	94	0.01
9	Naphthalene	1.126	1.018	9.6	96	0.00
10	Hexachlorobutadiene	0.253	0.237	6.3	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.640	0.570	10.9	95	0.00
12	2-Methylnaphthalene	0.740	0.662	10.5	95	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.194	0.165	14.9	93	0.00
15 S	2-Fluorobiphenyl	1.681	1.500	10.8	95	0.00
16	Acenaphthylene	1.826	1.599	12.4	97	0.00
17	Acenaphthene	1.294	1.163	10.1	95	0.00
18	Fluorene	1.646	1.492	9.4	96	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
20	4,6-Dinitro-2-methylphenol	0.065	0.042	35.4#	71	0.01
21	4-Bromophenyl-phenylether	0.260	0.232	10.8	95	0.00
22	Hexachlorobenzene	0.307	0.297	3.3	99	0.00
23	Atrazine	0.180	0.148	17.8	89	0.00
24	Pentachlorophenol	0.110	0.092	16.4	93	0.00
25	Phenanthrene	1.283	1.168	9.0	96	0.00
26	Anthracene	1.064	0.916	13.9	94	0.00
27 SURR	Fluoranthene-d10	1.170	1.075	8.1	97	0.00
28	Fluoranthene	1.450	1.328	8.4	96	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
30	Pyrene	1.890	1.830	3.2	97	0.00
31 S	Terphenyl-d14	0.893	0.815	8.7	95	0.00
32	Benzo(a)anthracene	1.285	1.073	16.5	95	0.00
33	Chrysene	1.647	1.587	3.6	98	0.00
34	Bis(2-ethylhexyl)phthalate	0.783	0.702	10.3	95	0.00
35 I	Perylene-d12	1.000	1.000	0.0	100	0.00
36	Indeno(1,2,3-cd)pyrene	1.599	1.380	13.7	98	0.00
37	Benzo(b)fluoranthene	1.519	1.293	14.9	99	0.00
38	Benzo(k)fluoranthene	1.928	1.643	14.8	92	0.00
39 C	Benzo(a)pyrene	1.302	1.222	6.1	106	0.00
40	Dibenzo(a,h)anthracene	1.204	1.043	13.4	106	0.00
41	Benzo(g,h,i)perylene	1.646	1.428	13.2	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4