

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032522\
 Data File : BN019174.D
 Acq On : 26 Mar 2022 03:09
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 26 04:26:27 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 25 19:03:12 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	96	0.00
2	1,4-Dioxane	0.400	0.360	10.0	92	0.00
3	n-Nitrosodimethylamine	0.400	0.342	14.5	88	0.00
4 S	2-Fluorophenol	0.400	0.379	5.3	95	0.00
5 S	Phenol-d6	0.400	0.350	12.5	94	0.00
6	bis(2-Chloroethyl)ether	0.400	0.351	12.3	90	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	97	0.00
8 S	Nitrobenzene-d5	0.400	0.328	18.0	94	0.01
9	Naphthalene	0.400	0.361	9.8	96	0.00
10	Hexachlorobutadiene	0.400	0.375	6.3	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.356	11.0	95	0.00
12	2-Methylnaphthalene	0.400	0.358	10.5	95	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.400	0.339	15.3	93	0.00
15 S	2-Fluorobiphenyl	0.400	0.357	10.8	95	0.00
16	Acenaphthylene	0.400	0.350	12.5	97	0.00
17	Acenaphthene	0.400	0.360	10.0	95	0.00
18	Fluorene	0.400	0.363	9.3	96	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	96	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.256	36.0#	71	0.01
21	4-Bromophenyl-phenylether	0.400	0.358	10.5	95	0.00
22	Hexachlorobenzene	0.400	0.387	3.3	99	0.00
23	Atrazine	0.400	0.329	17.8	89	0.00
24	Pentachlorophenol	0.400	0.335	16.3	93	0.00
25	Phenanthrene	0.400	0.364	9.0	96	0.00
26	Anthracene	0.400	0.344	14.0	94	0.00
27 SURR	Fluoranthene-d10	0.400	0.368	8.0	97	0.00
28	Fluoranthene	0.400	0.366	8.5	96	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	97	0.00
30	Pyrene	0.400	0.387	3.3	97	0.00
31 S	Terphenyl-d14	0.400	0.365	8.8	95	0.00
32	Benzo(a)anthracene	0.400	0.334	16.5	95	0.00
33	Chrysene	0.400	0.385	3.8	98	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.359	10.3	95	0.00
35 I	Perylene-d12	0.400	0.400	0.0	100	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.345	13.8	98	0.00
37	Benzo(b)fluoranthene	0.400	0.340	15.0	99	0.00
38	Benzo(k)fluoranthene	0.400	0.341	14.8	92	0.00
39 C	Benzo(a)pyrene	0.400	0.376	6.0	106	0.00
40	Dibenzo(a,h)anthracene	0.400	0.346	13.5	106	0.00
41	Benzo(g,h,i)perylene	0.400	0.347	13.3	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4