



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032522\
 Data File : BN019155.D
 Acq On : 25 Mar 2022 13:42
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032522.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Mar 25 19:03:12 2022






AutoFind: Scans 2496, 2497, 2498; Background Corrected with Scan 2489

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.0	156492	PASS
68	69	0.00	2	1.8	2789	PASS
69	198	0.00	100	39.9	152322	PASS
70	69	0.00	2	0.5	688	PASS
127	198	10	80	46.4	177003	PASS
197	198	0.00	2	0.3	1251	PASS
198	198	100	100	100.0	381291	PASS
199	198	5	9	6.8	26061	PASS
275	198	10	60	25.0	95264	PASS
365	198	1	100	3.1	11972	PASS
441	198	0.01	100	9.7	36979	PASS
442	442	50	100	100.0	221099	PASS
443	442	15	24	19.3	42757	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032522\
Data File : BN019155.D
Acq On : 25 Mar 2022 13:42
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Mar 28 18:59:12 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 28 18:59:06 2022
Response via : Initial Calibration



TIC: BN019155.D\data.ms

(70) Pentachlorophenol (C)

16.875min (0.000) 20586.23 ng




response 446701

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	63.18
264.00	61.60	64.17
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032522\
Data File : BN019155.D
Acq On : 25 Mar 2022 13:42
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Mar 28 18:59:12 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 28 18:59:06 2022
Response via : Initial Calibration



TIC: BN019155.D\data.ms

(77) Benzidine

19.463min (0.000) 0.00 ng m

response 1039596

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.01
183.00	13.20	11.27
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/28/2022	BNA_N	<u>BN019155.D</u>
Compound Name	Response	Retention Time
DDT	1319045	20.71
DDD	6356	20.26
DDE	211	19.76
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
6567	1325612	0.50

Instrument :
BNA_N
ClientSampleId :
DFTPP