

Data Path : Z:\HPCHEM1\BNA N\DATA\BN032618\
 Data File : BN000624.D
 Acq On : 26 Mar 2018 17:36
 Operator : JU/SJ
 Sample : J1986-08
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Mar 27 04:06:01 2018
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN031918.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 27 03:53:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.41	152	136177	20.00	ng/ul	0.00
18) Naphthalene-d8	11.25	136	589535	20.00	ng/ul	0.00
35) Acenaphthene-d10	15.03	164	332806	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.76	188	734506	20.00	ng/ul	0.00
75) Chrysene-d12	21.88	240	744569	20.00	ng/ul	0.00
83) Perylene-d12	24.36	264	690168	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.58	96	13199	3.46	ng/uL	0.00
5) Phenol-d5	7.55	99	296639	23.35	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.71	67	174063	23.75	ng/ul	0.00
9) 2-Chlorophenol-d4	7.94	132	221699	23.08	ng/ul	0.00
13) 4-Methylphenol-d8	9.12	113	233301	23.58	ng/ul	0.00
19) Nitrobenzene-d5	9.60	128	115497	24.12	ng/ul	0.00
22) 2-Nitrophenol-d4	10.32	143	126626	25.53	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.87	165	209726	23.59	ng/ul	0.00
29) 4-Chloroaniline-d4	11.38	131	63922	7.00	ng/ul	0.00
43) Dimethylphthalate-d6	14.42	166	714157	27.16	ng/ul	0.00
46) Acenaphthylene-d8	14.73	160	883842	25.48	ng/ul	0.00
51) 4-Nitrophenol-d4	15.22	143	128398	26.30	ng/ul	0.00
57) Fluorene-d10	16.01	176	591282	26.39	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	16.13	200	83256	19.79	ng/ul	0.00
70) Anthracene-d10	17.85	188	912615	25.85	ng/ul	0.00
76) Pyrene-d10	20.11	212	1004298	24.43	ng/ul	0.00
87) Benzo(a)pyrene-d12	24.20	264	833594	25.59	ng/ul	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
44) Dimethylphthalate	14.47	163	131089	5.040	ng/ul#	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA N\DATA\BN032618\
Data File : BN000624.D
Acq On : 26 Mar 2018 17:36
Operator : JU/SJ
Sample : J1986-08
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Mar 27 04:06:01 2018
Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN031918.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 27 03:53:32 2018
Response via : Initial Calibration





#44

Dimethylphthalate

Concen: 5.040 ng/ul

RT: 14.47 min Scan# 1875

Delta R.T. -0.00 min


Lab File: BN000624.D

Acq: 26 Mar 2018 17:36

Instrument :

BNA_N

ClientSampleId :



Tot Ion:	163	Resp:	131089
Ion	Ratio	Lower	Upper
163	100		
194	3.8	4.0	6.0#
164	10.0	8.4	12.6

