

Data Path : Z:\HPCHEM1\BNA N\DATA\BN032618\
 Data File : BN000615.D
 Acq On : 26 Mar 2018 12:28
 Operator : JU/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD02034

Quant Time: Mar 27 02:15:42 2018
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN031918.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Mar 26 05:47:13 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	40	0.00
2	1,4-Dioxane	0.603	0.540	10.4	39	0.00
3 S	1,4-Dioxane-d8	0.560	0.502	10.4	38	0.00
4	Benzaldehyde	0.761	0.818	-7.5	40	0.00
5 S	Phenol-d5	1.866	1.899	-1.8	40	0.00
6	Phenol	1.904	1.925	-1.1	40	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.076	1.085	-0.8	40	0.00
8	Bis(2-Chloroethyl)ether	1.511	1.498	0.9	39	0.00
9 S	2-Chlorophenol-d4	1.411	1.396	1.1	40	0.00
10	2-Chlorophenol	1.440	1.440	0.0	40	0.00
11	2-Methylphenol	1.396	1.448	-3.7	40	0.00
12	2,2'-oxybis(1-Chloropropane	1.628	1.669	-2.5	40	0.00
13 S	4-Methylphenol-d8	1.453	1.521	-4.7	40	0.00
14	Acetophenone	2.160	2.354	-9.0	41	0.00
15 P	N-Nitroso-di-n-propylamine	1.130	1.190	-5.3	41	0.00
16	4-Methylphenol	1.495	1.574	-5.3	40	0.00
17	Hexachloroethane	0.601	0.588	2.2	40	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	40	0.00
19 S	Nitrobenzene-d5	0.162	0.156	3.7	40	0.00
20	Nitrobenzene	0.401	0.391	2.5	41	0.00
21	Isophorone	0.764	0.788	-3.1	42	0.00
22 S	2-Nitrophenol-d4	0.168	0.166	1.2	41	0.00
23 C	2-Nitrophenol	0.181	0.178	1.7	41	0.00
24	2,4-Dimethylphenol	0.390	0.392	-0.5	41	0.00
25	Bis(2-Chloroethoxy)methane	0.483	0.473	2.1	41	0.00
26 S	2,4-Dichlorophenol-d3	0.302	0.302	0.0	41	0.00
27 C	2,4-Dichlorophenol	0.293	0.291	0.7	41	0.00
28	Naphthalene	1.050	1.030	1.9	41	0.00
29 S	4-Chloroaniline-d4	0.310	0.401	-29.4#	41	0.00
30	4-Chloroaniline	0.314	0.402	-28.0#	40	0.00
31 C	Hexachlorobutadiene	0.162	0.165	-1.9	43	0.00
32	Caprolactam	0.107	0.126	-17.8	45	0.00
33 C	4-Chloro-3-methylphenol	0.344	0.371	-7.8	43	0.00
34	2-Methylnaphthalene	0.701	0.718	-2.4	42	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	43	0.00
36	1,2,4,5-Tetrachlorobenzene	0.601	0.566	5.8	43	0.00
37	Hexachlorocyclopentadiene	0.313	0.321	-2.6	52	0.00
38 C	2,4,6-Trichlorophenol	0.408	0.389	4.7	43	0.00
39	2,4,5-Trichlorophenol	0.425	0.413	2.8	43	0.00
40	1,1'-Biphenyl	1.703	1.605	5.8	43	0.00
41	2-Chloronaphthalene	1.312	1.240	5.5	43	0.00
42	2-Nitroaniline	0.398	0.406	-2.0	44	0.00
43 S	Dimethylphthalate-d6	1.580	1.628	-3.0	45	0.00
44	Dimethylphthalate	1.563	1.605	-2.7	45	0.00

Data Path : Z:\HPCHEM1\BNA N\DATA\BN032618\
 Data File : BN000615.D
 Acq On : 26 Mar 2018 12:28
 Operator : JU/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD02034

Quant Time: Mar 27 02:15:42 2018
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN031918.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Mar 26 05:47:13 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.315	0.327	-3.8	44	0.00
46 S	Acenaphthylene-d8	2.084	2.047	1.8	44	0.00
47	Acenaphthylene	2.042	1.990	2.5	43	0.00
48	3-Nitroaniline	0.304	0.343	-12.8	44	0.00
49 C	Acenaphthene	1.392	1.360	2.3	43	0.00
50	2,4-Dinitrophenol	0.149	0.138	7.4	44	0.00
51 S	4-Nitrophenol-d4	0.293	0.287	2.0	41	0.00
52	4-Nitrophenol	0.227	0.233	-2.6	43	0.00
53	Dibenzofuran	1.891	1.874	0.9	44	0.00
54	2,4-Dinitrotoluene	0.447	0.474	-6.0	45	0.00
55	2,3,4,6-Tetrachlorophenol	0.325	0.322	0.9	43	0.00
56	Diethylphthalate	1.595	1.701	-6.6	46	0.00
57 S	Fluorene-d10	1.346	1.384	-2.8	45	0.00
58	Fluorene	1.503	1.535	-2.1	45	0.00
59	4-Chlorophenyl-phenylether	0.694	0.711	-2.4	45	0.00
60	4-Nitroaniline	0.342	0.347	-1.5	45	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	45	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.115	0.096	16.5	40	0.00
63	4,6-Dinitro-2-methylphenol	0.121	0.102	15.7	40	0.00
64	N-Nitrosodiphenylamine	0.659	0.616	6.5	45	0.00
65	4-Bromophenyl-phenylether	0.204	0.196	3.9	46	0.00
66	Hexachlorobenzene	0.215	0.216	-0.5	48	0.00
67	Atrazine	0.199	0.221	-11.1	47	0.00
68 C	Pentachlorophenol	0.114	0.111	2.6	47	0.00
69	Phenanthrene	1.152	1.124	2.4	46	0.00
70 S	Anthracene-d10	0.961	0.939	2.3	46	0.00
71	Anthracene	1.181	1.150	2.6	46	0.00
72	Carbazole	1.013	1.054	-4.0	47	0.00
73	Di-n-butylphthalate	1.378	1.443	-4.7	48	0.00
74 C	Fluoranthene	1.145	1.284	-12.1	49	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	52	0.00
76 S	Pyrene-d10	1.104	0.988	10.5	49	0.00
77	Pyrene	1.461	1.276	12.7	49	0.00
78	Butylbenzylphthalate	0.726	0.655	9.8	49	0.00
79	3,3'-Dichlorobenzidine	0.391	0.428	-9.5	55	0.00
80	Benzo(a)anthracene	1.333	1.243	6.8	52	0.00
81	Bis(2-ethylhexyl)phthalate	1.053	0.970	7.9	50	0.00
82	Chrysene	1.249	1.163	6.9	52	0.00
83 I	Perylene-d12	1.000	1.000	0.0	62	0.00
84	Di-n-octyl phthalate	1.683	1.530	9.1	53	0.00
85	Benzo(b)fluoranthene	1.264	1.167	7.7	61	0.00
86	Benzo(k)fluoranthene	1.220	1.118	8.4	58	0.00
87 S	Benzo(a)pyrene-d12	0.944	0.919	2.6	62	0.00

Data Path : Z:\HPCHEM1\BNA N\DATA\BN032618\
Data File : BN000615.D
Acq On : 26 Mar 2018 12:28
Operator : JU/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTD02034

Quant Time: Mar 27 02:15:42 2018
Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN031918.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Mar 26 05:47:13 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.189	1.132	4.8	61	0.00
89	Indeno(1,2,3-cd)pyrene	1.282	1.382	-7.8	73	0.00
90	Dibenzo(a,h)anthracene	1.068	1.151	-7.8	73	0.00
91	Benzo(a,h,i)perylene	1.066	1.166	-9.4	76	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0