

Data Path : Z:\HPCHEM1\BNA N\DATA\BN032618\
 Data File : BN000626.D
 Acq On : 26 Mar 2018 18:45
 Operator : JU/SJ
 Sample : J1986-14
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BEEH3

Quant Time: Mar 27 04:06:30 2018
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN031918.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 27 03:53:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.41	152	124690	20.00	ng/ul	0.00
18) Naphthalene-d8	11.25	136	566359	20.00	ng/ul	0.00
35) Acenaphthene-d10	15.03	164	326970	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.75	188	713200	20.00	ng/ul	0.00
75) Chrysene-d12	21.88	240	731325	20.00	ng/ul	0.00
83) Perylene-d12	24.36	264	776056	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.58	96	14670	4.20	ng/uL	0.00
5) Phenol-d5	7.55	99	290832	25.00	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.71	67	181847	27.10	ng/ul	0.00
9) 2-Chlorophenol-d4	7.94	132	226881	25.79	ng/ul	0.00
13) 4-Methylphenol-d8	9.12	113	216084	23.85	ng/ul	0.00
19) Nitrobenzene-d5	9.59	128	121293	26.37	ng/ul	0.00
22) 2-Nitrophenol-d4	10.32	143	132418	27.79	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.87	165	219029	25.64	ng/ul	0.00
29) 4-Chloroaniline-d4	11.38	131	31127	3.55	ng/ul	0.00
43) Dimethylphthalate-d6	14.42	166	805607	31.19	ng/ul	0.00
46) Acenaphthylene-d8	14.73	160	788074	23.13	ng/ul	0.00
51) 4-Nitrophenol-d4	15.22	143	121706	25.37	ng/ul	0.00
57) Fluorene-d10	16.01	176	660671	30.01	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	16.13	200	88165	21.58	ng/ul	0.00
70) Anthracene-d10	17.85	188	896912	26.16	ng/ul	0.00
76) Pyrene-d10	20.11	212	1045722	25.90	ng/ul	0.00
87) Benzo(a)pyrene-d12	24.20	264	770251	21.03	ng/ul	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
44) Dimethylphthalate	14.47	163	156656	6.131	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA N\DATA\BN032618\
Data File : BN000626.D
Acq On : 26 Mar 2018 18:45
Operator : JU/SJ
Sample : J1986-14
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BEEH3

Quant Time: Mar 27 04:06:30 2018
Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN031918.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 27 03:53:32 2018
Response via : Initial Calibration





#44

Dimethylphthalate

Concen: 6.131 ng/ul

RT: 14.47 min Scan# 1875

Delta R.T. 0.00 min

Lab File: BN000626.D

Acq: 26 Mar 2018 18:45

Instrument :

BNA_N

ClientSampleId :

BEEH3


Tot Ion:163 Resp: 156656

Ion Ratio Lower Upper

163 100

194 4.3 4.0 6.0

164 10.1 8.4 12.6



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E

