

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032621\
 Data File : BN014119.D
 Acq On : 26 Mar 2021 11:51
 Operator : CG/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Quant Time: Mar 26 13:03:36 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 26 13:01:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.692	152	6890	0.40	ng	0.00
7) Naphthalene-d8	10.467	136	25101	0.40	ng	0.00
13) Acenaphthene-d10	14.308	164	12499	0.40	ng	0.00
19) Phenanthrene-d10	17.055	188	23920	0.40	ng	0.00
29) Chrysene-d12	21.229	240	18990	0.40	ng	# 0.00
36) Perylene-d12	23.431	264	17560	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.284	112	3615	0.23	ng	0.00
5) Phenol-d6	6.855	99	4572	0.23	ng	0.00
8) Nitrobenzene-d5	8.832	82	3264	0.20	ng	0.00
11) 2-Methylnaphthalene-d10	12.053	152	7451	0.19	ng	0.00
14) 2,4,6-Tribromophenol	15.805	330	738	0.18	ng	0.00
15) 2-Fluorobiphenyl	12.938	172	9264	0.20	ng	0.00
27) Fluoranthene-d10	19.079	212	11633	0.18	ng	0.00
31) Terphenyl-d14	19.680	244	8659	0.21	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.231	88	1798	0.20	ng	98
3) n-Nitrosodimethylamine	3.561	42	1200	0.20	ng	# 95
6) bis(2-Chloroethyl)ether	7.121	93	3572	0.20	ng	100
9) Naphthalene	10.505	128	11975	0.19	ng	99
10) Hexachlorobutadiene	10.796	225	2236	0.19	ng	# 99
12) 2-Methylnaphthalene	12.124	142	7785	0.19	ng	99
16) Acenaphthylene	14.029	152	8734	0.18	ng	99
17) Acenaphthene	14.372	154	6588	0.19	ng	100
18) Fluorene	15.361	166	8037	0.18	ng	100
20) 4,6-Dinitro-2-methylph...	15.437	198	477	0.18	ng	# 50
21) 4-Bromophenyl-phenylether	16.252	248	2336	0.19	ng	99
22) Hexachlorobenzene	16.362	284	2671	0.19	ng	99
23) Atrazine	16.520	200	1489	0.18	ng	98
24) Pentachlorophenol	16.703	266	727	0.22	ng	95
25) Phenanthrene	17.092	178	12301	0.19	ng	100
26) Anthracene	17.177	178	9829	0.18	ng	100
28) Fluoranthene	19.109	202	12112	0.17	ng	99
30) Pyrene	19.471	202	12443	0.21	ng	100
32) Benzo(a)anthracene	21.208	228	10237	0.19	ng	97
33) Chrysene	21.261	228	11258	0.19	ng	99
34) Bis(2-ethylhexyl)phtha...	21.144	149	3243	0.16	ng	100
35) Indeno(1,2,3-cd)pyrene	25.629	276	12416	0.20	ng	99
37) Benzo(b)fluoranthene	22.815	252	11520	0.18	ng	93
38) Benzo(k)fluoranthene	22.815	252	11520	0.18	ng	93
39) Benzo(a)pyrene	23.335	252	10150	0.19	ng	# 84
40) Dibenzo(a,h)anthracene	25.646	278	10418	0.18	ng	95
41) Benzo(g,h,i)perylene	26.310	276	11230	0.18	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032621\
Data File : BN014119.D
Acq On : 26 Mar 2021 11:51
Operator : CG/JU
Sample : SSTDICC0.2
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Quant Time: Mar 26 13:03:36 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 26 13:01:59 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.692 min Scan# 610
 Delta R.T. 0.000 min
 Lab File: BN014119.D
 Acq: 26 Mar 2021 11:51




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion:152 Resp: 6890
 Ion Ratio Lower Upper
 152 100
 150 152.3 121.2 181.8
 115 58.5 46.2 69.2



1,4-Dioxane
 Concen: 0.20 ng
 RT: 3.231 min Scan# 56
 Delta R.T. 0.000 min
 Lab File: BN014119.D
 Acq: 26 Mar 2021 11:51

Tgt Ion: 88 Resp: 1798
 Ion Ratio Lower Upper
 88 100
 58 78.1 63.1 94.7
 43 35.3 26.2 39.2









n-Nitrosodimethylamine
Concen: 0.20 ng
RT: 3.561 min Scan# 97
Delta R.T. 0.000 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion: 42 Resp: 1200
Ion Ratio Lower Upper
42 100
74 112.9 94.6 141.8
44 9.8 4.2 6.2#



2-Fluorophenol
Concen: 0.23 ng
RT: 5.284 min Scan# 311
Delta R.T. 0.000 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Tgt Ion: 112 Resp: 3615
Ion Ratio Lower Upper
112 100
64 60.1 48.3 72.5
63 31.5 25.4 38.0









Phenol-d6
 Concen: 0.23 ng
 RT: 6.855 min Scan# 506
 Delta R.T. 0.000 min
 Lab File: BN014119.D
 Acq: 26 Mar 2021 11:51




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Tgt Ion:	99	Resp:	4572
Ion	Ratio	Lower	Upper
99	100		
42	21.6	18.2	27.4
71	32.9	26.7	40.1




bis(2-Chloroethyl)ether
 Concen: 0.20 ng
 RT: 7.121 min Scan# 539
 Delta R.T. 0.000 min
 Lab File: BN014119.D
 Acq: 26 Mar 2021 11:51

Tgt Ion:	93	Resp:	3572
Ion	Ratio	Lower	Upper
93	100		
63	81.2	65.1	97.7
95	33.3	26.6	39.8






Abundance Scan 857 (10.467 min): BN014120.D\data.ms (-857) (-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.467 min Scan# 857
Delta R.T. 0.000 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Tgt Ion:	136	Resp:	25101
Ion	Ratio	Lower	Upper
136	100		
137	11.0	9.1	13.7
54	5.1	4.3	6.5
68	4.3	3.4	5.2





Abundance Scan 728 (8.832 min): BN014120.D\data.ms (-724) (-)




Nitrobenzene-d5
Concen: 0.20 ng
RT: 8.832 min Scan# 728
Delta R.T. 0.000 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Tgt Ion:	82	Resp:	3264
Ion	Ratio	Lower	Upper
82	100		
128	42.8	33.5	50.3
54	52.4	40.3	60.5




Abundance Scan 728 (8.832 min): BN014119.D\data.ms (-712) (-)






Abundance Scan 860 (10.505 min): BN014120.D\data.ms (-855) (-)

Naphthalene
Concen: 0.19 ng
RT: 10.505 min Scan# 860
Delta R.T. 0.000 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2







Tgt Ion:	128	Resp:	11975
Ion	Ratio	Lower	Upper
128	100		
129	11.7	9.2	13.8
127	13.8	10.8	16.2




Abundance Scan 883 (10.796 min): BN014120.D\data.ms (-875) (-)

Hexachlorobutadiene
Concen: 0.19 ng
RT: 10.796 min Scan# 883
Delta R.T. 0.000 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Tgt Ion:	225	Resp:	2236
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.0	51.6	77.4






Abundance Scan 1062 (12.053 min): BN014120.D\data.ms (-1019) (-)




2-Methylnaphthalene-d10
Concen: 0.19 ng
RT: 12.053 min Scan# 1062
Delta R.T. 0.000 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Tgt Ion:152 Resp: 7451
Ion Ratio Lower Upper
152 100
151 20.1 16.1 24.1






Abundance Scan 1078 (12.124 min): BN014120.D\data.ms (-1017) (-)



2-Methylnaphthalene
Concen: 0.19 ng
RT: 12.124 min Scan# 1078
Delta R.T. 0.000 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51



Tgt Ion:142 Resp: 7785
Ion Ratio Lower Upper
142 100
141 89.3 71.1 106.7
115 37.2 29.0 43.6





Abundance Scan 1306 (14.308 min): BN014120.D\data.ms (-1302) (-)

Acenaphthene-d10
Concen: 0.40 ng
RT: 14.308 min Scan# 1306
Delta R.T. 0.000 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2





Tgt Ion:	164	Resp:	12499
Ion	Ratio	Lower	Upper
164	100		
162	102.6	81.4	122.2
160	47.3	37.5	56.3





Abundance Scan 1424 (15.805 min): BN014120.D\data.ms (-1416) (-)


2,4,6-Tribromophenol
Concen: 0.18 ng
RT: 15.805 min Scan# 1424
Delta R.T. 0.000 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51



Tgt Ion:	330	Resp:	738
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	37.4	31.2	46.8





Abundance Scan 1198 (12.938 min): BN014120.D\data.ms (-1198) (-)




2-Fluorobiphenyl
Concen: 0.20 ng
RT: 12.938 min Scan# 1198
Delta R.T. 0.000 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	172	Resp:	9264
Ion	Ratio	Lower	Upper
172	100		
171	33.5	26.9	40.3
170	22.3	17.8	26.8



Abundance Scan 1198 (12.938 min): BN014119.D\data.ms (-1190) (-)





Abundance Scan 1284 (14.029 min): BN014120.D\data.ms (-1215) (-)




Acenaphthylene
Concen: 0.18 ng
RT: 14.029 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Tgt Ion:	152	Resp:	8734
Ion	Ratio	Lower	Upper
152	100		
151	19.7	15.4	23.2
153	13.2	10.3	15.5




Abundance Scan 1284 (14.029 min): BN014119.D\data.ms (-1268) (-)






Abundance Scan 1311 (14.371 min): BN014120.D\data.ms (-1305) (-)

Acenaphthene
Concen: 0.19 ng
RT: 14.372 min Scan# 1311
Delta R.T. 0.000 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2







Tgt Ion:	154	Resp:	6588
Ion	Ratio	Lower	Upper
154	100		
153	111.1	89.0	133.4
152	54.5	43.1	64.7




Abundance Scan 1389 (15.361 min): BN014120.D\data.ms (-1314) (-)

Fluorene
Concen: 0.18 ng
RT: 15.361 min Scan# 1389
Delta R.T. 0.000 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Tgt Ion:	166	Resp:	8037
Ion	Ratio	Lower	Upper
166	100		
165	97.5	77.9	116.9
167	13.4	10.6	16.0





Abundance Scan 1526 (17.055 min): BN014120.D\data.ms (-1510) (-)




Phenanthrene-d10
Concen: 0.40 ng
RT: 17.055 min Scan# 1526
Delta R.T. 0.001 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	188	Resp:	23920
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	6.5	5.2	7.8



Abundance Scan 1526 (17.055 min): BN014119.D\data.ms (-1510) (-)





Abundance Scan 1395 (15.437 min): BN014120.D\data.ms (-1320) (-)




4,6-Dinitro-2-methylphenol
Concen: 0.18 ng
RT: 15.437 min Scan# 1395
Delta R.T. 0.000 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51


Tgt Ion:	198	Resp:	477
Ion	Ratio	Lower	Upper
198	100		
51	106.4	47.8	71.6#
105	67.0	35.4	53.2#



Abundance Scan 1395 (15.437 min): BN014119.D\data.ms (-1387) (-)






Abundance Scan 1460 (16.252 min): BN014120.D\data.ms (-1423) (-)




4-Bromophenyl-phenylether
Concen: 0.19 ng
RT: 16.252 min Scan# 1460
Delta R.T. 0.001 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Tgt Ion:	248	Resp:	2336
Ion	Ratio	Lower	Upper
248	100		
250	97.7	77.0	115.6
141	48.9	39.7	59.5






Abundance Scan 1469 (16.362 min): BN014120.D\data.ms (-1423) (-)




Hexachlorobenzene
Concen: 0.19 ng
RT: 16.362 min Scan# 1469
Delta R.T. 0.001 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Tgt Ion:	284	Resp:	2671
Ion	Ratio	Lower	Upper
284	100		
142	36.8	29.2	43.8
249	30.4	24.7	37.1





Abundance Scan 1482 (16.519 min): BN014120.D\data.ms (-1425) (-)




Atrazine
Concen: 0.18 ng
RT: 16.520 min Scan# 1482
Delta R.T. 0.001 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	200	Resp:	1489
Ion	Ratio	Lower	Upper
200	100		
173	26.5	20.9	31.3
215	51.4	39.5	59.3



Abundance Scan 1482 (16.520 min): BN014119.D\data.ms (-1441) (-)





Abundance Scan 1497 (16.702 min): BN014120.D\data.ms (-1424) (-)




Pentachlorophenol
Concen: 0.22 ng
RT: 16.703 min Scan# 1497
Delta R.T. 0.001 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51


Tgt Ion:	266	Resp:	727
Ion	Ratio	Lower	Upper
266	100		
264	64.1	50.2	75.2
268	59.7	52.5	78.7



Abundance Scan 1497 (16.703 min): BN014119.D\data.ms (-1489) (-)





Abundance Scan 1529 (17.091 min): BN014120.D\data.ms (-1525) (-)




Phenanthrene
Concen: 0.19 ng
RT: 17.092 min Scan# 1529
Delta R.T. 0.001 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Tgt Ion:	178	Resp:	12301
Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.0	22.6
179	15.6	12.4	18.6


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2



Abundance Scan 1529 (17.092 min): BN014119.D\data.ms (-1504) (-)





Abundance Scan 1536 (17.177 min): BN014120.D\data.ms (-1525) (-)




Anthracene
Concen: 0.18 ng
RT: 17.177 min Scan# 1536
Delta R.T. 0.001 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51


Tgt Ion:	178	Resp:	9829
Ion	Ratio	Lower	Upper
178	100		
176	18.4	14.7	22.1
179	15.6	12.3	18.5



Abundance Scan 1536 (17.177 min): BN014119.D\data.ms (-1528) (-)





Abundance Scan 1821 (19.081 min): BN014120.D\data.ms (-1821) (-)




Fluoranthene-d10
Concen: 0.18 ng
RT: 19.079 min Scan# 1819
Delta R.T. -0.002 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:212 Resp: 11633
Ion Ratio Lower Upper
212 100
106 12.9 10.1 15.1
104 7.3 5.5 8.3



Abundance Scan 1819 (19.079 min): BN014119.D\data.ms (-1779) (-)





Abundance Scan 1827 (19.111 min): BN014120.D\data.ms (-1827) (-)




Fluoranthene
Concen: 0.17 ng
RT: 19.109 min Scan# 1825
Delta R.T. -0.002 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51


Tgt Ion:202 Resp: 12112
Ion Ratio Lower Upper
202 100
101 11.5 8.8 13.2
203 17.2 13.8 20.6



Abundance Scan 1825 (19.109 min): BN014119.D\data.ms (-1785) (-)






Abundance Scan 2122 (21.226 min): BN014120.D\data.ms (-2122) (-)




Chrysene-d12
Concen: 0.40 ng
RT: 21.229 min Scan# 2121
Delta R.T. 0.003 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Tgt Ion:	240	Resp:	18990
Ion	Ratio	Lower	Upper
240	100		
120	8.0	8.4	12.6#
236	25.8	21.0	31.6






Abundance Scan 1900 (19.473 min): BN014120.D\data.ms (-1930) (-)



Pyrene
Concen: 0.21 ng
RT: 19.471 min Scan# 1898
Delta R.T. -0.002 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51


Tgt Ion:	202	Resp:	12443
Ion	Ratio	Lower	Upper
202	100		
200	20.4	16.2	24.4
203	17.4	13.9	20.9






Abundance Scan 1942 (19.682 min): BN014120.D\data.ms (-1934) (-)

Terphenyl-d14
Concen: 0.21 ng
RT: 19.680 min Scan# 1940
Delta R.T. -0.002 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2




Tgt Ion:	244	Resp:	8659
Ion	Ratio	Lower	Upper
244	100		
212	8.2	6.0	9.0
122	12.3	8.6	13.0






Abundance Scan 2121 (21.215 min): BN014120.D\data.ms (-2125) (-)


Benzo(a)anthracene
Concen: 0.19 ng
RT: 21.208 min Scan# 2119
Delta R.T. -0.008 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51



Tgt Ion:	228	Resp:	10237
Ion	Ratio	Lower	Upper
228	100		
226	27.9	20.6	31.0
229	19.7	16.2	24.2






Abundance Scan 2125 (21.258 min): BN014120.D\data.ms (-2125) (-)




Chrysene
Concen: 0.19 ng
RT: 21.261 min Scan# 2124
Delta R.T. 0.003 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Tgt Ion:	228	Resp:	11258
Ion	Ratio	Lower	Upper
228	100		
226	30.2	24.4	36.6
229	20.0	15.5	23.3






Abundance Scan 2115 (21.152 min): BN014120.D\data.ms (-2104) (-)



Bis(2-ethylhexyl)phthalate
Concen: 0.16 ng
RT: 21.144 min Scan# 2113
Delta R.T. -0.008 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51


Tgt Ion:	149	Resp:	3243
Ion	Ratio	Lower	Upper
149	100		
167	26.9	21.4	32.2
279	4.4	3.5	5.3





Abundance Scan 3176 (25.633 min): BN014120.D\data.ms (-3135) (-)

Indeno(1,2,3-cd)pyrene
Concen: 0.20 ng
RT: 25.629 min Scan# 3173
Delta R.T. -0.004 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2



Tgt Ion:	276	Resp:	12416
Ion	Ratio	Lower	Upper
276	100		
138	24.0	19.4	29.2
277	25.0	19.8	29.6




Abundance Scan 3173 (25.629 min): BN014119.D\data.ms (-3145) (-)





Abundance Scan 2533 (23.432 min): BN014120.D\data.ms (-2526) (-)


Perylene-d12
Concen: 0.40 ng
RT: 23.431 min Scan# 2531
Delta R.T. -0.001 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51




Tgt Ion:	264	Resp:	17560
Ion	Ratio	Lower	Upper
264	100		
260	24.2	19.8	29.8
265	37.1	30.7	46.1



Abundance Scan 2531 (23.431 min): BN014119.D\data.ms (-2487) (-)





Abundance Scan 2340 (22.771 min): BN014120.D\data.ms (-2340) (-)




Benzo(b)fluoranthene
Concen: 0.18 ng
RT: 22.815 min Scan# 2351
Delta R.T. 0.044 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Tgt Ion:252 Resp: 11520
Ion Ratio Lower Upper
252 100
253 30.0 20.6 30.8
125 13.5 9.4 14.2


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2



Abundance Scan 2351 (22.815 min): BN014119.D\data.ms (-2280) (-)





Abundance Scan 2352 (22.812 min): BN014120.D\data.ms (-2346) (-)




Benzo(k)fluoranthene
Concen: 0.18 ng
RT: 22.815 min Scan# 2351
Delta R.T. 0.003 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Tgt Ion:252 Resp: 11520
Ion Ratio Lower Upper
252 100
253 30.0 20.6 30.8
125 13.5 9.5 14.3




Abundance Scan 2351 (22.815 min): BN014119.D\data.ms (-2263) (-)






Abundance Scan 2505 (23.336 min): BN014120.D\data.ms (-2439) (-)

Benzo(a)pyrene
Concen: 0.19 ng
RT: 23.335 min Scan# 2503
Delta R.T. -0.001 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2




Tgt Ion:	252	Resp:	10150
Ion	Ratio	Lower	Upper
252	100		
253	34.9	21.9	32.9#
125	21.3	10.8	16.2#






Abundance Scan 3179 (25.643 min): BN014120.D\data.ms (-3140) (-)


Dibenzo(a,h)anthracene
Concen: 0.18 ng
RT: 25.646 min Scan# 3178
Delta R.T. 0.003 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51



Tgt Ion:	278	Resp:	10418
Ion	Ratio	Lower	Upper
278	100		
139	17.6	13.4	20.0
279	30.3	21.3	31.9




Abundance Scan 3372 (26.303 min): BN014120.D\data.ms (-3349) (-)



Benzo(g,h,i)perylene
Concen: 0.18 ng
RT: 26.310 min Scan# 3372
Delta R.T. 0.006 min
Lab File: BN014119.D
Acq: 26 Mar 2021 11:51

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Abundance Scan 3372 (26.310 min): BN014119.D\data.ms



Tgt Ion	Ratio	Lower	Upper
276	100		
277	25.3	20.1	30.1
138	22.4	17.0	25.4

Abundance Scan 3372 (26.310 min): BN014119.D\data.ms (-3312) (-)

