

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032621\
 Data File : BN014124.D
 Acq On : 26 Mar 2021 14:53
 Operator : CG/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Mar 26 15:22:38 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 26 13:01:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.692	152	8293	0.40	ng	0.00
7) Naphthalene-d8	10.454	136	27322	0.40	ng	#-0.01
13) Acenaphthene-d10	14.308	164	14358	0.40	ng	0.00
19) Phenanthrene-d10	17.055	188	26193	0.40	ng	0.00
29) Chrysene-d12	21.229	240	23712	0.40	ng	# 0.00
36) Perylene-d12	23.431	264	18525	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.284	112	95972	5.03	ng	0.00
5) Phenol-d6	6.855	99	124719	5.14	ng	0.00
8) Nitrobenzene-d5	8.832	82	95222	5.26	ng	0.00
11) 2-Methylnaphthalene-d10	12.053	152	203188	4.77	ng	0.00
14) 2,4,6-Tribromophenol	15.805	330	25300	5.40	ng	0.00
15) 2-Fluorobiphenyl	12.938	172	253324	4.72	ng	0.00
27) Fluoranthene-d10	19.079	212	344443	4.73	ng	0.00
31) Terphenyl-d14	19.680	244	253967	4.88	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.231	88	46880	4.29	ng	99
3) n-Nitrosodimethylamine	3.537	42	42578	5.81	ng	# 95
6) bis(2-Chloroethyl)ether	7.112	93	96549	4.53	ng	99
9) Naphthalene	10.505	128	323762	4.77	ng	99
10) Hexachlorobutadiene	10.796	225	58168	4.58	ng	# 99
12) 2-Methylnaphthalene	12.124	142	213918	4.69	ng	99
16) Acenaphthylene	14.029	152	299110	5.43	ng	100
17) Acenaphthene	14.372	154	195732	4.84	ng	97
18) Fluorene	15.361	166	234244	4.60	ng	100
20) 4,6-Dinitro-2-methylph...	15.437	198	21147	7.20	ng	# 67
21) 4-Bromophenyl-phenylether	16.252	248	68019	5.07	ng	94
22) Hexachlorobenzene	16.362	284	73598	4.75	ng	99
23) Atrazine	16.520	200	50988	5.55	ng	97
24) Pentachlorophenol	16.703	266	27747	7.57	ng	98
25) Phenanthrene	17.092	178	350198	4.89	ng	100
26) Anthracene	17.177	178	315716	5.36	ng	100
28) Fluoranthene	19.109	202	372099	4.68	ng	100
30) Pyrene	19.471	202	372365	5.12	ng	100
32) Benzo(a)anthracene	21.208	228	331105	4.89	ng	97
33) Chrysene	21.261	228	345527	4.73	ng	99
34) Bis(2-ethylhexyl)phtha...	21.144	149	93045	3.69	ng	100
35) Indeno(1,2,3-cd)pyrene	25.632	276	346369	4.37	ng	100
37) Benzo(b)fluoranthene	22.770	252	323767	4.82	ng	93
38) Benzo(k)fluoranthene	22.770	252	323767	4.89	ng	93
39) Benzo(a)pyrene	23.335	252	286668	4.97	ng	# 91
40) Dibenzo(a,h)anthracene	25.646	278	291220	4.74	ng	95
41) Benzo(g,h,i)perylene	26.306	276	304677	4.73	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032621\
Data File : BN014124.D
Acq On : 26 Mar 2021 14:53
Operator : CG/JU
Sample : SSTDIC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC5.0

Quant Time: Mar 26 15:22:38 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 26 13:01:59 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.692 min Scan# 610
 Delta R.T. 0.000 min
 Lab File: BN014124.D
 Acq: 26 Mar 2021 14:53




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:152 Resp: 8293
 Ion Ratio Lower Upper
 152 100
 150 150.2 121.2 181.8
 115 55.8 46.2 69.2



1,4-Dioxane
 Concen: 4.29 ng
 RT: 3.231 min Scan# 56
 Delta R.T. 0.000 min
 Lab File: BN014124.D
 Acq: 26 Mar 2021 14:53

Tgt Ion: 88 Resp: 46880
 Ion Ratio Lower Upper
 88 100
 58 80.0 63.1 94.7
 43 33.1 26.2 39.2









n-Nitrosodimethylamine
Concen: 5.81 ng
RT: 3.537 min Scan# 94
Delta R.T. -0.024 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 42 Resp: 42578
Ion Ratio Lower Upper
42 100
74 113.2 94.6 141.8
44 6.6 4.2 6.2#



2-Fluorophenol
Concen: 5.03 ng
RT: 5.284 min Scan# 311
Delta R.T. 0.000 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Tgt Ion: 112 Resp: 95972
Ion Ratio Lower Upper
112 100
64 60.0 48.3 72.5
63 31.8 25.4 38.0









Phenol-d6
 Concen: 5.14 ng
 RT: 6.855 min Scan# 506
 Delta R.T. 0.000 min
 Lab File: BN014124.D
 Acq: 26 Mar 2021 14:53




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Tgt Ion:	99	Resp:	124719
Ion	Ratio	Lower	Upper
99	100		
42	21.9	18.2	27.4
71	33.2	26.7	40.1




bis(2-Chloroethyl)ether
 Concen: 4.53 ng
 RT: 7.112 min Scan# 538
 Delta R.T. -0.008 min
 Lab File: BN014124.D
 Acq: 26 Mar 2021 14:53

Tgt Ion:	93	Resp:	96549
Ion	Ratio	Lower	Upper
93	100		
63	80.5	65.1	97.7
95	32.8	26.6	39.8





Abundance Scan 857 (10.467 min): BN014120.D\data.ms (-857#7-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.454 min Scan# 856
Delta R.T. -0.013 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	136	Resp:	27322
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.1	13.7
54	10.8	4.3	6.5#
68	8.0	3.4	5.2#



Abundance Scan 856 (10.454 min): BN014124.D\data.ms (-841) (-)





Abundance Scan 728 (8.832 min): BN014120.D\data.ms (-724#8)




Nitrobenzene-d5
Concen: 5.26 ng
RT: 8.832 min Scan# 728
Delta R.T. 0.000 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53


Tgt Ion:	82	Resp:	95222
Ion	Ratio	Lower	Upper
82	100		
128	43.7	33.5	50.3
54	48.8	40.3	60.5



Abundance Scan 728 (8.832 min): BN014124.D\data.ms (-712) (-)






Abundance Scan 860 (10.505 min): BN014120.D\data.ms (-855) (-)




Naphthalene
Concen: 4.77 ng
RT: 10.505 min Scan# 860
Delta R.T. 0.000 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:	128	Resp:	323762
Ion	Ratio	Lower	Upper
128	100		
129	10.9	9.2	13.8
127	13.1	10.8	16.2






Abundance Scan 883 (10.796 min): BN014120.D\data.ms (-875) (-)




Hexachlorobutadiene
Concen: 4.58 ng
RT: 10.796 min Scan# 883
Delta R.T. 0.000 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Tgt Ion:	225	Resp:	58168
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.8	51.6	77.4





Abundance Scan 1062 (12.053 min): BN014120.D\data.ms (-1019) (-)




2-Methylnaphthalene-d10
Concen: 4.77 ng
RT: 12.053 min Scan# 1062
Delta R.T. 0.000 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 203188
Ion Ratio Lower Upper
152 100
151 19.9 16.1 24.1



Abundance Scan 1062 (12.053 min): BN014124.D\data.ms (-1017) (-)





Abundance Scan 1078 (12.124 min): BN014120.D\data.ms (-1017) (-)





2-Methylnaphthalene
Concen: 4.69 ng
RT: 12.124 min Scan# 1078
Delta R.T. 0.000 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Tgt Ion:142 Resp: 213918
Ion Ratio Lower Upper
142 100
141 88.3 71.1 106.7
115 34.6 29.0 43.6



Abundance Scan 1078 (12.124 min): BN014124.D\data.ms (-1055) (-)









Scan 1306 (14.308 min): BN014120.D\data.ms (-1302) (-)
 Acenaphthene-d10
 Concen: 0.40 ng
 RT: 14.308 min Scan# 1306
 Delta R.T. 0.000 min
 Lab File: BN014124.D
 Acq: 26 Mar 2021 14:53




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Tgt Ion:	164	Resp:	14358
Ion	Ratio	Lower	Upper
164	100		
162	101.5	81.4	122.2
160	45.8	37.5	56.3




Scan 1424 (15.805 min): BN014120.D\data.ms (-1414) (-)
 2,4,6-Tribromophenol
 Concen: 5.40 ng
 RT: 15.805 min Scan# 1424
 Delta R.T. 0.000 min
 Lab File: BN014124.D
 Acq: 26 Mar 2021 14:53

Tgt Ion:	330	Resp:	25300
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	38.2	31.2	46.8





Abundance Scan 1198 (12.938 min): BN014120.D\data.ms (-1198) (-)




2-Fluorobiphenyl
Concen: 4.72 ng
RT: 12.938 min Scan# 1198
Delta R.T. 0.000 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	172	Resp:	253324
Ion	Ratio	Lower	Upper
172	100		
171	33.2	26.9	40.3
170	21.7	17.8	26.8



Abundance Scan 1198 (12.938 min): BN014124.D\data.ms (-1198) (-)





Abundance Scan 1284 (14.029 min): BN014120.D\data.ms (-1284) (-)




Acenaphthylene
Concen: 5.43 ng
RT: 14.029 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Tgt Ion:	152	Resp:	299110
Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.4	23.2
153	13.0	10.3	15.5




Abundance Scan 1284 (14.029 min): BN014124.D\data.ms (-1268) (-)






Abundance Scan 1311 (14.371 min): BN014120.D\data.ms (-1305) (-)

Acenaphthene
Concen: 4.84 ng
RT: 14.372 min Scan# 1311
Delta R.T. 0.000 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0







Tgt Ion:154 Resp: 195732
Ion Ratio Lower Upper
154 100
153 108.2 89.0 133.4
152 51.8 43.1 64.7




Abundance Scan 1389 (15.361 min): BN014120.D\data.ms (-1314) (-)

Fluorene
Concen: 4.60 ng
RT: 15.361 min Scan# 1389
Delta R.T. 0.000 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Tgt Ion:166 Resp: 234244
Ion Ratio Lower Upper
166 100
165 97.2 77.9 116.9
167 13.3 10.6 16.0






Abundance Scan 1526 (17.055 min): BN014120.D\data.ms (-1510) (-)




Phenanthrene-d10
Concen: 0.40 ng
RT: 17.055 min Scan# 1526
Delta R.T. 0.001 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:188 Resp: 26193
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.8 5.2 7.8






Abundance Scan 1395 (15.437 min): BN014120.D\data.ms (-1387) (-)




4,6-Dinitro-2-methylphenol
Concen: 7.20 ng
RT: 15.437 min Scan# 1395
Delta R.T. 0.000 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Tgt Ion:198 Resp: 21147
Ion Ratio Lower Upper
198 100
51 28.8 47.8 71.6#
105 29.3 35.4 53.2#






Abundance Scan 1460 (16.252 min): BN014120.D\data.ms (-1423) (-)




4-Bromophenyl-phenylether
Concen: 5.07 ng
RT: 16.252 min Scan# 1460
Delta R.T. 0.001 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:	248	Resp:	68019
Ion	Ratio	Lower	Upper
248	100		
250	100.3	77.0	115.6
141	42.5	39.7	59.5






Abundance Scan 1469 (16.362 min): BN014120.D\data.ms (-1423) (-)



Hexachlorobenzene
Concen: 4.75 ng
RT: 16.362 min Scan# 1469
Delta R.T. 0.001 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53


Tgt Ion:	284	Resp:	73598
Ion	Ratio	Lower	Upper
284	100		
142	35.8	29.2	43.8
249	30.8	24.7	37.1





Abundance Scan 1482 (16.519 min): BN014120.D\data.ms (-1425) (-)

Atrazine
Concen: 5.55 ng
RT: 16.520 min Scan# 1482
Delta R.T. 0.001 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53


Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0



Tgt Ion:	200	Resp:	50988
Ion	Ratio	Lower	Upper
200	100		
173	22.6	20.9	31.3
215	48.9	39.5	59.3





Abundance Scan 1482 (16.520 min): BN014124.D\data.ms (-1441) (-)




Abundance Scan 1497 (16.702 min): BN014120.D\data.ms (-1424) (-)

Pentachlorophenol
Concen: 7.57 ng
RT: 16.703 min Scan# 1497
Delta R.T. 0.001 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53


Tgt Ion:	266	Resp:	27747
Ion	Ratio	Lower	Upper
266	100		
264	62.0	50.2	75.2
268	63.3	52.5	78.7



Abundance Scan 1497 (16.703 min): BN014124.D\data.ms (-1489) (-)






Abundance Scan 1529 (17.091 min): BN014120.D\data.ms (-1525) (-)




Phenanthrene
Concen: 4.89 ng
RT: 17.092 min Scan# 1529
Delta R.T. 0.001 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:178 Resp: 350198
Ion Ratio Lower Upper
178 100
176 18.7 15.0 22.6
179 15.2 12.4 18.6






Abundance Scan 1536 (17.177 min): BN014120.D\data.ms (-1525) (-)




Anthracene
Concen: 5.36 ng
RT: 17.177 min Scan# 1536
Delta R.T. 0.001 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Tgt Ion:178 Resp: 315716
Ion Ratio Lower Upper
178 100
176 18.2 14.7 22.1
179 15.3 12.3 18.5






Abundance Scan 1821 (19.081 min): BN014120.D\data.ms (-1821) (-)




Fluoranthene-d10
Concen: 4.73 ng
RT: 19.079 min Scan# 1819
Delta R.T. -0.002 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:212 Resp: 344443
Ion Ratio Lower Upper
212 100
106 12.6 10.1 15.1
104 7.0 5.5 8.3






Abundance Scan 1827 (19.111 min): BN014120.D\data.ms (-1827) (-)




Fluoranthene
Concen: 4.68 ng
RT: 19.109 min Scan# 1825
Delta R.T. -0.002 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Tgt Ion:202 Resp: 372099
Ion Ratio Lower Upper
202 100
101 10.9 8.8 13.2
203 17.4 13.8 20.6






Abundance Scan 2122 (21.226 min): BN014120.D\data.ms (-2122) (-)




Chrysene-d12
Concen: 0.40 ng
RT: 21.229 min Scan# 2121
Delta R.T. 0.003 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:240 Resp: 23712
Ion Ratio Lower Upper
240 100
120 7.3 8.4 12.6#
236 25.4 21.0 31.6






Abundance Scan 1900 (19.473 min): BN014120.D\data.ms (-1930) (-)



Pyrene
Concen: 5.12 ng
RT: 19.471 min Scan# 1898
Delta R.T. -0.002 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53


Tgt Ion:202 Resp: 372365
Ion Ratio Lower Upper
202 100
200 20.4 16.2 24.4
203 17.4 13.9 20.9






Abundance Scan 1942 (19.682 min): BN014120.D\data.ms (-1934) (-)

Terphenyl-d14
Concen: 4.88 ng
RT: 19.680 min Scan# 1940
Delta R.T. -0.002 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0




Tgt Ion:	244	Resp:	253967
Ion	Ratio	Lower	Upper
244	100		
212	7.3	6.0	9.0
122	11.3	8.6	13.0






Abundance Scan 2121 (21.215 min): BN014120.D\data.ms (-2119) (-)


Benzo(a)anthracene
Concen: 4.89 ng
RT: 21.208 min Scan# 2119
Delta R.T. -0.008 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53



Tgt Ion:	228	Resp:	331105
Ion	Ratio	Lower	Upper
228	100		
226	27.4	20.6	31.0
229	18.9	16.2	24.2




Abundance Scan 2125 (21.258 min): BN014120.D\data.ms (-2125) (-)





Chrysene
Concen: 4.73 ng
RT: 21.261 min Scan# 2124
Delta R.T. 0.003 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53


Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0



Tgt Ion:	228	Resp:	345527
Ion	Ratio	Lower	Upper
228	100		
226	29.7	24.4	36.6
229	19.3	15.5	23.3






Abundance Scan 2115 (21.152 min): BN014120.D\data.ms (-2115) (-)



Bis(2-ethylhexyl)phthalate
Concen: 3.69 ng
RT: 21.144 min Scan# 2113
Delta R.T. -0.008 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53


Tgt Ion:	149	Resp:	93045
Ion	Ratio	Lower	Upper
149	100		
167	27.0	21.4	32.2
279	4.0	3.5	5.3





Abundance Scan 3176 (25.633 min): BN014120.D\data.ms (-3145) (-)

Indeno(1,2,3-cd)pyrene
Concen: 4.37 ng
RT: 25.632 min Scan# 3174
Delta R.T. -0.001 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53


Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0



Tgt Ion	Ratio	Lower	Upper
276	100		
138	24.0	19.4	29.2
277	24.9	19.8	29.6






Abundance Scan 3174 (25.632 min): BN014124.D\data.ms (-3145) (-)




Abundance Scan 2533 (23.432 min): BN014120.D\data.ms (-2487) (-)

Perylene-d12
Concen: 0.40 ng
RT: 23.431 min Scan# 2531
Delta R.T. -0.001 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53


Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.2	19.8	29.8
265	36.3	30.7	46.1



Abundance Scan 2531 (23.431 min): BN014124.D\data.ms (-2487) (-)






Abundance Scan 2340 (22.771 min): BN014120.D\data.ms (-225) (-)




Benzo(b)fluoranthene
Concen: 4.82 ng
RT: 22.770 min Scan# 2338
Delta R.T. -0.001 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Tgt Ion:252 Resp: 323767
Ion Ratio Lower Upper
252 100
253 22.1 20.6 30.8
125 9.6 9.4 14.2

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0






Abundance Scan 2352 (22.812 min): BN014120.D\data.ms (-226) (-)



Benzo(k)fluoranthene
Concen: 4.89 ng
RT: 22.770 min Scan# 2338
Delta R.T. -0.042 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53


Tgt Ion:252 Resp: 323767
Ion Ratio Lower Upper
252 100
253 22.1 20.6 30.8
125 9.6 9.5 14.3






Abundance Scan 2505 (23.336 min): BN014120.D\data.ms (-2439) (-)

Benzo(a)pyrene
Concen: 4.97 ng
RT: 23.335 min Scan# 2503
Delta R.T. -0.001 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0




Tgt Ion:	252	Resp:	286668
Ion	Ratio	Lower	Upper
252	100		
253	21.9	21.9	32.9
125	10.6	10.8	16.2#






Abundance Scan 3179 (25.643 min): BN014120.D\data.ms (-3140) (-)


Dibenzo(a,h)anthracene
Concen: 4.74 ng
RT: 25.646 min Scan# 3178
Delta R.T. 0.003 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53



Tgt Ion:	278	Resp:	291220
Ion	Ratio	Lower	Upper
278	100		
139	15.4	13.4	20.0
279	23.7	21.3	31.9




Abundance Scan 3372 (26.303 min): BN014120.D\data.ms (-3372) (-)



Benzo(g,h,i)perylene
Concen: 4.73 ng
RT: 26.306 min Scan# 3371
Delta R.T. 0.003 min
Lab File: BN014124.D
Acq: 26 Mar 2021 14:53

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Abundance Scan 3371 (26.306 min): BN014124.D\data.ms



Tgt Ion:276 Resp: 304677
Ion Ratio Lower Upper
276 100
277 23.9 20.1 30.1
138 20.5 17.0 25.4

Abundance Scan 3371 (26.306 min): BN014124.D\data.ms (-3312) (-)

