




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032621\
Data File : BN014117.D
Acq On : 26 Mar 2021 10:39
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 26 11:51:01 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 26 11:50:50 2021
Response via : Initial Calibration






TIC: BN014117.D\data.ms

(70) Pentachlorophenol (C)**16.704min (0.000) 30129.83 ng****response 298638**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	56.12
264.00	61.60	60.63
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032621\
Data File : BN014117.D
Acq On : 26 Mar 2021 10:39
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 26 11:51:01 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 26 11:50:50 2021
Response via : Initial Calibration



TIC: BN014117.D\data.ms

(77) Benzidine

19.298min (0.000) 0.00 ng

response 1520214

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.50
183.00	13.20	11.69
0.00	0.00	0.00