

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032621\
 Data File : BN014125.D
 Acq On : 26 Mar 2021 16:06
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN032621

Quant Time: Mar 26 16:39:21 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 26 15:25:59 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	111	0.00
2	1,4-Dioxane	0.400	0.405	-1.3	103	0.00
3	n-Nitrosodimethylamine	0.400	0.347	13.3	100	0.00
4 S	2-Fluorophenol	0.400	0.408	-2.0	106	0.00
5 S	Phenol-d6	0.400	0.401	-0.3	106	0.00
6	bis(2-Chloroethyl)ether	0.400	0.420	-5.0	110	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	110	0.00
8 S	Nitrobenzene-d5	0.400	0.379	5.3	105	0.00
9	Naphthalene	0.400	0.413	-3.2	110	0.00
10	Hexachlorobutadiene	0.400	0.415	-3.7	111	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.409	-2.2	109	0.00
12	2-Methylnaphthalene	0.400	0.409	-2.2	109	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	109	0.00
14 S	2,4,6-Tribromophenol	0.400	0.342	14.5	99	0.00
15 S	2-Fluorobiphenyl	0.400	0.403	-0.8	109	0.00
16	Acenaphthylene	0.400	0.370	7.5	107	0.00
17	Acenaphthene	0.400	0.389	2.8	107	0.00
18	Fluorene	0.400	0.392	2.0	107	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	107	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.330	17.5	100	0.00
21	4-Bromophenyl-phenylether	0.400	0.402	-0.5	106	0.00
22	Hexachlorobenzene	0.400	0.415	-3.7	109	0.00
23	Atrazine	0.400	0.371	7.3	104	0.00
24	Pentachlorophenol	0.400	0.289	27.8#	89	0.00
25	Phenanthrene	0.400	0.412	-3.0	108	0.00
26	Anthracene	0.400	0.390	2.5	107	0.00
27 SURR	Fluoranthene-d10	0.400	0.401	-0.3	109	0.00
28	Fluoranthene	0.400	0.403	-0.8	110	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	112	0.00
30	Pyrene	0.400	0.420	-5.0	109	0.00
31 S	Terphenyl-d14	0.400	0.430	-7.5	111	0.00
32	Benzo(a)anthracene	0.400	0.397	0.8	110	0.00
33	Chrysene	0.400	0.426	-6.5	112	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.368	8.0	106	-0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.416	-4.0	113	0.00
36 I	Perylene-d12	0.400	0.400	0.0	112	0.00
37	Benzo(b)fluoranthene	0.400	0.417	-4.2	115	0.00
38	Benzo(k)fluoranthene	0.400	0.410	-2.5	114	0.00
39 C	Benzo(a)pyrene	0.400	0.416	-4.0	118	0.00
40	Dibenzo(a,h)anthracene	0.400	0.404	-1.0	113	0.00
41	Benzo(g,h,i)perylene	0.400	0.405	-1.3	111	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN032621