

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032621\
 Data File : BN014121.D
 Acq On : 26 Mar 2021 13:04
 Operator : CG/JU
 Sample : SSTDIC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC0.8

Quant Time: Mar 26 13:37:18 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 26 13:01:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.692	152	7244	0.40	ng	0.00
7) Naphthalene-d8	10.454	136	25727	0.40	ng	#-0.01
13) Acenaphthene-d10	14.308	164	12936	0.40	ng	0.00
19) Phenanthrene-d10	17.055	188	25120	0.40	ng	0.00
29) Chrysene-d12	21.226	240	22014	0.40	ng	0.00
36) Perylene-d12	23.432	264	18591	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.284	112	14952	0.90	ng	0.00
5) Phenol-d6	6.855	99	18851	0.89	ng	0.00
8) Nitrobenzene-d5	8.832	82	14605	0.86	ng	0.00
11) 2-Methylnaphthalene-d10	12.053	152	31823	0.79	ng	0.00
14) 2,4,6-Tribromophenol	15.805	330	3231	0.77	ng	0.00
15) 2-Fluorobiphenyl	12.938	172	42387	0.88	ng	0.00
27) Fluoranthene-d10	19.081	212	53314	0.76	ng	0.00
31) Terphenyl-d14	19.682	244	42986	0.89	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.231	88	7399	0.77	ng	96
3) n-Nitrosodimethylamine	3.553	42	5778	0.90	ng	# 94
6) bis(2-Chloroethyl)ether	7.120	93	15534	0.83	ng	100
9) Naphthalene	10.505	128	52343	0.82	ng	99
10) Hexachlorobutadiene	10.796	225	9600	0.80	ng	# 99
12) 2-Methylnaphthalene	12.124	142	34071	0.79	ng	99
16) Acenaphthylene	14.029	152	39824	0.80	ng	100
17) Acenaphthene	14.371	154	29666	0.81	ng	100
18) Fluorene	15.361	166	35949	0.78	ng	100
20) 4,6-Dinitro-2-methylph...	15.437	198	2393	0.85	ng	# 80
21) 4-Bromophenyl-phenylether	16.252	248	10557	0.82	ng	98
22) Hexachlorobenzene	16.361	284	11786	0.79	ng	99
23) Atrazine	16.519	200	6979	0.79	ng	98
24) Pentachlorophenol	16.702	266	3449	0.98	ng	98
25) Phenanthrene	17.091	178	54746	0.80	ng	100
26) Anthracene	17.177	178	45301	0.80	ng	99
28) Fluoranthene	19.111	202	57983	0.76	ng	100
30) Pyrene	19.473	202	57639	0.85	ng	100
32) Benzo(a)anthracene	21.215	228	48323	0.77	ng	100
33) Chrysene	21.258	228	55103	0.81	ng	99
34) Bis(2-ethylhexyl)phtha...	21.152	149	13721	0.59	ng	100
35) Indeno(1,2,3-cd)pyrene	25.633	276	55306	0.75	ng	100
37) Benzo(b)fluoranthene	22.771	252	53310	0.79	ng	96
38) Benzo(k)fluoranthene	22.812	252	52631	0.79	ng	96
39) Benzo(a)pyrene	23.332	252	45699	0.79	ng	95
40) Dibenzo(a,h)anthracene	25.646	278	46763	0.76	ng	98
41) Benzo(g,h,i)perylene	26.303	276	49891	0.77	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032621\
Data File : BN014121.D
Acq On : 26 Mar 2021 13:04
Operator : CG/JU
Sample : SSTDIC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Quant Time: Mar 26 13:37:18 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 26 13:01:59 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.692 min Scan# 610
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:152 Resp: 7244
Ion Ratio Lower Upper
152 100
150 152.7 121.2 181.8
115 57.5 46.2 69.2



#2
1,4-Dioxane
Concen: 0.77 ng
RT: 3.231 min Scan# 56
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Tgt Ion: 88 Resp: 7399
Ion Ratio Lower Upper
88 100
58 82.2 63.1 94.7
43 34.6 26.2 39.2









#3
n-Nitrosodimethylamine
Concen: 0.90 ng
RT: 3.553 min Scan# 96
Delta R.T. -0.008 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 42 Resp: 5778
Ion Ratio Lower Upper
42 100
74 112.2 94.6 141.8
44 10.4 4.2 6.2#



#4
2-Fluorophenol
Concen: 0.90 ng
RT: 5.284 min Scan# 311
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Tgt Ion: 112 Resp: 14952
Ion Ratio Lower Upper
112 100
64 60.1 48.3 72.5
63 31.8 25.4 38.0









#5
Phenol-d6
Concen: 0.89 ng
RT: 6.855 min Scan# 506
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 99 Resp: 18851
Ion Ratio Lower Upper
99 100
42 22.3 18.2 27.4
71 33.1 26.7 40.1



#6
bis(2-Chloroethyl)ether
Concen: 0.83 ng
RT: 7.120 min Scan# 539
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Tgt Ion: 93 Resp: 15534
Ion Ratio Lower Upper
93 100
63 81.1 65.1 97.7
95 32.9 26.6 39.8









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.454 min Scan# 856
Delta R.T. -0.013 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	136	Resp:	25727
Ion Ratio	Lower	Upper	
136	100		
137	11.2	9.1	13.7
54	11.2	4.3	6.5#
68	8.2	3.4	5.2#



#8
Nitrobenzene-d5
Concen: 0.86 ng
RT: 8.832 min Scan# 728
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Tgt Ion:	82	Resp:	14605
Ion Ratio	Lower	Upper	
82	100		
128	41.6	33.5	50.3
54	49.1	40.3	60.5









#9
Naphthalene
Concen: 0.82 ng
RT: 10.505 min Scan# 860
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:128 Resp: 52343
Ion Ratio Lower Upper
128 100
129 11.0 9.2 13.8
127 13.1 10.8 16.2



#10
Hexachlorobutadiene
Concen: 0.80 ng
RT: 10.796 min Scan# 883
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Tgt Ion:225 Resp: 9600
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.0 51.6 77.4









#11
2-Methylnaphthalene-d10
Concen: 0.79 ng
RT: 12.053 min Scan# 1062
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:152 Resp: 31823
Ion Ratio Lower Upper
152 100
151 20.1 16.1 24.1



#12
2-Methylnaphthalene
Concen: 0.79 ng
RT: 12.124 min Scan# 1078
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Tgt Ion:142 Resp: 34071
Ion Ratio Lower Upper
142 100
141 89.8 71.1 106.7
115 35.7 29.0 43.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.308 min Scan# 1306
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:164 Resp: 12936
Ion Ratio Lower Upper
164 100
162 103.3 81.4 122.2
160 47.4 37.5 56.3



#14
2,4,6-Tribromophenol
Concen: 0.77 ng
RT: 15.805 min Scan# 1424
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Tgt Ion:330 Resp: 3231
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 38.6 31.2 46.8









#15
2-Fluorobiphenyl
Concen: 0.88 ng
RT: 12.938 min Scan# 1198
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:172 Resp: 42387
Ion Ratio Lower Upper
172 100
171 33.4 26.9 40.3
170 21.9 17.8 26.8



#16
Acenaphthylene
Concen: 0.80 ng
RT: 14.029 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Tgt Ion:152 Resp: 39824
Ion Ratio Lower Upper
152 100
151 19.5 15.4 23.2
153 13.1 10.3 15.5









#17
Acenaphthene
Concen: 0.81 ng
RT: 14.371 min Scan# 1311
Delta R.T. -0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:154 Resp: 29666
Ion Ratio Lower Upper
154 100
153 110.8 89.0 133.4
152 53.6 43.1 64.7



#18
Fluorene
Concen: 0.78 ng
RT: 15.361 min Scan# 1389
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Tgt Ion:166 Resp: 35949
Ion Ratio Lower Upper
166 100
165 97.7 77.9 116.9
167 13.5 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.055 min Scan# 1526
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:188	Resp:	25120
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	6.8	5.2



#20
4,6-Dinitro-2-methylphenol
Concen: 0.85 ng
RT: 15.437 min Scan# 1395
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Tgt Ion:198	Resp:	2393
Ion Ratio	Lower	Upper
198	100	
51	42.0	47.8
105	35.0	35.4









#21
4-Bromophenyl-phenylether
Concen: 0.82 ng
RT: 16.252 min Scan# 1460
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04




Instrument :
BNA_N
ClientSampled :
SSTDICC0.8


Tgt Ion:	248	Resp:	10557
Ion	Ratio	Lower	Upper
248	100		
250	98.0	77.0	115.6
141	50.3	39.7	59.5



#22
Hexachlorobenzene
Concen: 0.79 ng
RT: 16.361 min Scan# 1469
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Tgt Ion:	284	Resp:	11786
Ion	Ratio	Lower	Upper
284	100		
142	35.6	29.2	43.8
249	30.7	24.7	37.1









#23
Atrazine
Concen: 0.79 ng
RT: 16.519 min Scan# 1482
Delta R.T. -0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Tgt Ion:200 Resp: 6979




Ion	Ratio	Lower	Upper
200	100		
173	23.6	20.9	31.3
215	48.8	39.5	59.3




#24
Pentachlorophenol
Concen: 0.98 ng
RT: 16.702 min Scan# 1497
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Tgt Ion:266 Resp: 3449

Ion	Ratio	Lower	Upper
266	100		
264	62.3	50.2	75.2
268	63.5	52.5	78.7









#25
Phenanthrene
Concen: 0.80 ng
RT: 17.091 min Scan# 1529
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:178 Resp: 54746
Ion Ratio Lower Upper
178 100
176 18.7 15.0 22.6
179 15.4 12.4 18.6



#26
Anthracene
Concen: 0.80 ng
RT: 17.177 min Scan# 1536
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Tgt Ion:178 Resp: 45301
Ion Ratio Lower Upper
178 100
176 18.1 14.7 22.1
179 15.2 12.3 18.5









#27
Fluoranthene-d10
Concen: 0.76 ng
RT: 19.081 min Scan# 1821
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:212 Resp: 53314
Ion Ratio Lower Upper
212 100
106 12.5 10.1 15.1
104 7.0 5.5 8.3



#28
Fluoranthene
Concen: 0.76 ng
RT: 19.111 min Scan# 1827
Delta R.T. -0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Tgt Ion:202 Resp: 57983
Ion Ratio Lower Upper
202 100
101 11.0 8.8 13.2
203 17.3 13.8 20.6









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.226 min Scan# 2122
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	240	Resp:	22014
Ion Ratio	Lower	Upper	
240	100		
120	11.0	8.4	12.6
236	25.9	21.0	31.6



#30
Pyrene
Concen: 0.85 ng
RT: 19.473 min Scan# 1900
Delta R.T. -0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Tgt Ion:	202	Resp:	57639
Ion Ratio	Lower	Upper	
202	100		
200	20.3	16.2	24.4
203	17.5	13.9	20.9









#31
 Terphenyl-d14
 Concen: 0.89 ng
 RT: 19.682 min Scan# 1942
 Delta R.T. 0.000 min
 Lab File: BN014121.D
 Acq: 26 Mar 2021 13:04




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:244	Resp:	42986
Ion Ratio	Lower	Upper
244	100	
212	7.4	6.0 9.0
122	10.6	8.6 13.0



#32
 Benzo(a)anthracene
 Concen: 0.77 ng
 RT: 21.215 min Scan# 2121
 Delta R.T. 0.000 min
 Lab File: BN014121.D
 Acq: 26 Mar 2021 13:04

Tgt Ion:228	Resp:	48323
Ion Ratio	Lower	Upper
228	100	
226	25.6	20.6 31.0
229	20.1	16.2 24.2









#33
Chrysene
Concen: 0.81 ng
RT: 21.258 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	228	Resp:	55103
Ion Ratio	Lower	Upper	
228	100		
226	30.5	24.4	36.6
229	18.5	15.5	23.3



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.59 ng
RT: 21.152 min Scan# 2115
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Tgt Ion:	149	Resp:	13721
Ion Ratio	Lower	Upper	
149	100		
167	26.6	21.4	32.2
279	4.3	3.5	5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.75 ng
RT: 25.633 min Scan# 3176
Delta R.T. -0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:276 Resp: 55306
Ion Ratio Lower Upper
276 100
138 24.1 19.4 29.2
277 24.8 19.8 29.6



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.432 min Scan# 2533
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Tgt Ion:264 Resp: 18591
Ion Ratio Lower Upper
264 100
260 24.4 19.8 29.8
265 35.9 30.7 46.1









#37
Benzo(b)fluoranthene
Concen: 0.79 ng
RT: 22.771 min Scan# 2340
Delta R.T. 0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:252 Resp: 53310
Ion Ratio Lower Upper
252 100
253 23.4 20.6 30.8
125 10.9 9.4 14.2



#38
Benzo(k)fluoranthene
Concen: 0.79 ng
RT: 22.812 min Scan# 2352
Delta R.T. -0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Tgt Ion:252 Resp: 52631
Ion Ratio Lower Upper
252 100
253 23.4 20.6 30.8
125 10.2 9.5 14.3









#39
Benzo(a)pyrene
Concen: 0.79 ng
RT: 23.332 min Scan# 2504
Delta R.T. -0.003 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:252	Resp:	45699
Ion Ratio	Lower	Upper
252	100	
253	24.5	21.9 32.9
125	12.3	10.8 16.2



#40
Dibenzo(a,h)anthracene
Concen: 0.76 ng
RT: 25.646 min Scan# 3180
Delta R.T. 0.003 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Tgt Ion:278	Resp:	46763
Ion Ratio	Lower	Upper
278	100	
139	16.4	13.4 20.0
279	24.8	21.3 31.9





#41
Benzo(g,h,i)perylene
Concen: 0.77 ng
RT: 26.303 min Scan# 3372
Delta R.T. -0.000 min
Lab File: BN014121.D
Acq: 26 Mar 2021 13:04

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Tgt Ion:276 Resp: 49891
Ion Ratio Lower Upper
276 100
277 24.0 20.1 30.1
138 21.3 17.0 25.4

