

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032621\
 Data File : BN014125.D
 Acq On : 26 Mar 2021 16:06
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN032621

Manual Integrations
 APPROVED

mohammad
 3/29/2021 3:28:44 PM

Quant Time: Mar 26 16:39:21 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 26 15:25:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.692	152	7167	0.40	ng	0.00
7) Naphthalene-d8	10.467	136	25959	0.40	ng	0.00
13) Acenaphthene-d10	14.308	164	13550	0.40	ng	0.00
19) Phenanthrene-d10	17.055	188	25024	0.40	ng	0.00
29) Chrysene-d12	21.226	240	20679	0.40	ng	0.00
36) Perylene-d12	23.428	264	18315	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.284	112	7501	0.41	ng	0.00
5) Phenol-d6	6.855	99	9352	0.40	ng	0.00
8) Nitrobenzene-d5	8.832	82	6693	0.38	ng	0.00
11) 2-Methylnaphthalene-d10	12.053	152	16135	0.41	ng	0.00
14) 2,4,6-Tribromophenol	15.805	330	1489	0.34	ng	0.00
15) 2-Fluorobiphenyl	12.938	172	20493	0.40	ng	0.00
27) Fluoranthene-d10	19.081	212	26273	0.40	ng	0.00
31) Terphenyl-d14	19.682	244	20489	0.43	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.231	88	3730	0.40	ng	97
3) n-Nitrosodimethylamine	3.561	42	2462	0.35	ng	97
6) bis(2-Chloroethyl)ether	7.120	93	7807	0.42	ng	100
9) Naphthalene	10.505	128	26444	0.41	ng	99
10) Hexachlorobutadiene	10.796	225	4867	0.41	ng	# 99
12) 2-Methylnaphthalene	12.124	142	17157	0.41	ng	99
16) Acenaphthylene	14.029	152	19198	0.37	ng	100
17) Acenaphthene	14.371	154	14588	0.39	ng	100
18) Fluorene	15.361	166	17783	0.39	ng	100
20) 4,6-Dinitro-2-methylph...	15.437	198	1027	0.33	ng	# 89
21) 4-Bromophenyl-phenylether	16.252	248	5175	0.40	ng	98
22) Hexachlorobenzene	16.361	284	5964	0.42	ng	99
23) Atrazine	16.519	200	3266	0.37	ng	99
24) Pentachlorophenol	16.702	266	1292	0.29	ng	99
25) Phenanthrene	17.091	178	27654	0.41	ng	100
26) Anthracene	17.176	178	22022	0.39	ng	100
28) Fluoranthene	19.111	202	28328	0.40	ng	100
30) Pyrene	19.473	202	28358	0.42	ng	100
32) Benzo(a)anthracene	21.215	228	22742	0.40	ng	100
33) Chrysene	21.258	228	26964	0.43	ng	100
34) Bis(2-ethylhexyl)phtha...	21.141	149	6442	0.37	ng	99
35) Indeno(1,2,3-cd)pyrene	25.632	276	27339	0.42	ng	99
37) Benzo(b)fluoranthene	22.768	252	26923m	0.42	ng	
38) Benzo(k)fluoranthene	22.812	252	25862	0.41	ng	99
39) Benzo(a)pyrene	23.332	252	23082	0.42	ng	# 97
40) Dibenzo(a,h)anthracene	25.646	278	23025	0.40	ng	98
41) Benzo(g,h,i)perylene	26.300	276	24553	0.41	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032621\
Data File : BN014125.D
Acq On : 26 Mar 2021 16:06
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
ICVBN032621

Manual Integrations
APPROVED

mohammad
3/29/2021 3:28:44 PM

Quant Time: Mar 26 16:39:21 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 26 15:25:59 2021
Response via : Initial Calibration





#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.692 min Scan# 610
 Delta R.T. 0.000 min
 Lab File: BN014125.D
 Acq: 26 Mar 2021 16:06





Tgt Ion: 152 Resp: 7167

Ion	Ratio	Lower	Upper
152	100		
150	151.9	121.2	181.8
115	57.1	46.2	69.2

Instrument :
 BNA_N
 ClientSampled :
 ICVBN032621

Manual Integrations
 APPROVED




mohammad
 3/29/2021 3:28:44 PM




#2
 1,4-Dioxane
 Concen: 0.40 ng
 RT: 3.231 min Scan# 56
 Delta R.T. -0.000 min
 Lab File: BN014125.D
 Acq: 26 Mar 2021 16:06

Tgt Ion: 88 Resp: 3730

Ion	Ratio	Lower	Upper
88	100		
58	82.3	63.1	94.7
43	33.8	26.2	39.2





#3
n-Nitrosodimethylamine
Concen: 0.35 ng
RT: 3.561 min Scan# 97
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06





Tgt Ion: 42 Resp: 2462

Ion	Ratio	Lower	Upper
42	100		
74	115.3	94.6	141.8
44	5.6	4.2	6.2

Instrument :
BNA_N
ClientSampled :
ICVBN032621

Manual Integrations
APPROVED




mohammad
3/29/2021 3:28:44 PM




#4
2-Fluorophenol
Concen: 0.41 ng
RT: 5.284 min Scan# 311
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion: 112 Resp: 7501

Ion	Ratio	Lower	Upper
112	100		
64	59.9	48.3	72.5
63	31.4	25.4	38.0









#5
Phenol-d6
Concen: 0.40 ng
RT: 6.855 min Scan# 506
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion: 99 Resp: 9352
Ion Ratio Lower Upper
99 100
42 21.9 18.2 27.4
71 33.7 26.7 40.1

Instrument :
BNA_N
ClientSampleId :
ICVBN032621




Manual Integrations
APPROVED


mohammad
3/29/2021 3:28:44 PM



#6
bis(2-Chloroethyl)ether
Concen: 0.42 ng
RT: 7.120 min Scan# 539
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06


Tgt Ion: 93 Resp: 7807
Ion Ratio Lower Upper
93 100
63 81.1 65.1 97.7
95 33.0 26.6 39.8





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.467 min Scan# 857
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06



Instrument :
BNA_N
ClientSampleId :
ICVBN032621




Tgt Ion:	136	Resp:	25959
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.1	13.7
54	5.2	4.3	6.5
68	4.3	3.4	5.2

Manual Integrations
APPROVED



mohammad
3/29/2021 3:28:44 PM




#8
Nitrobenzene-d5
Concen: 0.38 ng
RT: 8.832 min Scan# 728
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06



Tgt Ion:	82	Resp:	6693
Ion	Ratio	Lower	Upper
82	100		
128	41.2	33.5	50.3
54	51.5	40.3	60.5









#9
Naphthalene
Concen: 0.41 ng
RT: 10.505 min Scan# 860
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion:	128	Resp:	26444
Ion	Ratio	Lower	Upper
128	100		
129	11.4	9.2	13.8
127	13.2	10.8	16.2

Instrument :
BNA_N
ClientSampleId :
ICVBN032621




Manual Integrations
APPROVED


mohammad
3/29/2021 3:28:44 PM



#10
Hexachlorobutadiene
Concen: 0.41 ng
RT: 10.796 min Scan# 883
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06





Tgt Ion:	225	Resp:	4867
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.9	51.6	77.4








#11
2-Methylnaphthalene-d10
Concen: 0.41 ng
RT: 12.053 min Scan# 1062
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion:152 Resp: 16135
Ion Ratio Lower Upper
152 100
151 20.0 16.1 24.1



#12
2-Methylnaphthalene
Concen: 0.41 ng
RT: 12.124 min Scan# 1078
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06


Tgt Ion:142 Resp: 17157
Ion Ratio Lower Upper
142 100
141 89.5 71.1 106.7
115 35.9 29.0 43.6




Instrument :
BNA_N
ClientSampleId :
ICVBN032621

Manual Integrations
APPROVED




mohammad
3/29/2021 3:28:44 PM




#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.308 min Scan# 1306
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06





Tgt Ion:164 Resp: 13550
Ion Ratio Lower Upper
164 100
162 101.5 81.4 122.2
160 46.1 37.5 56.3



#14
2,4,6-Tribromophenol
Concen: 0.34 ng
RT: 15.805 min Scan# 1424
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06




Tgt Ion:330 Resp: 1489
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 37.4 31.2 46.8



Instrument :
BNA_N
ClientSampleId :
ICVBN032621

Manual Integrations
APPROVED

mohammad
3/29/2021 3:28:44 PM







#15
2-Fluorobiphenyl
Concen: 0.40 ng
RT: 12.938 min Scan# 1198
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion:172 Resp: 20493
Ion Ratio Lower Upper
172 100
171 33.1 26.9 40.3
170 21.6 17.8 26.8

Instrument :
BNA_N
ClientSampleId :
ICVBN032621




Manual Integrations
APPROVED


mohammad
3/29/2021 3:28:44 PM



#16
Acenaphthylene
Concen: 0.37 ng
RT: 14.029 min Scan# 1284
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06


Tgt Ion:152 Resp: 19198
Ion Ratio Lower Upper
152 100
151 19.0 15.4 23.2
153 12.9 10.3 15.5





#17
Acenaphthene
Concen: 0.39 ng
RT: 14.371 min Scan# 1311
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06




Instrument :
BNA_N
ClientSampleId :
ICVBN032621



Tgt Ion:154 Resp: 14588
Ion Ratio Lower Upper
154 100
153 111.5 89.0 133.4
152 54.5 43.1 64.7




Manual Integrations
APPROVED


mohammad
3/29/2021 3:28:44 PM



#18
Fluorene
Concen: 0.39 ng
RT: 15.361 min Scan# 1389
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion:166 Resp: 17783
Ion Ratio Lower Upper
166 100
165 97.3 77.9 116.9
167 13.6 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.055 min Scan# 1526
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion:	188	Resp:	25024
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	6.7	5.2	7.8

Instrument :
BNA_N
ClientSampleId :
ICVBN032621




Manual Integrations
APPROVED


mohammad
3/29/2021 3:28:44 PM




#20
4,6-Dinitro-2-methylphenol
Concen: 0.33 ng
RT: 15.437 min Scan# 1395
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion:	198	Resp:	1027
Ion Ratio	Lower	Upper	
198	100		
51	73.1	47.8	71.6
105	46.1	35.4	53.2









#21
4-Bromophenyl-phenylether
Concen: 0.40 ng
RT: 16.252 min Scan# 1460
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06





Tgt Ion:248 Resp: 5175
Ion Ratio Lower Upper
248 100
250 98.6 77.0 115.6
141 49.8 39.7 59.5



#22
Hexachlorobenzene
Concen: 0.42 ng
RT: 16.361 min Scan# 1469
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06




Tgt Ion:284 Resp: 5964
Ion Ratio Lower Upper
284 100
142 35.9 29.2 43.8
249 31.1 24.7 37.1




Instrument :
BNA_N
ClientSampleId :
ICVBN032621

Manual Integrations
APPROVED




mohammad
3/29/2021 3:28:44 PM




#23
Atrazine
Concen: 0.37 ng
RT: 16.519 min Scan# 1482
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06





Tgt Ion:200 Resp: 3266
Ion Ratio Lower Upper
200 100
173 24.8 20.9 31.3
215 49.6 39.5 59.3



#24
Pentachlorophenol
Concen: 0.29 ng
RT: 16.702 min Scan# 1497
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06




Tgt Ion:266 Resp: 1292
Ion Ratio Lower Upper
266 100
264 63.2 50.2 75.2
268 65.0 52.5 78.7




Instrument :
BNA_N
ClientSampleId :
ICVBN032621

Manual Integrations
APPROVED




mohammad
3/29/2021 3:28:44 PM




#25
Phenanthrene
Concen: 0.41 ng
RT: 17.091 min Scan# 1529
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06





Tgt Ion:178 Resp: 27654
Ion Ratio Lower Upper
178 100
176 19.0 15.0 22.6
179 15.6 12.4 18.6



#26
Anthracene
Concen: 0.39 ng
RT: 17.176 min Scan# 1536
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06




Tgt Ion:178 Resp: 22022
Ion Ratio Lower Upper
178 100
176 18.3 14.7 22.1
179 15.2 12.3 18.5



Instrument :
BNA_N
ClientSampleId :
ICVBN032621

Manual Integrations
APPROVED

mohammad
3/29/2021 3:28:44 PM







#27
Fluoranthene-d10
Concen: 0.40 ng
RT: 19.081 min Scan# 1821
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion:212 Resp: 26273
Ion Ratio Lower Upper
212 100
106 12.7 10.1 15.1
104 7.2 5.5 8.3

Instrument :
BNA_N
ClientSampleId :
ICVBN032621




Manual Integrations
APPROVED


mohammad
3/29/2021 3:28:44 PM



#28
Fluoranthene
Concen: 0.40 ng
RT: 19.111 min Scan# 1827
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion:202 Resp: 28328
Ion Ratio Lower Upper
202 100
101 10.9 8.8 13.2
203 17.4 13.8 20.6









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.226 min Scan# 2122
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion:240 Resp: 20679
Ion Ratio Lower Upper
240 100
120 10.1 8.4 12.6
236 26.5 21.0 31.6

Instrument :
BNA_N
ClientSampleId :
ICVBN032621




Manual Integrations
APPROVED


mohammad
3/29/2021 3:28:44 PM



#30
Pyrene
Concen: 0.42 ng
RT: 19.473 min Scan# 1900
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion:202 Resp: 28358
Ion Ratio Lower Upper
202 100
200 20.4 16.2 24.4
203 17.4 13.9 20.9









#31
Terphenyl-d14
Concen: 0.43 ng
RT: 19.682 min Scan# 1942
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion:244 Resp: 20489
Ion Ratio Lower Upper
244 100
212 7.6 6.0 9.0
122 11.2 8.6 13.0

Instrument :
BNA_N
ClientSampleId :
ICVBN032621




Manual Integrations
APPROVED


mohammad
3/29/2021 3:28:44 PM



#32
Benzo(a)anthracene
Concen: 0.40 ng
RT: 21.215 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion:228 Resp: 22742
Ion Ratio Lower Upper
228 100
226 25.9 20.6 31.0
229 19.9 16.2 24.2









#33
Chrysene
Concen: 0.43 ng
RT: 21.258 min Scan# 2125
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion: 228 Resp: 26964
Ion Ratio Lower Upper
228 100
226 30.4 24.4 36.6
229 19.1 15.5 23.3

Instrument :
BNA_N
ClientSampleId :
ICVBN032621




Manual Integrations
APPROVED


mohammad
3/29/2021 3:28:44 PM



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.37 ng
RT: 21.141 min Scan# 2114
Delta R.T. -0.011 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion: 149 Resp: 6442
Ion Ratio Lower Upper
149 100
167 27.6 21.4 32.2
279 4.2 3.5 5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.42 ng
RT: 25.632 min Scan# 3176
Delta R.T. -0.000 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion:	276	Resp:	27339
Ion Ratio	Lower	Upper	
276	100		
138	23.7	19.4	29.2
277	24.8	19.8	29.6

Instrument :
BNA_N
ClientSampleId :
ICVBN032621




Manual Integrations
APPROVED


mohammad
3/29/2021 3:28:44 PM



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.428 min Scan# 2532
Delta R.T. -0.003 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion:	264	Resp:	18315
Ion Ratio	Lower	Upper	
264	100		
260	24.5	19.8	29.8
265	36.3	30.7	46.1





#37

Benzo(b)fluoranthene

Concen: 0.42 ng m





RT: 22.768 min Scan# 2339

Delta R.T. -0.003 min

Lab File: BN014125.D

Acq: 26 Mar 2021 16:06

Tgt Ion:252	Resp:	26923
Ion Ratio	Lower	Upper
252	100	
253	25.5	20.6 30.8
125	14.3	9.4 14.2#



#38

Benzo(k)fluoranthene

Concen: 0.41 ng




RT: 22.812 min Scan# 2352

Delta R.T. -0.000 min

Lab File: BN014125.D

Acq: 26 Mar 2021 16:06

Tgt Ion:252	Resp:	25862
Ion Ratio	Lower	Upper
252	100	
253	25.2	20.6 30.8
125	11.3	9.5 14.3



Instrument :


BNA_N

ClientSampleId :

ICVBN032621

Manual Integrations
APPROVED

mohammad
3/29/2021 3:28:44 PM







#39
Benzo(a)pyrene
Concen: 0.42 ng
RT: 23.332 min Scan# 2504
Delta R.T. -0.003 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion	Ratio	Lower	Upper
252	100		
253	27.5	21.9	32.9
125	16.5	10.8	16.2#

Instrument :
BNA_N
ClientSampleId :
ICVBN032621




Manual Integrations
APPROVED


mohammad
3/29/2021 3:28:44 PM



#40
Dibenzo(a,h)anthracene
Concen: 0.40 ng
RT: 25.646 min Scan# 3180
Delta R.T. 0.003 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06


Tgt Ion	Ratio	Lower	Upper
278	100		
139	16.2	13.4	20.0
279	25.2	21.3	31.9





#41
Benzo(g,h,i)perylene
Concen: 0.41 ng
RT: 26.300 min Scan# 3371
Delta R.T. -0.003 min
Lab File: BN014125.D
Acq: 26 Mar 2021 16:06

Tgt Ion:	276	Resp:	24553
Ion Ratio	Lower	Upper	
276	100		
277	24.1	20.1	30.1
138	21.2	17.0	25.4



Instrument :
BNA_N
ClientSampleId :
ICVBN032621

Manual Integrations
APPROVED

mohammad
3/29/2021 3:28:44 PM