

Data Path : Z:\HPCHEM1\BNA N\Data\BN032718\  
 Data File : BN000642.D  
 Acq On : 27 Mar 2018 12:28  
 Operator : JU/SJ  
 Sample : SSTDCCC020  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 LabSampleId :  
 SSTD02037

Quant Time: Mar 28 01:30:47 2018  
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN031918.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Tue Mar 27 03:53:32 2018  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|      | Compound                    | AvqRF | CCRF  | %Dev   | Area% | Dev(min) |
|------|-----------------------------|-------|-------|--------|-------|----------|
| 1 I  | 1,4-Dichlorobenzene-d4      | 1.000 | 1.000 | 0.0    | 53    | 0.00     |
| 2    | 1,4-Dioxane                 | 0.603 | 0.586 | 2.8    | 57    | 0.00     |
| 3 S  | 1,4-Dioxane-d8              | 0.560 | 0.556 | 0.7    | 57    | 0.00     |
| 4    | Benzaldehyde                | 0.761 | 0.832 | -9.3   | 54    | 0.00     |
| 5 S  | Phenol-d5                   | 1.866 | 1.944 | -4.2   | 55    | 0.00     |
| 6    | Phenol                      | 1.904 | 1.985 | -4.3   | 55    | 0.00     |
| 7 S  | Bis-(2-Chloroethyl)ether-d8 | 1.076 | 1.144 | -6.3   | 57    | 0.00     |
| 8    | Bis(2-Chloroethyl)ether     | 1.511 | 1.586 | -5.0   | 56    | 0.00     |
| 9 S  | 2-Chlorophenol-d4           | 1.411 | 1.443 | -2.3   | 56    | 0.00     |
| 10   | 2-Chlorophenol              | 1.440 | 1.481 | -2.8   | 56    | 0.00     |
| 11   | 2-Methylphenol              | 1.396 | 1.474 | -5.6   | 55    | 0.00     |
| 12   | 2,2'-oxybis(1-Chloropropane | 1.628 | 1.764 | -8.4   | 57    | 0.00     |
| 13 S | 4-Methylphenol-d8           | 1.453 | 1.533 | -5.5   | 55    | 0.00     |
| 14   | Acetophenone                | 2.160 | 2.402 | -11.2  | 56    | 0.00     |
| 15 P | N-Nitroso-di-n-propylamine  | 1.130 | 1.218 | -7.8   | 56    | 0.00     |
| 16   | 4-Methylphenol              | 1.495 | 1.586 | -6.1   | 55    | 0.00     |
| 17   | Hexachloroethane            | 0.601 | 0.612 | -1.8   | 56    | 0.00     |
| 18 I | Naphthalene-d8              | 1.000 | 1.000 | 0.0    | 54    | 0.00     |
| 19 S | Nitrobenzene-d5             | 0.162 | 0.161 | 0.6    | 55    | 0.00     |
| 20   | Nitrobenzene                | 0.401 | 0.409 | -2.0   | 57    | 0.00     |
| 21   | Isophorone                  | 0.764 | 0.810 | -6.0   | 57    | 0.00     |
| 22 S | 2-Nitrophenol-d4            | 0.168 | 0.173 | -3.0   | 57    | 0.00     |
| 23 C | 2-Nitrophenol               | 0.181 | 0.186 | -2.8   | 56    | 0.00     |
| 24   | 2,4-Dimethylphenol          | 0.390 | 0.397 | -1.8   | 55    | 0.00     |
| 25   | Bis(2-Chloroethoxy)methane  | 0.483 | 0.499 | -3.3   | 57    | 0.00     |
| 26 S | 2,4-Dichlorophenol-d3       | 0.302 | 0.312 | -3.3   | 56    | 0.00     |
| 27 C | 2,4-Dichlorophenol          | 0.293 | 0.304 | -3.8   | 56    | 0.00     |
| 28   | Naphthalene                 | 1.050 | 1.080 | -2.9   | 57    | 0.00     |
| 29 S | 4-Chloroaniline-d4          | 0.310 | 0.404 | -30.3# | 54    | 0.00     |
| 30   | 4-Chloroaniline             | 0.314 | 0.403 | -28.3# | 54    | 0.00     |
| 31 C | Hexachlorobutadiene         | 0.162 | 0.170 | -4.9   | 59    | 0.00     |
| 32   | Caprolactam                 | 0.107 | 0.123 | -15.0  | 58    | 0.00     |
| 33 C | 4-Chloro-3-methylphenol     | 0.344 | 0.342 | 0.6    | 52    | 0.00     |
| 34   | 2-Methylnaphthalene         | 0.701 | 0.745 | -6.3   | 57    | 0.00     |
| 35 I | Acenaphthene-d10            | 1.000 | 1.000 | 0.0    | 56    | 0.00     |
| 36   | 1,2,4,5-Tetrachlorobenzene  | 0.601 | 0.597 | 0.7    | 60    | 0.00     |
| 37   | Hexachlorocyclopentadiene   | 0.313 | 0.320 | -2.2   | 67    | 0.00     |
| 38 C | 2,4,6-Trichlorophenol       | 0.408 | 0.400 | 2.0    | 57    | 0.00     |
| 39   | 2,4,5-Trichlorophenol       | 0.425 | 0.409 | 3.8    | 56    | 0.00     |
| 40   | 1,1'-Biphenyl               | 1.703 | 1.707 | -0.2   | 59    | 0.00     |
| 41   | 2-Chloronaphthalene         | 1.312 | 1.300 | 0.9    | 59    | 0.00     |
| 42   | 2-Nitroaniline              | 0.398 | 0.418 | -5.0   | 59    | 0.00     |
| 43 S | Dimethylphthalate-d6        | 1.580 | 1.687 | -6.8   | 61    | 0.00     |
| 44   | Dimethylphthalate           | 1.563 | 1.657 | -6.0   | 60    | 0.00     |

Data Path : Z:\HPCHEM1\BNA N\Data\BN032718\  
 Data File : BN000642.D  
 Acq On : 27 Mar 2018 12:28  
 Operator : JU/SJ  
 Sample : SSTDCCC020  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 LabSampleId :  
 SSTD02037

Quant Time: Mar 28 01:30:47 2018  
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN031918.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Tue Mar 27 03:53:32 2018  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|      | Compound                    | AvqRF | CCRF  | %Dev  | Area% | Dev(min) |
|------|-----------------------------|-------|-------|-------|-------|----------|
| 45   | 2,6-Dinitrotoluene          | 0.315 | 0.335 | -6.3  | 59    | 0.00     |
| 46 S | Acenaphthylene-d8           | 2.084 | 2.119 | -1.7  | 59    | 0.00     |
| 47   | Acenaphthylene              | 2.042 | 2.074 | -1.6  | 59    | 0.00     |
| 48   | 3-Nitroaniline              | 0.304 | 0.348 | -14.5 | 58    | 0.00     |
| 49 C | Acenaphthene                | 1.392 | 1.418 | -1.9  | 59    | 0.00     |
| 50   | 2,4-Dinitrophenol           | 0.149 | 0.157 | -5.4  | 65    | 0.00     |
| 51 S | 4-Nitrophenol-d4            | 0.293 | 0.267 | 8.9   | 50    | 0.00     |
| 52   | 4-Nitrophenol               | 0.227 | 0.215 | 5.3   | 52    | 0.00     |
| 53   | Dibenzofuran                | 1.891 | 1.958 | -3.5  | 60    | 0.00     |
| 54   | 2,4-Dinitrotoluene          | 0.447 | 0.493 | -10.3 | 60    | 0.00     |
| 55   | 2,3,4,6-Tetrachlorophenol   | 0.325 | 0.318 | 2.2   | 55    | 0.00     |
| 56   | Diethylphthalate            | 1.595 | 1.731 | -8.5  | 60    | 0.00     |
| 57 S | Fluorene-d10                | 1.346 | 1.402 | -4.2  | 59    | 0.00     |
| 58   | Fluorene                    | 1.503 | 1.577 | -4.9  | 60    | 0.00     |
| 59   | 4-Chlorophenyl-phenylether  | 0.694 | 0.742 | -6.9  | 61    | 0.00     |
| 60   | 4-Nitroaniline              | 0.342 | 0.383 | -12.0 | 64    | 0.00     |
| 61 I | Phenanthrene-d10            | 1.000 | 1.000 | 0.0   | 58    | 0.00     |
| 62 S | 4,6-Dinitro-2-methylphenol- | 0.115 | 0.107 | 7.0   | 56    | 0.00     |
| 63   | 4,6-Dinitro-2-methylphenol  | 0.121 | 0.111 | 8.3   | 56    | 0.00     |
| 64   | N-Nitrosodiphenylamine      | 0.659 | 0.640 | 2.9   | 59    | 0.00     |
| 65   | 4-Bromophenyl-phenylether   | 0.204 | 0.207 | -1.5  | 62    | 0.00     |
| 66   | Hexachlorobenzene           | 0.215 | 0.226 | -5.1  | 64    | 0.00     |
| 67   | Atrazine                    | 0.199 | 0.211 | -6.0  | 57    | 0.00     |
| 68 C | Pentachlorophenol           | 0.114 | 0.116 | -1.8  | 62    | 0.00     |
| 69   | Phenanthrene                | 1.152 | 1.160 | -0.7  | 61    | 0.00     |
| 70 S | Anthracene-d10              | 0.961 | 0.976 | -1.6  | 61    | 0.00     |
| 71   | Anthracene                  | 1.181 | 1.186 | -0.4  | 61    | 0.00     |
| 72   | Carbazole                   | 1.013 | 1.061 | -4.7  | 60    | 0.00     |
| 73   | Di-n-butylphthalate         | 1.378 | 1.403 | -1.8  | 60    | 0.00     |
| 74 C | Fluoranthene                | 1.145 | 1.278 | -11.6 | 62    | 0.00     |
| 75 I | Chrysene-d12                | 1.000 | 1.000 | 0.0   | 59    | 0.00     |
| 76 S | Pyrene-d10                  | 1.104 | 1.126 | -2.0  | 63    | 0.00     |
| 77   | Pyrene                      | 1.461 | 1.433 | 1.9   | 62    | 0.00     |
| 78   | Butylbenzylphthalate        | 0.726 | 0.716 | 1.4   | 60    | 0.00     |
| 79   | 3,3'-Dichlorobenzidine      | 0.391 | 0.270 | 30.9# | 39    | 0.00     |
| 80   | Benzo(a)anthracene          | 1.333 | 1.295 | 2.9   | 61    | 0.00     |
| 81   | Bis(2-ethylhexyl)phthalate  | 1.053 | 1.022 | 2.9   | 60    | 0.00     |
| 82   | Chrysene                    | 1.249 | 1.213 | 2.9   | 61    | 0.00     |
| 83 I | Perylene-d12                | 1.000 | 1.000 | 0.0   | 55    | 0.00     |
| 84   | Di-n-octyl phthalate        | 1.683 | 1.936 | -15.0 | 60    | 0.00     |
| 85   | Benzo(b)fluoranthene        | 1.264 | 1.326 | -4.9  | 62    | 0.00     |
| 86   | Benzo(k)fluoranthene        | 1.220 | 1.235 | -1.2  | 57    | 0.00     |
| 87 S | Benzo(a)pyrene-d12          | 0.944 | 0.972 | -3.0  | 59    | 0.00     |

Data Path : Z:\HPCHEM1\BNA N\Data\BN032718\  
Data File : BN000642.D  
Acq On : 27 Mar 2018 12:28  
Operator : JU/SJ  
Sample : SSTDCCC020  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_N  
LabSampleId :  
SSTD02037

Quant Time: Mar 28 01:30:47 2018  
Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN031918.M  
Quant Title : SVOA CALIBRATION  
QLast Update : Tue Mar 27 03:53:32 2018  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min  
Max. RRF Dev : 20% Max. Rel. Area : 150%

|      | Compound               | AvgRF | CCRF  | %Dev | Area% | Dev(min) |
|------|------------------------|-------|-------|------|-------|----------|
| 88 C | Benzo(a)pyrene         | 1.189 | 1.201 | -1.0 | 58    | 0.00     |
| 89   | Indeno(1,2,3-cd)pyrene | 1.282 | 1.169 | 8.8  | 56    | 0.00     |
| 90   | Dibenzo(a,h)anthracene | 1.068 | 0.981 | 8.1  | 55    | 0.00     |
| 91   | Benzo(a,h,i)perylene   | 1.066 | 0.939 | 11.9 | 55    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0