


Data Path : Z:\HPCHEM1\BNA_N\Data\BN032718\
 Data File : BN000641.D
 Acq On : 27 Mar 2018 11:54
 Operator : JU/SJ
 Sample : DFTPP87
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP87

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN031918.M
 Title : SVOA CALIBRATION
 Last Update : Wed Mar 28 01:35:46 2018



AutoFind: Scans 2499, 2500, 2501; Background Corrected with Scan 2493

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.7	81360	PASS
68	69	0.00	2	1.7	1706	PASS
69	198	0.00	100	40.6	98015	PASS
70	69	0.00	2	0.7	673	PASS
127	198	10	80	50.3	121320	PASS
197	198	0.00	2	0.7	1668	PASS
198	198	100	100	100.0	241130	PASS
199	198	5	9	6.9	16647	PASS
275	198	10	60	24.7	59458	PASS
365	198	1	100	3.1	7502	PASS
441	443	0.01	100	80.7	29034	PASS
442	198	50	100	76.1	183405	PASS
443	442	15	24	19.6	35978	PASS

m/z	Abundance
37.00	395.0
38.00	931.0
39.00	5511.0
40.00	546.0
41.00	557.0
43.00	505.0
44.00	515.0
49.00	746.0
50.00	20208.0
51.10	79768.0
52.00	4042.0
55.00	914.0
56.00	2410.0
57.00	5923.0
61.00	1014.0
62.00	1285.0
63.10	3286.0
64.00	424.0
65.00	1828.0
67.00	387.0
68.10	1534.0
69.00	94616.0
70.00	657.0
71.00	439.0
73.00	1200.0
74.00	9352.0
75.00	15065.0
76.10	5488.0
77.10	113264.0
78.00	7930.0
79.00	7637.0
80.00	5566.0
81.00	7971.0
82.00	1839.0
83.00	2179.0
84.00	474.0
85.00	2346.0
86.00	2403.0
87.00	935.0
88.00	523.0
89.00	306.0
91.00	1786.0
92.00	1928.0
93.00	12697.0
94.00	999.0
95.00	460.0
96.00	586.0
97.10	516.0
98.00	9535.0
99.00	8362.0
100.00	747.0
101.00	4222.0
103.00	1272.0
104.00	2744.0
105.00	2566.0
106.10	986.0
107.00	33472.0
108.00	5560.0
109.00	1085.0
110.00	61344.0
111.00	9652.0
112.00	1127.0
112.90	498.0
116.00	1775.0
117.00	26984.0
118.00	2307.0
119.00	363.0
120.00	398.0
122.00	1887.0
123.00	3283.0
124.00	1454.0
125.00	1497.0
126.10	708.0
127.00	110784.0
128.00	8784.0
129.00	41408.0
130.00	3710.0
131.00	761.0
132.00	487.0
133.00	321.0
134.00	1152.0
135.00	3663.0
136.00	1571.0
137.00	1966.0
138.00	384.0
140.00	493.0
141.00	5495.0
142.00	2009.0
143.00	1173.0
144.00	447.0
145.00	408.0

Instrument :
BNA_N
ClientSampleId :
DFTPP87

146.00	944.0
147.00	2980.0
148.00	6791.0
149.00	1608.0
150.00	493.0
151.10	602.0
151.80	527.0
153.00	1670.0
154.00	1291.0
155.00	3215.0
156.00	4401.0
157.00	744.0
157.90	933.0
159.00	723.0
160.00	1656.0
161.00	2559.0
162.00	682.0
164.00	333.0
165.00	2103.0
166.00	1493.0
167.00	11056.0
168.00	6108.0
169.00	941.0
169.90	312.0
171.00	516.0
172.00	890.0
173.10	1247.0
174.00	2130.0
175.00	4077.0
176.10	1268.0
177.00	2152.0
178.00	688.0
179.00	7597.0
180.00	5182.0
181.00	2519.0
182.00	418.0
183.90	605.0
185.00	3783.0
186.00	28352.0
187.00	8083.0
188.00	962.0
189.00	1745.0
190.00	397.0
191.10	936.0
192.00	2423.0
193.00	3057.0
194.00	709.0
195.00	343.0
196.00	5965.0
197.10	1341.0
198.00	201984.0
199.00	14386.0
200.00	1162.0
201.40	878.0
203.00	1481.0
204.00	6799.0
205.10	12241.0
206.10	52144.0
207.10	8117.0
208.00	1878.0
209.00	714.0
210.00	742.0
211.00	1956.0
211.80	450.0
214.90	648.0
216.00	1059.0
217.00	13729.0
218.00	1802.0
221.00	7582.0
221.90	1751.0
223.00	2811.0
224.10	27248.0
225.10	7638.0
226.10	788.0
227.00	11313.0
228.00	1574.0
229.00	2331.0
230.00	386.0
231.00	951.0
233.90	705.0
235.00	961.0
236.00	576.0
237.00	1108.0
239.00	539.0
240.00	313.0
241.00	539.0
242.10	1562.0
243.10	1585.0
244.10	21464.0
245.10	2826.0
246.00	3913.0
247.00	776.0
249.00	636.0
252.00	346.0
252.90	434.0

Instrument :
BNA_N
ClientSampleId :
DFTPP87

254.10	1012.0
255.00	105208.0
256.00	16147.0
257.00	1369.0
258.00	6120.0
259.00	1057.0
265.00	2413.0
266.00	413.0
272.10	402.0
273.00	3398.0
274.00	8281.0
275.10	44936.0
276.10	6203.0
277.00	4104.0
278.00	637.0
283.10	468.0
284.00	318.0
285.00	734.0
293.00	818.0
295.00	342.0
296.00	12729.0
297.00	1871.0
303.10	1317.0
304.10	424.0
314.10	633.0
315.00	1404.0
316.10	730.0
321.00	429.0
323.10	3885.0
324.10	810.0
326.90	685.0
328.00	524.0
333.00	484.0
334.00	2309.0
335.00	584.0
341.00	419.0
346.00	1071.0
352.10	1201.0
353.00	859.0
354.00	1291.0
365.00	5342.0
366.00	856.0
372.10	2032.0
373.10	562.0
383.00	595.0
402.00	714.0
403.10	1224.0
404.00	424.0
421.00	822.0
422.00	1055.0
423.10	6310.0
424.10	1639.0
441.10	17992.0
442.10	118216.0
443.10	23576.0
444.10	2435.0

Instrument :
BNA_N
ClientSampleId :
DFTPP87

m/z	Abundance
37.00	347.0
38.00	1130.0
39.10	6652.0
40.00	502.0
41.00	646.0
43.10	459.0
44.00	450.0
49.10	898.0
50.00	23960.0
51.10	94728.0
52.10	4633.0
55.00	1106.0
56.00	2963.0
57.00	7347.0
58.00	330.0
61.00	1339.0
62.00	1607.0
63.00	4145.0
64.00	599.0
65.00	2046.0
67.00	419.0
68.00	2028.0
69.00	115000.0
70.00	666.0
71.00	487.0
73.10	1471.0
74.00	12410.0
75.00	18240.0
76.10	6654.0
77.10	139456.0
78.10	9759.0
79.00	8945.0
80.00	6793.0
81.00	9823.0
82.00	2676.0
83.00	2688.0
84.00	380.0
85.00	2722.0
86.00	2950.0
87.00	1328.0
88.00	610.0
91.00	2464.0
92.00	2716.0
93.00	15672.0
94.00	1188.0
95.10	541.0
96.10	779.0
97.10	527.0
98.00	11942.0
99.00	10358.0
100.00	810.0
101.00	4993.0
103.00	1741.0
104.00	3610.0
105.00	3203.0
106.00	1144.0
107.00	43352.0
108.00	6495.0
109.00	1375.0
110.00	78576.0
111.00	11697.0
112.00	1519.0
113.00	525.0
115.00	314.0
116.00	2383.0
117.00	35128.0
118.00	2750.0
119.00	503.0
120.00	568.0
121.10	306.0
122.00	2825.0
123.00	4047.0
124.00	1767.0
125.00	1692.0
126.20	794.0
127.00	143040.0
128.00	10530.0
129.00	53208.0
130.00	4947.0
131.00	941.0
132.00	429.0
133.00	372.0
134.00	1661.0
135.00	4538.0
136.00	1772.0
137.10	2441.0
138.10	428.0
139.00	351.0
140.00	560.0
141.00	6847.0
142.10	2411.0

Instrument :
BNA_N
ClientSampleId :
DFTPP87

143.00	1762.0
144.00	475.0
144.90	533.0
146.00	1429.0
147.00	3833.0
148.00	8874.0
149.00	1937.0
150.10	558.0
151.10	1006.0
151.90	733.0
153.00	1997.0
154.00	1556.0
155.00	3970.0
156.00	5644.0
157.10	1296.0
158.00	1154.0
159.00	1161.0
160.00	2219.0
161.00	3388.0
162.10	1010.0
163.10	371.0
163.90	414.0
165.00	2793.0
166.00	2185.0
167.00	14962.0
168.00	7671.0
169.10	1358.0
170.00	475.0
170.90	662.0
172.00	1361.0
173.00	1722.0
174.00	3028.0
175.00	5904.0
176.00	1569.0
177.00	2691.0
178.00	908.0
179.00	10596.0
180.00	7148.0
181.00	3398.0
182.00	545.0
184.00	782.0
185.10	5034.0
186.00	39560.0
187.00	11428.0
188.00	1353.0
189.00	2517.0
190.00	421.0
191.10	1007.0
192.00	3339.0
193.10	3737.0
194.00	904.0
195.00	542.0
196.00	8267.0
197.00	1877.0
198.00	285184.0
199.00	19792.0
200.00	1530.0
201.50	1138.0
203.00	2010.0
204.00	10008.0
205.10	17136.0
206.10	72536.0
207.10	10384.0
208.00	2598.0
209.00	842.0
210.00	991.0
211.10	2674.0
211.70	672.0
215.00	866.0
216.10	1666.0
217.00	18864.0
218.00	2464.0
221.10	10065.0
221.90	2758.0
223.00	4185.0
224.10	40296.0
225.10	10328.0
226.10	1195.0
227.00	16560.0
228.00	2358.0
229.00	3494.0
230.00	493.0
231.00	1396.0
234.00	1064.0
235.00	1215.0
236.00	912.0
237.00	1374.0
239.00	761.0
240.00	529.0
241.00	926.0
242.00	2257.0
243.10	2413.0
244.10	31032.0
245.10	4342.0
246.00	6015.0

Instrument :
BNA_N
ClientSampleId :
DFTPP87

247.00	1321.0
249.00	1294.0
251.00	363.0
252.10	328.0
253.00	844.0
255.00	157184.0
256.10	23528.0
257.00	1995.0
258.00	9619.0
259.00	1479.0
260.00	312.0
264.00	350.0
265.00	3756.0
266.00	664.0
272.10	495.0
273.10	4750.0
274.00	12774.0
275.10	70000.0
276.00	9152.0
277.10	6110.0
278.00	934.0
283.10	708.0
284.10	465.0
285.00	972.0
289.00	329.0
292.00	319.0
293.00	1351.0
294.00	317.0
295.10	563.0
296.00	19944.0
297.00	2912.0
303.10	2376.0
304.00	654.0
308.00	318.0
314.00	900.0
315.00	2045.0
316.00	1089.0
321.00	625.0
322.10	388.0
323.10	6198.0
324.10	1140.0
327.00	1184.0
328.00	656.0
332.00	446.0
333.00	606.0
334.10	3908.0
335.00	1138.0
341.00	713.0
346.00	1580.0
352.00	1936.0
353.00	1398.0
354.10	2117.0
355.10	436.0
365.00	8653.0
366.00	1355.0
371.10	435.0
372.10	3359.0
373.00	939.0
383.00	853.0
390.00	438.0
402.00	1280.0
403.10	1954.0
404.10	813.0
421.00	1734.0
422.10	1678.0
423.10	11790.0
424.10	2737.0
425.00	359.0
441.10	33240.0
442.10	206976.0
443.10	42592.0
444.10	3806.0

Instrument :
BNA_N
ClientSampleId :
DFTPP87

m/z	Abundance
38.00	875.0
39.10	4942.0
40.00	422.0
41.00	617.0
43.00	473.0
44.00	421.0
49.00	649.0
50.00	17448.0
51.10	69584.0
52.10	3575.0
55.00	744.0
56.00	2299.0
57.00	5434.0
57.90	317.0
61.00	955.0
62.00	1295.0
63.00	3165.0
64.00	463.0
65.00	1543.0
67.00	378.0
68.00	1556.0
69.00	85760.0
70.00	697.0
71.00	478.0
73.00	1308.0
74.00	8823.0
75.00	14102.0
76.00	4944.0
77.10	104296.0
78.10	6800.0
79.00	6786.0
80.00	5102.0
81.00	7299.0
82.00	2015.0
83.00	1821.0
84.00	366.0
85.00	2438.0
86.00	2318.0
87.00	1005.0
87.90	500.0
91.00	1758.0
92.00	1961.0
93.00	11695.0
94.10	765.0
95.00	441.0
96.00	679.0
97.00	464.0
98.00	9274.0
99.00	6971.0
100.00	601.0
101.00	4182.0
103.00	1293.0
104.00	2698.0
105.00	2494.0
106.00	928.0
107.00	32296.0
108.00	4730.0
109.10	1178.0
110.00	59800.0
111.00	8966.0
112.00	1019.0
112.90	457.0
116.00	1739.0
117.00	27680.0
118.00	2229.0
120.00	469.0
122.00	2486.0
123.00	3346.0
124.00	1379.0
125.00	1401.0
126.10	720.0
127.00	110136.0
128.00	8337.0
129.00	42696.0
130.00	3521.0
131.00	721.0
131.90	436.0
133.00	340.0
134.00	1460.0
135.00	3418.0
136.00	1472.0
137.00	2111.0
140.00	468.0
141.00	5391.0
142.00	2053.0
143.00	1338.0
144.00	382.0
145.00	449.0
146.00	1053.0
147.00	3133.0
148.00	7125.0

Instrument :
BNA_N
ClientSampleId :
DFTPP87

149.00	1551.0
150.00	496.0
151.10	754.0
151.70	404.0
153.00	1738.0
154.00	1394.0
155.10	3341.0
156.10	4577.0
157.10	1091.0
157.90	1101.0
159.00	968.0
160.00	1716.0
161.10	2382.0
161.90	673.0
164.00	346.0
165.00	2259.0
166.00	1678.0
167.00	12135.0
168.00	6586.0
169.00	1143.0
170.10	451.0
171.00	390.0
172.00	973.0
173.00	1501.0
174.00	2442.0
175.00	4814.0
176.00	1174.0
177.00	2192.0
178.00	893.0
179.00	8748.0
180.00	5906.0
181.00	2752.0
182.10	445.0
182.90	322.0
184.00	834.0
185.00	4335.0
186.00	31824.0
187.00	9211.0
188.00	910.0
189.00	2064.0
190.00	398.0
191.00	1019.0
192.00	2777.0
193.00	3234.0
194.00	811.0
195.00	504.0
196.00	6887.0
197.10	1787.0
198.00	236224.0
199.00	15764.0
200.00	1186.0
201.50	1086.0
203.00	1686.0
204.00	8176.0
205.10	14929.0
206.10	60080.0
207.10	8521.0
208.00	2194.0
209.00	729.0
210.00	1117.0
211.00	2369.0
215.00	724.0
216.10	1520.0
217.00	16640.0
218.00	2019.0
221.10	8767.0
221.80	2198.0
223.00	3422.0
224.10	33648.0
225.10	8916.0
226.10	907.0
227.00	14589.0
228.00	1829.0
229.00	2834.0
229.90	321.0
231.00	1238.0
234.00	848.0
235.00	968.0
235.90	700.0
237.00	1218.0
239.00	614.0
239.90	461.0
241.00	735.0
242.10	2053.0
243.10	1859.0
244.10	26232.0
245.10	3501.0
246.00	5271.0
247.00	950.0
249.00	1009.0
251.00	310.0
252.90	574.0
255.00	138496.0
256.10	19816.0
257.00	1612.0

Instrument :
BNA_N
ClientSampleId :
DFTPP87

258.00	7806.0
259.00	1385.0
265.10	3312.0
266.00	451.0
272.20	396.0
273.00	4071.0
274.10	11306.0
275.10	63440.0
276.10	8022.0
277.00	5797.0
278.00	909.0
283.00	606.0
284.00	405.0
285.00	884.0
293.00	1309.0
295.10	394.0
296.00	18056.0
297.00	2658.0
303.10	2154.0
304.10	602.0
314.00	885.0
315.00	1899.0
316.10	1148.0
321.00	679.0
323.10	5421.0
324.10	1142.0
326.90	1075.0
328.00	626.0
332.00	480.0
333.10	497.0
334.10	3769.0
335.00	1100.0
341.10	672.0
346.00	1430.0
352.10	1769.0
353.10	1265.0
354.10	2015.0
355.00	532.0
365.00	8511.0
366.00	1245.0
371.00	511.0
372.10	3317.0
373.10	854.0
383.10	979.0
390.00	425.0
402.10	1423.0
403.10	1976.0
404.00	679.0
421.00	1691.0
422.10	1740.0
423.10	11496.0
424.10	2612.0
425.10	314.0
441.10	35872.0
442.10	225024.0
443.10	41768.0
444.10	3949.0

Instrument :
BNA_N
ClientSampleId :
DFTPP87

Scan 2493 (18.101 min): BN000641.D
DFTPP87

m/z	Abundance
40.00	340.0
41.00	468.0
43.00	524.0
43.90	434.0
55.00	535.0
57.00	580.0
69.10	443.0
71.10	374.0
73.10	480.0
83.10	305.0
207.10	793.0

Instrument :
BNA_N
ClientSampleId :
DFTPP87