

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032725\
 Data File : BN036771.D
 Acq On : 28 Mar 2025 13:58
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Mar 28 14:23:03 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 16:06:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.695	152	1636	0.400	ng	-0.03
7) Naphthalene-d8	10.487	136	4177	0.400	ng	#-0.02
13) Acenaphthene-d10	14.334	164	2615	0.400	ng	-0.03
19) Phenanthrene-d10	17.086	188	5571	0.400	ng	-0.02
29) Chrysene-d12	21.277	240	4555	0.400	ng	#-0.02
35) Perylene-d12	23.516	264	4167	0.400	ng	#-0.04
System Monitoring Compounds						
4) 2-Fluorophenol	5.290	112	1384	0.363	ng	-0.02
5) Phenol-d6	6.872	99	1653	0.351	ng	-0.03
8) Nitrobenzene-d5	8.854	82	1480	0.326	ng	-0.02
11) 2-Methylnaphthalene-d10	12.080	152	2256	0.363	ng	-0.03
14) 2,4,6-Tribromophenol	15.833	330	464	0.391	ng	-0.02
15) 2-Fluorobiphenyl	12.963	172	5285	0.347	ng	-0.03
27) Fluoranthene-d10	19.118	212	5637	0.395	ng	-0.02
31) Terphenyl-d14	19.722	244	3716	0.341	ng	-0.02
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.225	88	752	0.414	ng	97
3) n-Nitrosodimethylamine	3.535	42	1423	0.388	ng	97
6) bis(2-Chloroethyl)ether	7.125	93	1671	0.343	ng	98
9) Naphthalene	10.530	128	4359	0.355	ng	98
10) Hexachlorobutadiene	10.818	225	1044	0.361	ng	# 99
12) 2-Methylnaphthalene	12.156	142	2874	0.368	ng	98
16) Acenaphthylene	14.056	152	4358	0.353	ng	99
17) Acenaphthene	14.398	154	2946	0.365	ng	100
18) Fluorene	15.393	166	4111	0.376	ng	98
20) 4,6-Dinitro-2-methylph...	15.478	198	428	0.446	ng	87
21) 4-Bromophenyl-phenylether	16.280	248	1300	0.372	ng	96
22) Hexachlorobenzene	16.391	284	1537	0.365	ng	96
23) Atrazine	16.553	200	1063	0.380	ng	97
24) Pentachlorophenol	16.739	266	702	0.365	ng	95
25) Phenanthrene	17.124	178	6367	0.381	ng	99
26) Anthracene	17.223	178	5569	0.369	ng	100
28) Fluoranthene	19.150	202	7617	0.406	ng	99
30) Pyrene	19.513	202	7795	0.350	ng	99
32) Benzo(a)anthracene	21.259	228	5566	0.351	ng	99
33) Chrysene	21.313	228	6726	0.389	ng	99
34) Bis(2-ethylhexyl)phtha...	21.196	149	3968	0.352	ng	# 99
36) Indeno(1,2,3-cd)pyrene	25.782	276	5233	0.348	ng	96
37) Benzo(b)fluoranthene	22.844	252	5720	0.377	ng	97
38) Benzo(k)fluoranthene	22.844	252	5720	0.360	ng	96
39) Benzo(a)pyrene	23.420	252	4867	0.381	ng	96
40) Dibenzo(a,h)anthracene	25.800	278	3914	0.334	ng	96
41) Benzo(g,h,i)perylene	26.475	276	4758	0.355	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032725\
Data File : BN036771.D
Acq On : 28 Mar 2025 13:58
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Quant Time: Mar 28 14:23:03 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 16:06:28 2025
Response via : Initial Calibration







#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.695 min Scan# 61
 Delta R.T. -0.029 min
 Lab File: BN036771.D
 Acq: 28 Mar 2025 13:58

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC







Tgt Ion:152 Resp: 1636
 Ion Ratio Lower Upper
 152 100
 150 156.3 123.7 185.5
 115 72.2 54.3 81.5



#2
 1,4-Dioxane
 Concen: 0.414 ng
 RT: 3.225 min Scan# 20
 Delta R.T. -0.014 min
 Lab File: BN036771.D
 Acq: 28 Mar 2025 13:58

Tgt Ion: 88 Resp: 752
 Ion Ratio Lower Upper
 88 100
 43 47.5 37.8 56.8
 58 80.9 67.4 101.2








#3
 n-Nitrosodimethylamine
 Concen: 0.388 ng
 RT: 3.535 min Scan# 61
 Delta R.T. -0.014 min
 Lab File: BN036771.D
 Acq: 28 Mar 2025 13:58




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion: 42 Resp: 1423
 Ion Ratio Lower Upper
 42 100
 74 72.9 60.6 90.8
 44 6.7 6.3 9.5



#4
 2-Fluorophenol
 Concen: 0.363 ng
 RT: 5.290 min Scan# 306
 Delta R.T. -0.022 min
 Lab File: BN036771.D
 Acq: 28 Mar 2025 13:58

Tgt Ion: 112 Resp: 1384
 Ion Ratio Lower Upper
 112 100
 64 69.4 53.1 79.7
 63 42.8 31.8 47.8









#5
Phenol-d6
Concen: 0.351 ng
RT: 6.872 min Scan# 51
Delta R.T. -0.029 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:	99	Resp:	1653
Ion	Ratio	Lower	Upper
99	100		
42	37.9	26.5	39.7
71	45.1	34.1	51.1



#6
bis(2-Chloroethyl)ether
Concen: 0.343 ng
RT: 7.125 min Scan# 560
Delta R.T. -0.029 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58

Tgt Ion:	93	Resp:	1671
Ion	Ratio	Lower	Upper
93	100		
63	87.3	67.7	101.5
95	31.8	25.6	38.4








#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.487 min Scan# 91
Delta R.T. -0.021 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:136	Resp:	4177
Ion	Ratio	Lower Upper
136	100	
137	12.6	10.3 15.5
54	11.4	11.5 17.3#
68	7.0	7.0 10.4



#8
Nitrobenzene-d5
Concen: 0.326 ng
RT: 8.854 min Scan# 775
Delta R.T. -0.021 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58

Tgt Ion: 82	Resp:	1480
Ion	Ratio	Lower Upper
82	100	
128	45.0	30.6 45.8
54	71.5	52.2 78.4








#9
Naphthalene
Concen: 0.355 ng
RT: 10.530 min Scan# 91
Delta R.T. -0.032 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:128 Resp: 4359
Ion Ratio Lower Upper
128 100
129 13.0 9.8 14.6
127 15.4 11.8 17.8



#10
Hexachlorobutadiene
Concen: 0.361 ng
RT: 10.818 min Scan# 959
Delta R.T. -0.032 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58

Tgt Ion:225 Resp: 1044
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 51.8 77.8








#11
2-Methylnaphthalene-d10
Concen: 0.363 ng
RT: 12.080 min Scan# 1136
Delta R.T. -0.030 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:152 Resp: 2256
Ion Ratio Lower Upper
152 100
151 21.0 17.0 25.6



#12
2-Methylnaphthalene
Concen: 0.368 ng
RT: 12.156 min Scan# 1151
Delta R.T. -0.025 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58

Tgt Ion:142 Resp: 2874
Ion Ratio Lower Upper
142 100
141 91.2 71.7 107.5
115 49.5 38.3 57.5








#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.334 min Scan# 1443
Delta R.T. -0.032 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:164 Resp: 2615
Ion Ratio Lower Upper
164 100
162 107.8 84.2 126.2
160 55.8 42.2 63.2



#14
2,4,6-Tribromophenol
Concen: 0.391 ng
RT: 15.833 min Scan# 1579
Delta R.T. -0.025 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58

Tgt Ion:330 Resp: 464
Ion Ratio Lower Upper
330 100
332 93.3 75.2 112.8
141 49.1 43.4 65.2








#15
2-Fluorobiphenyl
Concen: 0.347 ng
RT: 12.963 min Scan# 11
Delta R.T. -0.025 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:172 Resp: 5285
Ion Ratio Lower Upper
172 100
171 37.1 29.5 44.3
170 26.1 20.2 30.4



#16
Acenaphthylene
Concen: 0.353 ng
RT: 14.056 min Scan# 1417
Delta R.T. -0.021 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58

Tgt Ion:152 Resp: 4358
Ion Ratio Lower Upper
152 100
151 19.5 16.2 24.4
153 13.1 10.6 15.8








#17
Acenaphthene
Concen: 0.365 ng
RT: 14.398 min Scan# 1449
Delta R.T. -0.032 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:154 Resp: 2946
Ion Ratio Lower Upper
154 100
153 117.7 94.1 141.1
152 61.6 49.8 74.6



#18
Fluorene
Concen: 0.376 ng
RT: 15.393 min Scan# 1542
Delta R.T. -0.021 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58

Tgt Ion:166 Resp: 4111
Ion Ratio Lower Upper
166 100
165 97.9 79.8 119.8
167 13.4 10.6 15.8








#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.086 min Scan# 1
Delta R.T. -0.025 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:188 Resp: 5571
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.6 8.8 13.2



#20
4,6-Dinitro-2-methylphenol
Concen: 0.446 ng
RT: 15.478 min Scan# 1550
Delta R.T. -0.021 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58

Tgt Ion:198 Resp: 428
Ion Ratio Lower Upper
198 100
51 115.3 107.9 161.9
105 66.0 56.2 84.2








#21
4-Bromophenyl-phenylether
Concen: 0.372 ng
RT: 16.280 min Scan# 1615
Delta R.T. -0.025 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:248 Resp: 1300
Ion Ratio Lower Upper
248 100
250 93.6 73.0 109.6
141 91.8 68.6 103.0



#22
Hexachlorobenzene
Concen: 0.365 ng
RT: 16.391 min Scan# 1624
Delta R.T. -0.025 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58

Tgt Ion:284 Resp: 1537
Ion Ratio Lower Upper
284 100
142 49.5 37.0 55.4
249 36.2 28.1 42.1








#23
Atrazine
Concen: 0.380 ng
RT: 16.553 min Scan# 1637
Delta R.T. -0.025 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:200 Resp: 1063
Ion Ratio Lower Upper
200 100
173 37.3 27.3 40.9
215 46.3 36.8 55.2



#24
Pentachlorophenol
Concen: 0.365 ng
RT: 16.739 min Scan# 1652
Delta R.T. -0.037 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58

Tgt Ion:266 Resp: 702
Ion Ratio Lower Upper
266 100
264 64.8 49.6 74.4
268 69.2 50.9 76.3





#25

Phenanthrene

Concen: 0.381 ng

RT: 17.124 min Scan# 1683

Delta R.T. -0.025 min

Lab File: BN036771.D

Acq: 28 Mar 2025 13:58

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:178 Resp: 6367

Ion Ratio Lower Upper

178 100

176 20.0 15.9 23.9

179 15.8 12.2 18.4



#26

Anthracene

Concen: 0.369 ng

RT: 17.223 min Scan# 1691

Delta R.T. -0.025 min

Lab File: BN036771.D

Acq: 28 Mar 2025 13:58




Tgt Ion:178 Resp: 5569

Ion Ratio Lower Upper

178 100

176 19.4 15.4 23.2

179 15.4 12.6 18.8








#27
Fluoranthene-d10
Concen: 0.395 ng
RT: 19.118 min Scan# 1925
Delta R.T. -0.023 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
212	100		
106	13.8	11.8	17.6
104	8.6	7.3	10.9



#28
Fluoranthene
Concen: 0.406 ng
RT: 19.150 min Scan# 1932
Delta R.T. -0.023 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58

Tgt Ion	Ratio	Lower	Upper
202	100		
101	11.0	9.4	14.0
203	16.8	13.5	20.3








#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.277 min Scan# 21
Delta R.T. -0.018 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:240 Resp: 4555
Ion Ratio Lower Upper
240 100
120 13.3 14.6 22.0#
236 30.0 24.1 36.1



#30
Pyrene
Concen: 0.350 ng
RT: 19.513 min Scan# 2010
Delta R.T. -0.023 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58

Tgt Ion:202 Resp: 7795
Ion Ratio Lower Upper
202 100
200 21.2 17.1 25.7
203 17.3 14.1 21.1








#31
 Terphenyl-d14
 Concen: 0.341 ng
 RT: 19.722 min Scan# 2060
 Delta R.T. -0.023 min
 Lab File: BN036771.D
 Acq: 28 Mar 2025 13:58




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:244 Resp: 3716
 Ion Ratio Lower Upper
 244 100
 212 12.0 9.6 14.4
 122 17.2 13.9 20.9



#32
 Benzo(a)anthracene
 Concen: 0.351 ng
 RT: 21.259 min Scan# 2256
 Delta R.T. -0.027 min
 Lab File: BN036771.D
 Acq: 28 Mar 2025 13:58

Tgt Ion:228 Resp: 5566
 Ion Ratio Lower Upper
 228 100
 226 28.8 22.5 33.7
 229 20.4 16.6 25.0





#33

Chrysene

Concen: 0.389 ng

RT: 21.313 min Scan# 21

Delta R.T. -0.018 min

Lab File: BN036771.D

Acq: 28 Mar 2025 13:58

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:228 Resp: 6726

Ion Ratio Lower Upper

228 100

226 32.1 25.3 37.9

229 20.0 15.8 23.8



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.352 ng

RT: 21.196 min Scan# 2249

Delta R.T. -0.018 min

Lab File: BN036771.D

Acq: 28 Mar 2025 13:58




Tgt Ion:149 Resp: 3968

Ion Ratio Lower Upper

149 100

167 26.1 20.7 31.1

279 3.3 3.6 5.4#








#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.516 min Scan# 21
 Delta R.T. -0.038 min
 Lab File: BN036771.D
 Acq: 28 Mar 2025 13:58




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:264 Resp: 4167
 Ion Ratio Lower Upper
 264 100
 260 27.7 22.6 33.8
 265 72.6 88.1 132.1#



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.348 ng
 RT: 25.782 min Scan# 3539
 Delta R.T. -0.055 min
 Lab File: BN036771.D
 Acq: 28 Mar 2025 13:58

Tgt Ion:276 Resp: 5233
 Ion Ratio Lower Upper
 276 100
 138 25.8 23.4 35.2
 277 25.2 20.0 30.0








#37
Benzo(b)fluoranthene
Concen: 0.377 ng
RT: 22.844 min Scan# 2534
Delta R.T. -0.029 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:252 Resp: 5720
Ion Ratio Lower Upper
252 100
253 28.6 23.9 35.9
125 20.5 17.4 26.2



#38
Benzo(k)fluoranthene
Concen: 0.360 ng
RT: 22.844 min Scan# 2534
Delta R.T. -0.073 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58

Tgt Ion:252 Resp: 5720
Ion Ratio Lower Upper
252 100
253 28.6 24.6 36.8
125 20.5 17.8 26.8








#39
Benzo(a)pyrene
Concen: 0.381 ng
RT: 23.420 min Scan# 21
Delta R.T. -0.035 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:252 Resp: 4867
Ion Ratio Lower Upper
252 100
253 32.6 27.8 41.8
125 26.1 22.7 34.1



#40
Dibenzo(a,h)anthracene
Concen: 0.334 ng
RT: 25.800 min Scan# 3545
Delta R.T. -0.055 min
Lab File: BN036771.D
Acq: 28 Mar 2025 13:58


Tgt Ion:278 Resp: 3914
Ion Ratio Lower Upper
278 100
139 24.4 20.8 31.2
279 33.5 28.8 43.2





#41
 Benzo(g,h,i)perylene
 Concen: 0.355 ng
 RT: 26.475 min Scan# 31
 Delta R.T. -0.055 min
 Lab File: BN036771.D
 Acq: 28 Mar 2025 13:58

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC



Tgt Ion:	276	Resp:	4758
Ion	Ratio	Lower	Upper
276	100		
277	25.7	22.2	33.4
138	27.0	24.1	36.1

