

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032725\
 Data File : BN036729.D
 Acq On : 27 Mar 2025 09:49
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 27 10:52:22 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 16:06:28 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	131	-0.02
2	1,4-Dioxane	0.400	0.384	4.0	112	-0.01
3	n-Nitrosodimethylamine	0.400	0.346	13.5	109	-0.01
4 S	2-Fluorophenol	0.400	0.377	5.8	117	-0.02
5 S	Phenol-d6	0.400	0.373	6.8	125	-0.03
6	bis(2-Chloroethyl)ether	0.400	0.367	8.3	121	-0.03
7 I	Naphthalene-d8	0.400	0.400	0.0	140	-0.02
8 S	Nitrobenzene-d5	0.400	0.341	14.8	126	-0.03
9	Naphthalene	0.400	0.357	10.8	122	-0.03
10	Hexachlorobutadiene	0.400	0.343	14.2	113	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.360	10.0	124	-0.03
12	2-Methylnaphthalene	0.400	0.359	10.3	123	-0.03
13 I	Acenaphthene-d10	0.400	0.400	0.0	141	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.368	8.0	125	-0.02
15 S	2-Fluorobiphenyl	0.400	0.343	14.2	117	-0.03
16	Acenaphthylene	0.400	0.341	14.8	117	-0.02
17	Acenaphthene	0.400	0.353	11.8	120	-0.03
18	Fluorene	0.400	0.354	11.5	118	-0.02
19 I	Phenanthrene-d10	0.400	0.400	0.0	148	-0.02
20	4,6-Dinitro-2-methylphenol	0.400	0.427	-6.7	138	-0.02
21	4-Bromophenyl-phenylether	0.400	0.360	10.0	122	-0.02
22	Hexachlorobenzene	0.400	0.354	11.5	117	-0.01
23	Atrazine	0.400	0.385	3.8	134	-0.02
24	Pentachlorophenol	0.400	0.412	-3.0	153	-0.04
25	Phenanthrene	0.400	0.384	4.0	131	-0.02
26	Anthracene	0.400	0.367	8.3	128	-0.02
27 SURR	Fluoranthene-d10	0.400	0.433	-8.2	147	-0.02
28	Fluoranthene	0.400	0.435	-8.7	149	-0.02
29 I	Chrysene-d12	0.400	0.400	0.0	189	-0.02
30	Pyrene	0.400	0.340	15.0	147	-0.02
31 S	Terphenyl-d14	0.400	0.342	14.5	151	-0.02
32	Benzo(a)anthracene	0.400	0.362	9.5	165	-0.03
33	Chrysene	0.400	0.370	7.5	165	-0.02
34	Bis(2-ethylhexyl)phthalate	0.400	0.436	-9.0	195	-0.02
35 I	Perylene-d12	0.400	0.400	0.0	190	-0.03
36	Indeno(1,2,3-cd)pyrene	0.400	0.345	13.8	153	-0.05
37	Benzo(b)fluoranthene	0.400	0.377	5.8	169	-0.03
38	Benzo(k)fluoranthene	0.400	0.349	12.8	156	-0.03
39 C	Benzo(a)pyrene	0.400	0.373	6.8	167	-0.03
40	Dibenzo(a,h)anthracene	0.400	0.343	14.2	157	-0.05
41	Benzo(g,h,i)perylene	0.400	0.333	16.8	147	-0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4