

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032725\
 Data File : BN036736.D
 Acq On : 27 Mar 2025 14:02
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 27 14:28:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 16:06:28 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	90	-0.02
2	1,4-Dioxane	0.400	0.442	-10.5	89	-0.01
3	n-Nitrosodimethylamine	0.400	0.377	5.8	82	0.00
4 S	2-Fluorophenol	0.400	0.373	6.8	80	-0.02
5 S	Phenol-d6	0.400	0.350	12.5	81	-0.03
6	bis(2-Chloroethyl)ether	0.400	0.352	12.0	80	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	96	-0.02
8 S	Nitrobenzene-d5	0.400	0.332	17.0	83	-0.02
9	Naphthalene	0.400	0.357	10.8	83	-0.03
10	Hexachlorobutadiene	0.400	0.365	8.8	82	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.356	11.0	84	-0.03
12	2-Methylnaphthalene	0.400	0.355	11.3	83	-0.03
13 I	Acenaphthene-d10	0.400	0.400	0.0	98	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.383	4.3	91	-0.02
15 S	2-Fluorobiphenyl	0.400	0.344	14.0	82	-0.02
16	Acenaphthylene	0.400	0.340	15.0	82	-0.02
17	Acenaphthene	0.400	0.360	10.0	85	-0.03
18	Fluorene	0.400	0.365	8.8	85	-0.02
19 I	Phenanthrene-d10	0.400	0.400	0.0	105	-0.02
20	4,6-Dinitro-2-methylphenol	0.400	0.444	-11.0	103	-0.01
21	4-Bromophenyl-phenylether	0.400	0.364	9.0	87	-0.01
22	Hexachlorobenzene	0.400	0.354	11.5	83	-0.01
23	Atrazine	0.400	0.387	3.3	95	-0.01
24	Pentachlorophenol	0.400	0.370	7.5	98	-0.02
25	Phenanthrene	0.400	0.382	4.5	92	-0.02
26	Anthracene	0.400	0.370	7.5	91	-0.02
27 SURR	Fluoranthene-d10	0.400	0.399	0.3	96	-0.02
28	Fluoranthene	0.400	0.407	-1.7	99	-0.02
29 I	Chrysene-d12	0.400	0.400	0.0	127	-0.02
30	Pyrene	0.400	0.348	13.0	101	-0.02
31 S	Terphenyl-d14	0.400	0.338	15.5	100	-0.02
32	Benzo(a)anthracene	0.400	0.354	11.5	109	-0.03
33	Chrysene	0.400	0.387	3.3	116	-0.02
34	Bis(2-ethylhexyl)phthalate	0.400	0.349	12.8	105	-0.02
35 I	Perylene-d12	0.400	0.400	0.0	134	-0.03
36	Indeno(1,2,3-cd)pyrene	0.400	0.362	9.5	113	-0.05
37	Benzo(b)fluoranthene	0.400	0.378	5.5	119	-0.03
38	Benzo(k)fluoranthene	0.400	0.419	-4.7	132	-0.03
39 C	Benzo(a)pyrene	0.400	0.368	8.0	116	-0.03
40	Dibenzo(a,h)anthracene	0.400	0.361	9.8	117	-0.05
41	Benzo(g,h,i)perylene	0.400	0.372	7.0	116	-0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4