

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032725\
 Data File : BN036755.D
 Acq On : 28 Mar 2025 01:52
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 28 05:36:33 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 16:06:28 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	-0.02
2	1,4-Dioxane	0.444	0.464	-4.5	80	-0.01
3	n-Nitrosodimethylamine	0.898	0.887	1.2	82	-0.01
4 S	2-Fluorophenol	0.932	0.839	10.0	73	-0.02
5 S	Phenol-d6	1.152	1.005	12.8	77	-0.03
6	bis(2-Chloroethyl)ether	1.190	1.034	13.1	75	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	93	-0.02
8 S	Nitrobenzene-d5	0.435	0.357	17.9	80	-0.02
9	Naphthalene	1.177	1.044	11.3	80	-0.03
10	Hexachlorobutadiene	0.277	0.253	8.7	80	-0.02
11 SURR	2-Methylnaphthalene-d10	0.595	0.516	13.3	79	-0.03
12	2-Methylnaphthalene	0.749	0.675	9.9	82	-0.03
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	-0.03
14 S	2,4,6-Tribromophenol	0.182	0.170	6.6	92	-0.02
15 S	2-Fluorobiphenyl	2.327	1.853	20.4	78	-0.03
16	Acenaphthylene	1.888	1.645	12.9	86	-0.02
17	Acenaphthene	1.236	1.101	10.9	87	-0.03
18	Fluorene	1.672	1.522	9.0	88	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	107	-0.02
20	4,6-Dinitro-2-methylphenol	0.086	0.076	11.6	106	-0.02
21	4-Bromophenyl-phenylether	0.251	0.227	9.6	89	-0.02
22	Hexachlorobenzene	0.303	0.272	10.2	86	-0.01
23	Atrazine	0.201	0.192	4.5	96	-0.02
24	Pentachlorophenol	0.138	0.130	5.8	102	-0.04
25	Phenanthrene	1.200	1.161	3.2	96	-0.02
26	Anthracene	1.083	1.013	6.5	94	-0.02
27 SURR	Fluoranthene-d10	1.025	1.003	2.1	96	-0.02
28	Fluoranthene	1.348	1.356	-0.6	100	-0.02
29 I	Chrysene-d12	1.000	1.000	0.0	143	-0.02
30	Pyrene	1.956	1.559	20.3	104	-0.02
31 S	Terphenyl-d14	0.958	0.759	20.8	105	-0.02
32	Benzo(a)anthracene	1.391	1.230	11.6	122	-0.03
33	Chrysene	1.520	1.471	3.2	130	-0.02
34	Bis(2-ethylhexyl)phthalate	0.990	0.859	13.2	117	-0.02
35 I	Perylene-d12	1.000	1.000	0.0	160#	-0.03
36	Indeno(1,2,3-cd)pyrene	1.444	1.338	7.3	138	-0.05
37	Benzo(b)fluoranthene	1.456	1.390	4.5	143	-0.03
38	Benzo(k)fluoranthene	1.527	1.443	5.5	142	-0.03
39 C	Benzo(a)pyrene	1.226	1.197	2.4	147	-0.03
40	Dibenzo(a,h)anthracene	1.124	1.038	7.7	142	-0.05
41	Benzo(g,h,i)perylene	1.286	1.220	5.1	141	-0.06

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4