


Data Path : Z:\HPCHEM1\BNA_N\DATA\BN032818\
Data File : BN000656.D
Acq On : 28 Mar 2018 13:26
Operator : JU/SJ
Sample : DFTPP88
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN031918.M
Title : SVOA CALIBRATION
Last Update : Thu Mar 29 07:07:58 2018



AutoFind: Scans 2499, 2500, 2501; Background Corrected with Scan 2494

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.5	42458	PASS
68	69	0.00	2	1.8	928	PASS
69	198	0.00	100	41.4	50832	PASS
70	69	0.00	2	0.8	386	PASS
127	198	10	80	50.4	61962	PASS
197	198	0.00	2	0.7	874	PASS
198	198	100	100	100.0	122930	PASS
199	198	5	9	6.9	8529	PASS
275	198	10	60	24.6	30224	PASS
365	198	1	100	3.1	3757	PASS
441	443	0.01	100	80.8	14373	PASS
442	198	50	100	74.4	91472	PASS
443	442	15	24	19.4	17781	PASS

m/z	Abundance
38.00	556.0
39.10	3233.0
40.00	439.0
41.00	308.0
49.10	468.0
50.00	11705.0
51.00	44856.0
52.00	2260.0
55.00	548.0
56.00	1399.0
57.00	3286.0
61.00	729.0
62.10	705.0
63.00	2022.0
64.00	315.0
65.00	938.0
68.00	925.0
69.00	52344.0
70.00	405.0
73.00	593.0
74.00	5489.0
75.00	8531.0
76.10	3084.0
77.10	62272.0
78.10	4644.0
79.00	4379.0
80.00	2856.0
81.00	4740.0
82.00	1208.0
83.00	1135.0
85.00	1155.0
86.00	1366.0
87.00	603.0
91.00	1251.0
92.00	1150.0
93.00	6989.0
94.00	474.0
96.00	480.0
98.00	5410.0
99.00	4486.0
99.90	402.0
101.00	2323.0
103.00	801.0
104.00	1450.0
105.00	1481.0
106.00	543.0
107.00	18184.0
108.00	3011.0
109.00	642.0
110.00	34128.0
111.00	5387.0
112.00	649.0
116.00	965.0
117.00	14678.0
118.00	1432.0
121.90	1131.0
123.00	2048.0
123.90	860.0
124.90	812.0
126.10	368.0
127.00	61512.0
128.00	4688.0
129.00	21848.0
130.00	2248.0
131.00	393.0
134.00	663.0
135.00	1833.0
136.00	775.0
137.00	1250.0
140.10	321.0
141.00	2822.0
142.00	1158.0
143.00	744.0
146.00	466.0
147.00	1707.0
148.00	3583.0
148.90	884.0
151.10	389.0
152.90	1033.0
154.00	948.0
155.00	1645.0
156.00	2544.0
157.00	524.0
158.00	552.0
159.00	434.0
160.00	937.0
161.00	1406.0
162.00	408.0
165.00	1217.0
166.00	812.0
167.00	6041.0

Instrument :
BNA_N
ClientSampleId :

168.00	3952.0
169.00	741.0
172.00	485.0
172.90	804.0
174.00	1232.0
175.00	2349.0
175.90	577.0
177.00	958.0
178.00	385.0
179.00	4109.0
180.00	3060.0
180.90	1248.0
184.00	350.0
185.00	1735.0
186.00	15371.0
187.00	4578.0
188.00	549.0
189.00	882.0
191.00	604.0
192.00	1249.0
193.00	1557.0
194.00	313.0
196.00	3224.0
197.20	711.0
198.00	111064.0
199.00	8489.0
200.00	598.0
201.60	523.0
203.00	625.0
204.00	3846.0
205.10	6736.0
206.10	28936.0
207.00	4187.0
208.00	1237.0
209.00	412.0
210.00	441.0
211.10	1118.0
215.00	326.0
216.00	653.0
217.00	7220.0
218.00	864.0
221.10	3702.0
221.90	858.0
223.00	1635.0
224.10	15166.0
225.10	3990.0
226.10	440.0
227.00	6003.0
228.00	923.0
229.00	1428.0
231.00	557.0
234.00	523.0
235.00	497.0
237.00	626.0
239.00	336.0
241.10	343.0
242.00	670.0
243.00	946.0
244.10	11376.0
245.10	1796.0
246.00	2278.0
247.00	448.0
249.00	489.0
254.10	693.0
255.00	58064.0
256.10	8487.0
257.00	705.0
258.00	3584.0
259.00	555.0
265.00	1313.0
273.00	1691.0
274.00	4644.0
275.10	24872.0
276.00	3328.0
277.00	2249.0
278.10	441.0
283.00	314.0
285.00	309.0
293.00	517.0
296.00	7318.0
297.10	1015.0
303.10	743.0
314.00	427.0
315.10	755.0
316.00	447.0
323.10	2112.0
324.10	471.0
327.00	427.0
334.00	1332.0
335.10	408.0
345.90	463.0
352.10	802.0
353.10	425.0
354.00	657.0
365.00	2728.0

Instrument :
BNA_N
ClientSampleId :

366.00	464.0
372.00	1195.0
402.00	426.0
403.00	822.0
421.00	461.0
422.00	573.0
423.10	3630.0
424.00	989.0
441.10	9343.0
442.10	62232.0
443.10	12361.0
444.10	1157.0

Instrument :
BNA_N
ClientSampleId :

m/z	Abundance
38.00	611.0
39.10	3452.0
40.00	489.0
41.00	359.0
49.00	475.0
50.00	13327.0
51.00	49832.0
52.00	2735.0
55.00	481.0
56.00	1598.0
57.00	3780.0
61.00	699.0
62.10	890.0
63.00	2372.0
64.00	384.0
65.10	1194.0
68.00	1033.0
69.00	61016.0
70.00	328.0
73.00	915.0
74.00	5812.0
75.00	9944.0
76.00	3625.0
77.10	73944.0
78.10	4768.0
79.00	4514.0
80.00	3584.0
81.00	4884.0
82.00	1451.0
83.00	1279.0
85.00	1391.0
86.00	1564.0
87.00	760.0
91.00	1307.0
92.00	1196.0
93.00	8003.0
94.00	640.0
96.00	527.0
98.00	6513.0
99.00	5182.0
100.00	425.0
101.00	2859.0
102.90	1011.0
104.00	1855.0
105.00	1810.0
106.00	519.0
107.00	21176.0
108.00	3460.0
109.00	706.0
110.00	40184.0
111.00	5835.0
112.00	666.0
116.00	1271.0
117.00	18568.0
118.00	1455.0
121.90	1444.0
123.00	2323.0
124.00	1026.0
124.90	1065.0
126.20	482.0
127.00	73696.0
128.00	5346.0
129.00	27840.0
130.00	2282.0
131.00	467.0
134.00	829.0
135.00	2349.0
136.00	905.0
137.00	1284.0
139.90	340.0
141.00	3739.0
142.00	1304.0
143.00	874.0
145.90	800.0
147.00	2010.0
148.00	4847.0
149.00	957.0
150.00	324.0
151.00	390.0
151.80	319.0
153.00	1188.0
154.00	1017.0
155.00	1854.0
156.10	2949.0
157.00	675.0
158.00	704.0
159.00	520.0
160.00	1065.0
161.00	1552.0
162.00	461.0
165.00	1486.0

Instrument :
BNA_N
ClientSampleId :

166.00	1086.0
167.00	7342.0
168.00	4671.0
169.00	832.0
172.00	713.0
173.00	904.0
174.00	1520.0
175.00	2972.0
176.00	822.0
177.00	1556.0
178.00	469.0
179.00	4965.0
180.00	3876.0
181.00	1526.0
183.90	426.0
185.10	2427.0
186.10	20448.0
187.00	5744.0
188.00	671.0
189.00	1429.0
191.10	712.0
192.00	1518.0
193.00	1920.0
194.00	455.0
196.00	4489.0
197.20	1149.0
198.00	146752.0
199.00	9969.0
200.00	702.0
201.40	537.0
203.00	958.0
204.00	5445.0
205.10	9298.0
206.10	37288.0
207.10	5590.0
208.00	1536.0
209.10	515.0
210.30	595.0
211.10	1498.0
215.00	454.0
216.00	755.0
217.00	9882.0
218.00	1312.0
221.00	5267.0
221.80	1475.0
223.00	2188.0
224.10	19760.0
225.10	5187.0
226.00	603.0
227.00	8681.0
228.10	1367.0
229.00	1791.0
231.00	731.0
233.90	545.0
235.10	586.0
236.00	424.0
237.00	735.0
239.00	307.0
241.00	451.0
242.00	1132.0
243.00	1192.0
244.10	15583.0
245.00	2332.0
246.00	3208.0
247.00	631.0
248.90	552.0
253.10	420.0
254.20	805.0
255.00	81400.0
256.10	11759.0
257.00	966.0
258.00	4460.0
259.00	742.0
265.00	1694.0
266.00	310.0
273.00	2304.0
274.00	6722.0
275.10	36336.0
276.00	4826.0
277.00	3086.0
278.00	502.0
282.90	351.0
285.10	592.0
293.00	876.0
296.00	10882.0
297.00	1596.0
303.10	1301.0
304.00	336.0
314.00	405.0
315.00	1119.0
316.00	635.0
321.00	312.0
323.10	3406.0
324.00	547.0
327.00	729.0

Instrument :
BNA_N
ClientSampleId :

328.00	378.0
334.10	1856.0
335.00	547.0
341.00	410.0
346.00	695.0
352.00	972.0
353.10	698.0
354.10	1130.0
365.00	4384.0
366.10	614.0
371.00	325.0
372.00	1760.0
373.00	509.0
383.00	493.0
402.00	727.0
403.10	975.0
421.00	823.0
422.00	876.0
423.10	6171.0
424.10	1369.0
441.10	17008.0
442.10	107920.0
443.10	21336.0
444.00	2095.0

Instrument :
BNA_N
ClientSampleId :

m/z	Abundance
38.10	440.0
39.10	2386.0
39.90	493.0
41.00	319.0
49.00	457.0
50.00	8234.0
51.00	32688.0
52.00	1775.0
55.00	495.0
56.00	1042.0
57.00	2287.0
60.90	546.0
62.00	509.0
63.00	1506.0
65.00	859.0
68.00	826.0
69.00	39136.0
70.10	426.0
73.00	655.0
74.00	3842.0
75.00	6722.0
76.00	2374.0
77.10	48184.0
78.10	3344.0
79.00	3487.0
80.00	2278.0
81.00	3369.0
82.00	827.0
83.00	839.0
85.00	875.0
86.00	1001.0
87.00	499.0
91.00	1004.0
92.00	973.0
93.00	5473.0
94.00	422.0
96.10	381.0
98.00	4152.0
99.00	3547.0
100.00	365.0
101.00	1989.0
102.90	732.0
103.90	1241.0
105.00	1270.0
106.10	407.0
107.00	15044.0
108.00	2295.0
109.00	580.0
110.00	27528.0
111.00	4399.0
112.00	594.0
116.00	876.0
117.00	12752.0
118.00	981.0
122.00	1081.0
123.00	1496.0
123.90	665.0
125.00	634.0
126.20	336.0
127.00	50680.0
128.00	3741.0
129.00	18872.0
130.00	1639.0
131.00	324.0
134.00	559.0
135.00	1873.0
136.00	695.0
137.00	971.0
141.00	2636.0
142.00	872.0
143.00	628.0
145.90	585.0
147.00	1422.0
148.00	3218.0
149.00	671.0
151.10	408.0
153.00	913.0
154.00	640.0
155.00	1591.0
156.00	2296.0
157.00	434.0
157.90	619.0
159.00	372.0
160.00	722.0
161.00	1224.0
162.00	438.0
165.00	1039.0
166.10	782.0
167.00	5844.0
168.00	3296.0
169.10	416.0

Instrument :
BNA_N
ClientSampleId :

171.90	491.0
173.00	559.0
174.00	1245.0
175.00	2008.0
176.00	769.0
177.00	1069.0
178.00	386.0
179.00	4091.0
180.00	2884.0
181.00	1333.0
183.90	365.0
185.00	2003.0
186.00	15093.0
187.00	4246.0
188.00	463.0
189.00	1075.0
190.90	535.0
192.00	1264.0
193.00	1357.0
193.90	326.0
196.00	3017.0
197.00	764.0
198.00	110976.0
199.00	7130.0
200.10	509.0
201.50	563.0
203.00	748.0
204.00	3763.0
205.00	6748.0
206.10	28144.0
207.10	3994.0
208.00	1213.0
209.00	341.0
210.20	457.0
211.00	1016.0
215.00	345.0
216.00	695.0
217.00	7482.0
218.00	956.0
221.10	3869.0
222.00	867.0
223.00	1807.0
224.10	15906.0
225.00	4033.0
226.00	465.0
227.00	6857.0
228.00	799.0
229.00	1356.0
231.00	509.0
233.90	337.0
235.00	531.0
236.00	328.0
237.00	482.0
239.00	327.0
241.00	411.0
242.00	980.0
243.10	941.0
244.10	12941.0
245.10	1692.0
246.00	2256.0
246.90	525.0
248.90	515.0
254.10	582.0
255.00	64944.0
256.00	8985.0
257.00	849.0
258.00	3832.0
259.00	665.0
265.00	1563.0
273.00	1837.0
274.00	5593.0
275.00	29464.0
276.10	3513.0
277.00	2612.0
278.00	506.0
285.10	410.0
293.00	582.0
296.00	8507.0
297.00	1059.0
303.00	959.0
314.00	369.0
315.00	922.0
316.10	459.0
323.10	2521.0
324.00	501.0
327.00	487.0
334.10	1871.0
335.00	409.0
341.00	349.0
346.00	715.0
352.10	912.0
353.00	659.0
354.00	1087.0
365.00	4159.0
366.00	639.0

Instrument :
BNA_N
ClientSampleId :

372.10	1556.0
373.00	426.0
383.10	407.0
402.10	649.0
403.00	912.0
403.90	327.0
421.00	774.0
422.00	858.0
423.10	6120.0
424.10	1277.0
441.10	16768.0
442.10	104264.0
443.10	19648.0
444.10	1540.0

Instrument :
BNA_N
ClientSampleId :

Scan 2494 (18.107 min): BN000656.D
DFTPP88

m/z	Abundance
40.00	362.0
44.00	303.0
207.10	660.0

Instrument :
BNA_N
ClientSampleId :