

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032822\
 Data File : BN019202.D
 Acq On : 28 Mar 2022 19:23
 Operator : CG/JU
 Sample : PB143655BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB143655BL

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 03/29/2022
 Supervised By : Jagrut Upadhyay 03/29/2022

Quant Time: Mar 29 06:24:17 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 25 19:03:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.847	152	3666	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	8860	0.400	ng	# 0.01
13) Acenaphthene-d10	14.484	164	5245	0.400	ng	0.00
19) Phenanthrene-d10	17.236	188	9559	0.400	ng	0.00
29) Chrysene-d12	21.422	240	7359	0.400	ng	# 0.00
35) Perylene-d12	23.787	264	6415	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.406	112	2806	0.324	ng	0.00
5) Phenol-d6	7.024	99	2095m	0.211	ng	0.02
8) Nitrobenzene-d5	9.025	82	1562	0.227	ng	0.03
11) 2-Methylnaphthalene-d10	12.257	152	4377	0.309	ng	0.02
14) 2,4,6-Tribromophenol	15.995	330	526	0.206	ng	0.02
15) 2-Fluorobiphenyl	13.127	172	6101m	0.277	ng	0.02
27) Fluoranthene-d10	19.261	212	9801	0.350	ng	0.00
31) Terphenyl-d14	19.855	244	6702	0.408	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032822\
Data File : BN019202.D
Acq On : 28 Mar 2022 19:23
Operator : CG/JU
Sample : PB143655BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB143655BL

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 03/29/2022
Supervised By :Jagrut Upadhyay 03/29/2022

Quant Time: Mar 29 06:24:17 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 25 19:03:12 2022
Response via : Initial Calibration

