

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032823\
 Data File : BN024565.D
 Acq On : 27 Mar 2023 16:16
 Operator : CG/JU
 Sample : 02067-11
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BFZ84

Quant Time: Mar 28 01:00:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN030323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 28 00:57:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.010	152	290771	20.000	ng/ul	0.00
20) Naphthalene-d8	10.828	136	1250170	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.651	164	756366	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.398	188	1650894	20.000	ng/ul	0.00
79) Chrysene-d12	21.586	240	1502614	20.000	ng/ul	0.00
88) Perylene-d12	24.086	264	1607191	20.000	ng/ul	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	33023	4.430	ng/uL	0.00
4) Pyridine-d5	3.822	84	103404	5.013	ng/ul	0.01
7) Phenol-d5	7.163	99	144189	5.528	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.334	67	495560	29.670	ng/ul	0.00
11) 2-Chlorophenol-d4	7.540	132	421024	21.020	ng/ul	0.00
15) 4-Methylphenol-d8	8.716	113	269503	12.998	ng/ul	0.00
21) Nitrobenzene-d5	9.181	128	269451	33.207	ng/ul	0.00
24) 2-Nitrophenol-d4	9.905	143	274608	31.035	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.446	165	465561	24.383	ng/ul	0.00
31) 4-Chloroaniline-d4	10.963	131	606349	21.285	ng/ul	0.00
46) Dimethylphthalate-d6	14.057	166	1844750	31.602	ng/ul	0.00
49) Acenaphthylene-d8	14.345	160	1870098	27.523	ng/ul	0.00
54) 4-Nitrophenol-d4	14.840	143	49802	4.465	ng/ul	0.00
60) Fluorene-d10	15.640	176	1566300	31.927	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.757	200	273418	30.740	ng/ul	0.00
73) Anthracene-d10	17.498	188	2881475	37.009	ng/ul	0.00
81) Pyrene-d10	19.786	212	3294922	37.520	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.921	264	3073473	38.469	ng/ul	0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032823\
Data File : BN024565.D
Acq On : 27 Mar 2023 16:16
Operator : CG/JU
Sample : 02067-11
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BFZ84

Quant Time: Mar 28 01:00:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN030323.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 28 00:57:32 2023
Response via : Initial Calibration

