


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032824\
Data File : BN030516.D
Acq On : 28 Mar 2024 09:17
Operator : MA/JU
Sample : DFTPP308
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP308

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
Title : SVOA CALIBRATION
Last Update : Fri Mar 29 02:15:53 2024



AutoFind: Scans 2458, 2459, 2460; Background Corrected with Scan 2452

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.1	239775	PASS
68	69	0.00	2	1.8	4267	PASS
69	198	0.00	100	37.7	243605	PASS
70	69	0.00	2	0.6	1379	PASS
127	198	10	80	48.2	310976	PASS
197	198	0.00	2	0.4	2763	PASS
198	198	100	100	100.0	645504	PASS
199	198	5	9	6.9	44635	PASS
275	198	10	60	24.4	157192	PASS
365	198	1	100	3.4	22241	PASS
441	198	0.01	100	9.8	63035	PASS
442	198	50	100	62.0	400021	PASS
443	442	15	24	18.8	75115	PASS

m/z	Abundance
39.10	204.0
40.00	2949.0
41.00	336.0
43.00	270.0
44.00	7757.0
44.90	271.0
48.90	466.0
51.00	268.0
55.00	228.0
67.10	213.0
73.00	2725.0
74.00	304.0
75.10	341.0
76.90	312.0
78.00	200.0
78.90	224.0
81.10	283.0
82.80	162.0
83.90	339.0
85.80	162.0
88.90	175.0
91.00	195.0
95.90	709.0
97.00	174.0
102.80	183.0
115.10	175.0
118.90	173.0
131.00	174.0
133.00	616.0
133.90	179.0
134.90	170.0
147.00	1291.0
149.00	197.0
176.90	216.0
190.90	552.0
191.90	203.0
192.90	273.0
194.00	153.0
207.00	4289.0
207.90	1069.0
208.90	644.0
221.00	544.0
222.00	160.0
267.00	264.0
281.00	1093.0
281.90	381.0
340.90	222.0
355.00	325.0

Instrument :
BNA_N
ClientSampleId :
DFTPP308

m/z	Abundance
36.00	258.0
37.00	1062.0
38.00	2833.0
39.00	16069.0
40.00	2886.0
41.00	695.0
41.90	183.0
42.90	353.0
44.00	6704.0
45.00	505.0
47.00	194.0
49.00	1842.0
50.00	54816.0
51.00	212608.0
52.00	11654.0
53.00	377.0
55.00	1419.0
56.00	6533.0
57.00	13447.0
58.00	546.0
58.90	343.0
59.90	263.0
60.90	2697.0
62.00	2732.0
63.00	8317.0
64.00	1215.0
65.00	4416.0
65.90	348.0
66.90	335.0
68.00	3680.0
69.00	213696.0
70.00	1153.0
70.90	202.0
71.80	189.0
73.00	2588.0
74.00	22648.0
75.00	36200.0
76.00	12412.0
77.00	255360.0
78.00	17760.0
79.00	15690.0
80.00	12811.0
81.00	17632.0
82.00	4059.0
83.00	4131.0
84.00	748.0
84.90	2793.0
85.90	4760.0
87.00	2560.0
87.90	986.0
88.90	431.0
91.00	3916.0
92.00	4555.0
93.00	28624.0
94.00	2112.0
94.90	528.0
96.00	1734.0
97.00	594.0
98.00	23448.0
99.00	20208.0
99.90	1736.0
101.00	9488.0
102.00	405.0
102.90	3128.0
104.00	5652.0
105.00	5548.0
106.00	2118.0
107.00	72216.0
108.00	11882.0
109.00	2044.0
110.00	133568.0
111.00	21072.0
112.00	2678.0
113.00	903.0
114.10	178.0
114.90	379.0
116.00	3775.0
117.00	70104.0
118.00	5806.0
118.90	885.0
119.90	1127.0
120.90	418.0
122.00	4511.0
122.90	7580.0
123.90	3259.0
125.00	3091.0
127.00	257664.0
128.00	20000.0
129.00	92296.0
130.00	8374.0
131.00	1785.0
132.00	913.0

Instrument :
BNA_N
ClientSampleId :
DFTPP308

132.70	732.0
134.00	3213.0
135.00	8502.0
136.00	3309.0
137.00	5361.0
137.90	919.0
138.90	491.0
140.00	1383.0
141.00	12179.0
142.00	4317.0
143.00	3133.0
143.90	651.0
144.90	769.0
146.00	2453.0
147.00	7498.0
148.00	16512.0
148.90	3896.0
149.90	912.0
151.00	1743.0
151.80	1124.0
153.00	4146.0
154.00	2977.0
155.00	6860.0
156.00	10319.0
157.00	2151.0
157.90	2483.0
158.90	1842.0
159.90	4183.0
161.00	5780.0
162.00	1578.0
162.90	517.0
163.90	592.0
165.00	4622.0
166.00	3752.0
167.00	25224.0
168.00	19392.0
169.00	2619.0
170.00	810.0
170.90	1186.0
172.00	1920.0
173.00	2752.0
174.00	5209.0
175.00	9333.0
176.00	2424.0
177.00	4872.0
177.90	1521.0
178.90	18160.0
180.00	13304.0
181.00	6387.0
181.90	951.0
182.80	557.0
184.00	1427.0
185.00	8385.0
186.00	65104.0
187.00	20024.0
188.00	2431.0
189.00	4446.0
189.90	898.0
191.00	2343.0
192.00	5782.0
193.00	7070.0
193.90	1357.0
195.00	726.0
196.00	13440.0
197.00	3467.0
197.90	482048.0
198.90	34816.0
200.00	2755.0
201.40	2428.0
202.90	2888.0
204.00	16856.0
205.00	29864.0
206.00	119096.0
207.00	22872.0
208.00	5093.0
209.00	1821.0
211.00	5264.0
212.90	358.0
214.90	1589.0
216.00	2685.0
217.00	32664.0
218.00	4156.0
219.00	550.0
221.00	18720.0
221.90	4647.0
223.00	7100.0
224.00	65616.0
225.00	17880.0
226.00	2232.0
227.00	28976.0
228.00	3984.0
228.90	5265.0
230.00	725.0
231.00	2431.0
231.90	337.0
232.90	755.0

Instrument :
BNA_N
ClientSampleId :
DFTPP308

234.00	1734.0
234.90	2099.0
235.90	1473.0
236.90	3242.0
237.90	446.0
238.90	1133.0
239.90	1096.0
241.00	1537.0
242.00	3718.0
243.00	3725.0
244.00	51200.0
245.00	7691.0
245.90	10300.0
246.90	2267.0
247.90	417.0
249.00	1721.0
250.00	393.0
251.00	417.0
251.90	547.0
253.00	1188.0
255.00	255744.0
256.00	36640.0
257.10	2748.0
258.00	15323.0
258.90	2545.0
259.90	425.0
260.90	444.0
263.90	429.0
265.00	5807.0
266.00	1004.0
266.80	219.0
268.00	157.0
269.90	251.0
271.00	852.0
271.90	665.0
273.00	6874.0
274.00	21032.0
275.00	102424.0
276.00	16720.0
276.90	9727.0
277.90	1915.0
278.90	373.0
281.00	1010.0
281.90	528.0
283.00	1323.0
283.90	753.0
285.00	1527.0
286.00	341.0
288.00	156.0
288.90	456.0
289.90	322.0
290.90	236.0
291.90	500.0
292.90	1824.0
293.90	739.0
294.90	669.0
296.00	32568.0
296.90	3850.0
298.10	289.0
301.00	427.0
302.00	564.0
303.00	3038.0
303.90	975.0
304.90	258.0
307.90	445.0
309.10	176.0
310.00	288.0
313.00	230.0
314.00	1587.0
314.90	2802.0
316.00	1676.0
317.00	424.0
321.00	906.0
322.10	537.0
323.00	9686.0
323.90	1530.0
325.00	347.0
326.00	344.0
326.90	1761.0
327.90	857.0
328.90	198.0
332.00	895.0
333.00	811.0
334.00	5078.0
335.00	1637.0
335.90	304.0
339.00	163.0
341.00	1387.0
342.10	263.0
342.90	170.0
345.90	1706.0
347.00	549.0
352.00	2559.0
353.00	1582.0
354.00	2608.0
355.00	832.0

Instrument :
BNA_N
ClientSampleId :
DFTPP308

355.90	195.0
358.90	233.0
364.90	13083.0
365.90	2060.0
366.90	178.0
369.90	194.0
370.90	572.0
372.00	4838.0
373.00	1225.0
373.90	172.0
382.90	1013.0
383.90	516.0
389.80	593.0
390.90	315.0
391.90	340.0
400.90	280.0
401.90	1703.0
403.00	2487.0
404.00	987.0
404.90	286.0
420.90	2024.0
421.90	1746.0
423.00	14163.0
423.90	3942.0
425.00	856.0
441.00	38120.0
442.00	212864.0
443.00	45840.0
444.10	4191.0
445.00	256.0

Instrument :
BNA_N
ClientSampleId :
DFTPP308

m/z	Abundance
35.90	299.0
37.00	1403.0
38.00	3197.0
39.00	20456.0
40.00	2983.0
41.00	745.0
43.00	299.0
44.00	6207.0
45.00	666.0
47.90	257.0
49.00	2026.0
50.00	73576.0
51.00	277248.0
52.00	14463.0
53.00	651.0
55.00	1554.0
56.00	7579.0
57.00	16744.0
57.90	858.0
59.00	323.0
60.00	282.0
61.00	3550.0
62.00	3992.0
63.00	11298.0
64.00	1646.0
65.00	5408.0
66.00	370.0
66.90	591.0
68.00	5049.0
69.00	280768.0
70.00	1716.0
70.80	199.0
73.00	3003.0
74.00	30128.0
75.00	45656.0
76.00	16087.0
77.00	340352.0
78.00	23792.0
79.00	22472.0
80.00	17080.0
81.00	23512.0
82.00	5728.0
83.00	5498.0
84.00	898.0
85.00	3418.0
86.00	6716.0
87.00	2504.0
87.90	1463.0
89.00	581.0
90.10	227.0
91.00	5208.0
92.00	6146.0
93.00	39976.0
94.00	2678.0
95.00	845.0
96.00	1947.0
97.00	728.0
98.00	31944.0
99.00	24808.0
100.00	2204.0
101.00	13545.0
101.90	774.0
103.00	4669.0
104.00	8500.0
105.00	7983.0
106.00	2631.0
107.00	105384.0
108.00	15322.0
109.10	2984.0
110.00	179520.0
111.00	28664.0
112.00	3674.0
112.90	1235.0
114.10	210.0
114.90	603.0
116.00	5398.0
117.00	98728.0
118.00	7546.0
118.90	1126.0
119.90	1114.0
121.00	563.0
122.00	6821.0
123.00	10728.0
123.90	4765.0
125.00	4663.0
127.00	356864.0
128.00	27624.0
129.00	140672.0
130.00	12148.0
131.00	2143.0
132.00	1195.0
132.80	826.0

Instrument :
BNA_N
ClientSampleId :
DFTPP308

134.00	4053.0
135.00	11891.0
136.00	4503.0
137.00	6418.0
137.90	1282.0
139.00	674.0
140.00	1698.0
141.00	18992.0
142.00	6357.0
142.90	4190.0
144.00	1311.0
144.90	937.0
146.00	3448.0
147.00	10011.0
148.00	23456.0
149.00	5460.0
150.00	1470.0
151.10	2273.0
151.90	1591.0
152.90	4826.0
154.00	4231.0
155.00	10604.0
156.00	15416.0
157.10	2889.0
158.00	2974.0
159.00	2882.0
159.90	5597.0
161.00	8808.0
162.00	2257.0
163.00	751.0
164.00	827.0
165.00	6771.0
166.00	5826.0
167.00	41392.0
168.00	24008.0
169.00	3669.0
170.00	1316.0
170.90	1712.0
172.00	3391.0
173.00	4471.0
174.00	7324.0
175.00	14829.0
176.00	4190.0
176.90	6511.0
177.90	2280.0
178.90	27224.0
180.00	21056.0
181.00	8606.0
181.90	1181.0
182.90	817.0
183.90	2409.0
185.00	13034.0
186.00	98392.0
187.00	31032.0
188.00	3193.0
189.00	6376.0
190.00	1033.0
191.00	3220.0
192.00	9037.0
193.00	10833.0
194.00	2267.0
195.00	1590.0
196.00	21080.0
197.00	4823.0
197.90	735424.0
198.90	51672.0
199.90	3859.0
201.50	3353.0
203.00	5139.0
204.00	26552.0
205.00	47056.0
206.00	192640.0
207.00	29528.0
208.00	7447.0
209.00	2699.0
210.00	3386.0
211.00	7644.0
213.00	499.0
213.90	216.0
214.90	2369.0
216.00	4680.0
217.00	50608.0
218.00	5807.0
218.90	536.0
221.00	28776.0
223.00	11627.0
224.00	108976.0
225.00	27352.0
226.00	3271.0
227.00	47992.0
228.00	6280.0
228.90	8973.0
229.90	1235.0
231.00	3766.0
232.00	754.0
233.00	744.0

Instrument :
BNA_N
ClientSampleId :
DFTPP308

234.00	2964.0
234.90	2908.0
236.00	2375.0
237.00	4465.0
237.80	696.0
238.90	1448.0
240.00	1413.0
240.90	2289.0
242.00	5714.0
243.00	6351.0
244.00	80160.0
245.00	11149.0
246.00	16840.0
247.00	3482.0
247.90	729.0
249.00	3029.0
249.90	731.0
250.90	673.0
252.00	977.0
253.00	1674.0
255.00	433344.0
256.00	62000.0
257.00	5376.0
257.90	22024.0
258.90	4162.0
260.00	765.0
261.00	760.0
263.00	202.0
264.10	913.0
265.00	10261.0
265.90	1423.0
267.00	311.0
269.90	534.0
270.90	946.0
271.90	1118.0
273.00	12716.0
274.00	31832.0
275.00	186560.0
276.00	25408.0
277.00	17216.0
278.00	3103.0
278.90	598.0
281.00	804.0
282.00	618.0
283.00	1598.0
283.80	1046.0
285.00	2688.0
286.00	469.0
288.00	263.0
288.80	720.0
289.90	399.0
291.00	461.0
291.90	787.0
293.00	4035.0
294.00	1072.0
296.00	53208.0
297.00	6776.0
298.00	459.0
301.00	862.0
302.00	776.0
303.00	6243.0
304.00	1853.0
305.00	306.0
306.90	155.0
307.90	872.0
309.00	516.0
309.90	663.0
312.10	247.0
313.00	490.0
314.00	2657.0
314.90	5834.0
316.10	3387.0
317.00	653.0
319.90	252.0
320.90	1430.0
322.00	950.0
323.00	17008.0
324.00	2785.0
325.00	219.0
325.90	344.0
326.90	3602.0
327.90	1925.0
329.00	272.0
331.90	1193.0
333.00	1780.0
334.00	11188.0
335.00	2810.0
335.90	317.0
338.90	273.0
340.10	189.0
341.00	2084.0
342.10	600.0
345.90	4202.0
347.00	984.0
350.80	271.0
351.00	254.0

Instrument :
BNA_N
ClientSampleId :
DFTPP308

351.90	4707.0
353.00	3585.0
354.00	5320.0
355.00	1235.0
355.90	167.0
359.10	430.0
364.90	25728.0
365.90	3785.0
366.70	359.0
370.00	509.0
370.90	1333.0
372.00	8389.0
373.00	2290.0
373.90	287.0
377.00	290.0
383.00	2208.0
384.00	659.0
384.80	205.0
390.00	1004.0
391.00	866.0
392.00	433.0
400.90	595.0
402.00	3680.0
403.00	4843.0
403.90	1666.0
405.00	379.0
410.00	184.0
415.00	179.0
421.00	4263.0
422.00	3888.0
423.00	27032.0
424.00	7426.0
425.00	835.0
441.00	62536.0
442.00	475264.0
443.00	81848.0
444.10	9051.0
445.00	458.0

Instrument :
BNA_N
ClientSampleId :
DFTPP308

m/z	Abundance
36.00	261.0
37.00	1101.0
38.00	2776.0
39.00	17528.0
40.00	3337.0
41.00	745.0
43.10	355.0
44.00	5717.0
45.00	631.0
48.00	257.0
49.00	1978.0
50.00	59208.0
51.00	230272.0
52.00	12064.0
53.00	467.0
55.00	1599.0
56.00	6561.0
57.00	13865.0
58.00	668.0
58.90	396.0
60.00	274.0
61.00	2427.0
62.00	2983.0
63.00	9006.0
64.00	1581.0
65.00	4574.0
65.90	330.0
67.00	405.0
68.00	4071.0
69.00	236352.0
70.00	1269.0
71.00	233.0
73.00	2684.0
74.00	26256.0
75.00	38864.0
76.00	14063.0
77.00	295680.0
78.00	19680.0
79.00	18208.0
79.90	13650.0
81.00	19880.0
82.00	4999.0
83.00	4330.0
83.90	863.0
85.00	2975.0
86.00	5710.0
86.90	2367.0
88.00	1167.0
88.90	486.0
90.10	153.0
91.00	4497.0
92.00	5518.0
93.00	33864.0
94.00	2272.0
94.90	612.0
96.00	2168.0
96.90	557.0
98.00	28288.0
99.00	20536.0
99.90	1818.0
101.00	11572.0
101.90	633.0
103.00	3838.0
103.90	7411.0
105.00	6994.0
106.00	2238.0
107.00	91416.0
108.00	14004.0
109.00	2462.0
110.00	161344.0
111.00	24624.0
112.00	2901.0
113.00	960.0
113.90	352.0
114.90	529.0
116.00	4955.0
117.00	88408.0
118.00	6530.0
119.00	849.0
120.00	1186.0
120.90	416.0
122.00	6393.0
123.00	8678.0
124.00	4318.0
124.90	4119.0
127.00	318400.0
128.00	24264.0
129.00	122040.0
130.00	10753.0
131.00	2256.0
132.00	964.0
132.90	687.0

Instrument :
BNA_N
ClientSampleId :
DFTPP308

134.00	3761.0
135.00	10144.0
136.00	4560.0
137.00	6092.0
138.00	1119.0
138.80	706.0
140.00	1771.0
141.00	15871.0
142.00	5381.0
142.90	3467.0
143.90	687.0
145.00	1070.0
145.90	3060.0
147.00	8867.0
148.00	22240.0
149.00	4414.0
150.00	1383.0
151.10	2330.0
151.80	1268.0
153.00	5078.0
154.00	3946.0
155.00	9723.0
156.00	14372.0
157.00	2690.0
157.90	2909.0
159.00	2455.0
160.00	5161.0
161.00	8154.0
162.00	2340.0
163.00	953.0
164.10	856.0
165.00	6227.0
166.00	4956.0
167.00	37624.0
168.00	22872.0
169.00	3016.0
170.00	1186.0
170.90	1386.0
172.00	2938.0
173.00	4094.0
174.00	7445.0
175.00	14309.0
176.00	3319.0
177.00	5910.0
178.00	2200.0
179.00	26392.0
180.00	18696.0
181.00	8709.0
181.90	1294.0
183.00	589.0
184.00	2236.0
185.00	13023.0
186.00	96448.0
187.00	27352.0
188.00	2887.0
188.90	6494.0
190.00	1135.0
191.00	3153.0
192.00	8278.0
192.90	9182.0
194.00	2016.0
195.00	1080.0
196.00	20288.0
197.90	719040.0
199.00	47416.0
200.00	3533.0
201.40	2912.0
203.00	4691.0
204.00	26560.0
205.10	43016.0
206.00	190272.0
207.00	26208.0
208.00	7316.0
209.00	2310.0
211.00	7350.0
212.90	538.0
214.00	280.0
214.90	2322.0
216.00	4026.0
217.00	49376.0
218.00	6058.0
219.00	791.0
221.00	29328.0
223.00	11186.0
224.00	107352.0
225.00	27560.0
226.00	3251.0
227.00	44008.0
228.00	6093.0
228.90	9354.0
230.00	1161.0
231.00	3512.0
232.00	624.0
233.00	881.0
234.00	2985.0
235.00	3157.0

Instrument :
BNA_N
ClientSampleId :
DFTPP308

235.90	2402.0
237.10	4159.0
238.00	641.0
239.00	1853.0
239.90	1389.0
240.90	2375.0
242.00	6053.0
243.10	6450.0
244.00	81432.0
245.00	11469.0
246.00	16105.0
246.90	3533.0
247.90	819.0
248.90	2837.0
249.90	517.0
250.90	638.0
252.00	718.0
253.00	1810.0
255.00	419648.0
256.00	61032.0
257.00	4750.0
258.00	23736.0
258.90	3999.0
259.90	785.0
260.90	805.0
262.00	164.0
262.90	202.0
263.90	647.0
265.00	10482.0
265.90	1516.0
266.90	226.0
267.80	173.0
269.90	475.0
270.90	1006.0
272.00	885.0
273.00	12707.0
274.00	34976.0
275.00	182592.0
276.00	26264.0
277.00	16013.0
278.00	2931.0
279.00	613.0
281.00	716.0
282.10	613.0
283.00	2046.0
284.00	1184.0
285.00	2816.0
286.00	696.0
289.00	691.0
289.90	567.0
290.90	539.0
292.10	788.0
292.90	3421.0
294.00	1080.0
296.00	57936.0
297.00	7753.0
298.00	623.0
298.90	167.0
301.00	676.0
302.00	1171.0
303.00	6135.0
304.00	1887.0
305.00	431.0
307.90	620.0
309.00	793.0
310.00	757.0
311.00	238.0
311.90	234.0
313.00	594.0
314.00	2689.0
314.90	6103.0
316.00	3516.0
317.00	684.0
319.90	278.0
321.00	1732.0
322.00	855.0
323.00	18464.0
324.00	3110.0
325.10	483.0
325.90	326.0
326.90	3811.0
327.90	1932.0
328.90	269.0
332.00	1395.0
332.90	1579.0
334.00	12926.0
335.00	2893.0
336.00	360.0
338.80	329.0
339.90	277.0
341.00	2127.0
342.10	535.0
346.00	4157.0
347.00	818.0
348.00	168.0
349.90	179.0

Instrument :
BNA_N
ClientSampleId :
DFTPP308

350.90	341.0
352.00	5230.0
352.90	3421.0
354.10	6104.0
355.00	1122.0
358.80	325.0
364.90	27912.0
366.00	4063.0
366.90	320.0
369.90	557.0
371.00	1216.0
372.00	9064.0
373.00	2244.0
374.00	287.0
376.90	221.0
382.90	2699.0
384.00	920.0
384.90	250.0
390.00	1643.0
391.00	703.0
392.00	836.0
400.90	636.0
402.00	4523.0
403.00	6229.0
404.00	1736.0
404.90	382.0
415.00	186.0
421.00	5317.0
422.00	3792.0
423.00	36320.0
424.00	7568.0
425.00	1024.0
439.30	226.0
441.00	88448.0
442.00	511936.0
443.00	97656.0
444.10	9475.0
445.00	544.0

Instrument :

BNA_N

ClientSampleId :

DFTPP308