

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032922\
 Data File : BN019221.D
 Acq On : 29 Mar 2022 12:45
 Operator : CG/JU
 Sample : PB143655BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB143655BL

Quant Time: Mar 29 17:28:19 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 25 19:03:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.840	152	4903	0.400	ng	0.00
7) Naphthalene-d8	10.647	136	12230	0.400	ng	# 0.00
13) Acenaphthene-d10	14.484	164	6556	0.400	ng	0.00
19) Phenanthrene-d10	17.225	188	10982	0.400	ng	0.00
29) Chrysene-d12	21.422	240	6613	0.400	ng	# 0.00
35) Perylene-d12	23.787	264	5780	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.399	112	3995	0.345	ng	0.00
5) Phenol-d6	7.002	99	3667	0.277	ng	0.00
8) Nitrobenzene-d5	9.003	82	2923	0.308	ng	0.01
11) 2-Methylnaphthalene-d10	12.238	152	6613	0.338	ng	0.00
14) 2,4,6-Tribromophenol	15.984	330	585	0.184	ng	0.01
15) 2-Fluorobiphenyl	13.116	172	8701	0.316	ng	0.01
27) Fluoranthene-d10	19.261	212	10703	0.333	ng	0.00
31) Terphenyl-d14	19.855	244	5604	0.380	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032922\
Data File : BN019221.D
Acq On : 29 Mar 2022 12:45
Operator : CG/JU
Sample : PB143655BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB143655BL

Quant Time: Mar 29 17:28:19 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 25 19:03:12 2022
Response via : Initial Calibration

