

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032922\
 Data File : BN019231.D
 Acq On : 29 Mar 2022 18:43
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 29 19:14:25 2022
 Quant Title :
 QLast Update : Fri Mar 25 19:03:12 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2	1,4-Dioxane	0.533	0.553	-3.8	102	0.00
3	n-Nitrosodimethylamine	0.595	0.545	8.4	91	0.00
4 S	2-Fluorophenol	0.945	0.935	1.1	95	0.00
5 S	Phenol-d6	1.082	0.918	15.2	88	0.00
6	bis(2-Chloroethyl)ether	1.168	1.097	6.1	92	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
8 S	Nitrobenzene-d5	0.310	0.281	9.4	92	0.01
9	Naphthalene	1.126	1.177	-4.5	98	0.00
10	Hexachlorobutadiene	0.253	0.290	-14.6	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.640	0.621	3.0	92	0.00
12	2-Methylnaphthalene	0.740	0.734	0.8	93	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	80	-0.01
14 S	2,4,6-Tribromophenol	0.194	0.160	17.5	74	0.00
15 S	2-Fluorobiphenyl	1.681	1.686	-0.3	88	0.00
16	Acenaphthylene	1.826	1.758	3.7	87	0.00
17	Acenaphthene	1.294	1.289	0.4	86	-0.01
18	Fluorene	1.646	1.539	6.5	81	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	67	0.00
20	4,6-Dinitro-2-methylphenol	0.065	0.048	26.2#	57	0.01
21	4-Bromophenyl-phenylether	0.260	0.270	-3.8	77	0.00
22	Hexachlorobenzene	0.307	0.362	-17.9	85	0.00
23	Atrazine	0.180	0.158	12.2	66	0.00
24	Pentachlorophenol	0.110	0.101	8.2	72	0.00
25	Phenanthrene	1.283	1.253	2.3	72	0.00
26	Anthracene	1.064	0.971	8.7	70	0.00
27 SURR	Fluoranthene-d10	1.170	1.136	2.9	72	0.00
28	Fluoranthene	1.450	1.413	2.6	72	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	61	0.00
30	Pyrene	1.890	2.130	-12.7	72	0.00
31 S	Terphenyl-d14	0.893	0.920	-3.0	68	0.00
32	Benzo(a)anthracene	1.285	1.234	4.0	69	0.00
33	Chrysene	1.647	1.693	-2.8	66	0.00
34	Bis(2-ethylhexyl)phthalate	0.783	0.823	-5.1	71	0.00
35 I	Perylene-d12	1.000	1.000	0.0	63	0.00
36	Indeno(1,2,3-cd)pyrene	1.599	1.539	3.8	69	0.00
37	Benzo(b)fluoranthene	1.519	1.480	2.6	72	0.00
38	Benzo(k)fluoranthene	1.928	1.851	4.0	65	0.00
39 C	Benzo(a)pyrene	1.302	1.351	-3.8	74	0.00
40	Dibenzo(a,h)anthracene	1.204	1.214	-0.8	77	0.00
41	Benzo(g,h,i)perylene	1.646	1.677	-1.9	69	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0