

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032922\
 Data File : BN019231.D
 Acq On : 29 Mar 2022 18:43
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 29 19:14:25 2022
 Quant Title :
 QLast Update : Fri Mar 25 19:03:12 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	92	0.00
2	1,4-Dioxane	0.400	0.416	-4.0	102	0.00
3	n-Nitrosodimethylamine	0.400	0.366	8.5	91	0.00
4 S	2-Fluorophenol	0.400	0.396	1.0	95	0.00
5 S	Phenol-d6	0.400	0.340	15.0	88	0.00
6	bis(2-Chloroethyl)ether	0.400	0.376	6.0	92	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	85	0.00
8 S	Nitrobenzene-d5	0.400	0.362	9.5	92	0.01
9	Naphthalene	0.400	0.418	-4.5	98	0.00
10	Hexachlorobutadiene	0.400	0.459	-14.8	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.388	3.0	92	0.00
12	2-Methylnaphthalene	0.400	0.397	0.8	93	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	80	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.330	17.5	74	0.00
15 S	2-Fluorobiphenyl	0.400	0.401	-0.3	88	0.00
16	Acenaphthylene	0.400	0.385	3.8	87	0.00
17	Acenaphthene	0.400	0.399	0.3	86	-0.01
18	Fluorene	0.400	0.374	6.5	81	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	67	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.294	26.5#	57	0.01
21	4-Bromophenyl-phenylether	0.400	0.416	-4.0	77	0.00
22	Hexachlorobenzene	0.400	0.471	-17.7	85	0.00
23	Atrazine	0.400	0.352	12.0	66	0.00
24	Pentachlorophenol	0.400	0.366	8.5	72	0.00
25	Phenanthrene	0.400	0.391	2.3	72	0.00
26	Anthracene	0.400	0.365	8.8	70	0.00
27 SURR	Fluoranthene-d10	0.400	0.388	3.0	72	0.00
28	Fluoranthene	0.400	0.390	2.5	72	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	61	0.00
30	Pyrene	0.400	0.451	-12.7	72	0.00
31 S	Terphenyl-d14	0.400	0.412	-3.0	68	0.00
32	Benzo(a)anthracene	0.400	0.384	4.0	69	0.00
33	Chrysene	0.400	0.411	-2.7	66	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.421	-5.2	71	0.00
35 I	Perylene-d12	0.400	0.400	0.0	63	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.385	3.8	69	0.00
37	Benzo(b)fluoranthene	0.400	0.390	2.5	72	0.00
38	Benzo(k)fluoranthene	0.400	0.384	4.0	65	0.00
39 C	Benzo(a)pyrene	0.400	0.415	-3.7	74	0.00
40	Dibenzo(a,h)anthracene	0.400	0.403	-0.8	77	0.00
41	Benzo(g,h,i)perylene	0.400	0.408	-2.0	69	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0