



4,6-Dinitro-2-methylphenol



Response = $7.940 \times 10^{-2} \times \text{Amt} - 2.720 \times 10^{-2}$
 Coef of Det (r^2) = 0.992093 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN033022.M
 Calibration Table Last Updated: Wed Mar 30 12:57:13 2022

Benzo (k) fluoranthene

Response Ratio



$$\text{Response} = 1.984\text{e}+000 * \text{Amt} + 7.021\text{e}-001$$


Coef of Det (r^2) = 0.999628 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN033022.M

Calibration Table Last Updated: Wed Mar 30 12:57:13 2022

Benzo (b) fluoranthene

Response Ratio



$$\text{Response} = 1.721\text{e}+000 * \text{Amt} + 2.770\text{e}-001$$


Coef of Det (r^2) = 0.998247 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN033022.M

Calibration Table Last Updated: Wed Mar 30 12:57:13 2022

Bis(2-ethylhexyl)phthalate

Response Ratio



$$\text{Response} = 9.044\text{e-}001 * \text{Amt} + 2.434\text{e-}001$$

Coef of Det (r^2) = 0.994283 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN033022.M

Calibration Table Last Updated: Wed Mar 30 12:57:13 2022