

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033022\
 Data File : BN019244.D
 Acq On : 30 Mar 2022 17:28
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 30 18:39:08 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN033022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 30 12:57:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	85	0.00
2	1,4-Dioxane	0.400	0.431	-7.7	86	0.00
3	n-Nitrosodimethylamine	0.400	0.392	2.0	81	0.00
4 S	2-Fluorophenol	0.400	0.377	5.8	76	0.00
5 S	Phenol-d6	0.400	0.341	14.8	73	0.00
6	bis(2-Chloroethyl)ether	0.400	0.378	5.5	75	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	84	0.00
8 S	Nitrobenzene-d5	0.400	0.339	15.3	76	0.00
9	Naphthalene	0.400	0.388	3.0	80	0.00
10	Hexachlorobutadiene	0.400	0.401	-0.3	82	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.385	3.8	80	0.00
12	2-Methylnaphthalene	0.400	0.379	5.3	79	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	82	0.00
14 S	2,4,6-Tribromophenol	0.400	0.315	21.3	69	0.00
15 S	2-Fluorobiphenyl	0.400	0.382	4.5	79	0.00
16	Acenaphthylene	0.400	0.354	11.5	76	0.00
17	Acenaphthene	0.400	0.378	5.5	78	0.00
18	Fluorene	0.400	0.380	5.0	78	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	81	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.350	12.5	66	0.01
21	4-Bromophenyl-phenylether	0.400	0.369	7.8	76	0.00
22	Hexachlorobenzene	0.400	0.399	0.3	79	0.00
23	Atrazine	0.400	0.319	20.3	67	0.00
24	Pentachlorophenol	0.400	0.303	24.3	68	0.00
25	Phenanthrene	0.400	0.375	6.3	76	0.00
26	Anthracene	0.400	0.347	13.3	74	0.00
27 SURR	Fluoranthene-d10	0.400	0.372	7.0	74	0.00
28	Fluoranthene	0.400	0.370	7.5	74	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	74	0.00
30	Pyrene	0.400	0.413	-3.2	75	0.00
31 S	Terphenyl-d14	0.400	0.431	-7.7	78	0.00
32	Benzo(a)anthracene	0.400	0.338	15.5	68	0.00
33	Chrysene	0.400	0.349	12.8	66	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.266	33.5#	57	0.00
35 I	Perylene-d12	0.400	0.400	0.0	73	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.406	-1.5	82	0.01
37	Benzo(b)fluoranthene	0.400	0.411	-2.7	76	0.00
38	Benzo(k)fluoranthene	0.400	0.381	4.8	75	0.00
39 C	Benzo(a)pyrene	0.400	0.365	8.8	72	0.00
40	Dibenzo(a,h)anthracene	0.400	0.378	5.5	75	0.01
41	Benzo(g,h,i)perylene	0.400	0.402	-0.5	78	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4