

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
 Data File : BN024651.D
 Acq On : 30 Mar 2023 20:12
 Operator : CG/JU
 Sample : 02038-05
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EYF65

Quant Time: Mar 31 02:13:09 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 31 02:10:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.000	152	8717	0.400	ng/u1	0.00
4) Naphthalene-d8	10.812	136	26558	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.638	164	15541	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.384	188	32507	0.400	ng/u1	0.00
17) Chrysene-d12	21.574	240	24369	0.400	ng/u1	0.00
23) Perylene-d12	24.047	264	25351	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.372	96	51046	5.705	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.401	152	15235	0.465	ng/u1	0.00
18) Fluoranthene-d10	19.408	212	33407	0.560	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
Data File : BN024651.D
Acq On : 30 Mar 2023 20:12
Operator : CG/JU
Sample : 02038-05
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EYF65

Quant Time: Mar 31 02:13:09 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 31 02:10:21 2023
Response via : Initial Calibration

