

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
 Data File : BN024664.D
 Acq On : 31 Mar 2023 04:39
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4254

Quant Time: Mar 31 05:09:13 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 31 02:10:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.000	152	11848	0.400	ng/u1	0.00
4) Naphthalene-d8	10.812	136	35391	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.638	164	19695	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.384	188	41370	0.400	ng/u1	0.00
17) Chrysene-d12	21.568	240	33583	0.400	ng/u1	0.00
23) Perylene-d12	24.038	264	26830	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.376	96	4864	0.400	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.396	152	17300	0.396	ng/u1	0.00
18) Fluoranthene-d10	19.408	212	33741	0.411	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.410	88	5080	0.397	ng/u1	96
5) Naphthalene	10.861	128	34088	0.392	ng/u1	100
7) 2-Methylnaphthalene	12.467	142	21282	0.393	ng/u1	100
8) 1-Methylnaphthalene	12.687	142	22321	0.394	ng/u1	100
10) Acenaphthylene	14.356	152	27479	0.395	ng/u1	100
11) Acenaphthene	14.698	153	23652	0.392	ng/u1	99
12) Fluorene	15.684	166	26445	0.383	ng/u1	99
14) Pentachlorophenol	17.029	266	3276	0.366	ng/u1	99
15) Phenanthrene	17.422	178	43693	0.385	ng/u1	99
16) Anthracene	17.515	178	36712	0.387	ng/u1	100
19) Fluoranthene	19.436	202	47190	0.399	ng/u1	98
20) Pyrene	19.799	202	47827	0.397	ng/u1	98
21) Benzo(a)anthracene	21.554	228	40585	0.402	ng/u1	99
22) Chrysene	21.606	228	42538	0.383	ng/u1	99
24) Benzo(b)fluoranthene	23.281	252	41122	0.369	ng/u1	99
25) Benzo(k)fluoranthene	23.328	252	39962	0.373	ng/u1	99
26) Benzo(a)pyrene	23.927	252	33692	0.384	ng/u1	99
27) Indeno(1,2,3-cd)pyrene	26.609	276	38530	0.369	ng/u1#	100
28) Dibenzo(a,h)anthracene	26.629	278	30352	0.363	ng/u1	99
29) Benzo(g,h,i)perylene	27.401	276	32301	0.360	ng/u1	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
Data File : BN024664.D
Acq On : 31 Mar 2023 04:39
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4254

Quant Time: Mar 31 05:09:13 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 31 02:10:21 2023
Response via : Initial Calibration

