

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
 Data File : BN024678.D
 Acq On : 31 Mar 2023 14:27
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4255

Quant Time: Apr 01 01:01:41 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Apr 01 01:00:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.000	152	11634	0.400	ng/ul	0.00
4) Naphthalene-d8	10.812	136	35348	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.638	164	19899	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.383	188	41732	0.400	ng/ul	0.00
17) Chrysene-d12	21.571	240	34596	0.400	ng/ul	0.00
23) Perylene-d12	24.058	264	27953	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.372	96	4491	0.376	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.396	152	17341	0.398	ng/ul	0.00
18) Fluoranthene-d10	19.408	212	34497	0.407	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.410	88	4783	0.381	ng/ul	98
5) Naphthalene	10.861	128	34189	0.394	ng/ul	100
7) 2-Methylnaphthalene	12.467	142	21420	0.396	ng/ul	100
8) 1-Methylnaphthalene	12.687	142	22451	0.397	ng/ul	100
10) Acenaphthylene	14.356	152	27995	0.398	ng/ul	100
11) Acenaphthene	14.698	153	23816	0.390	ng/ul	99
12) Fluorene	15.684	166	26685	0.383	ng/ul	99
14) Pentachlorophenol	17.033	266	3116	0.345	ng/ul	98
15) Phenanthrene	17.426	178	44077	0.385	ng/ul	100
16) Anthracene	17.514	178	37466	0.392	ng/ul	100
19) Fluoranthene	19.436	202	48417	0.397	ng/ul	98
20) Pyrene	19.798	202	49064	0.395	ng/ul	98
21) Benzo(a)anthracene	21.553	228	41843	0.403	ng/ul	100
22) Chrysene	21.609	228	44045	0.385	ng/ul	100
24) Benzo(b)fluoranthene	23.293	252	43440	0.374	ng/ul	99
25) Benzo(k)fluoranthene	23.348	252	40735	0.365	ng/ul	99
26) Benzo(a)pyrene	23.942	252	35937	0.393	ng/ul	97
27) Indeno(1,2,3-cd)pyrene	26.646	276	40120	0.369	ng/ul#	100
28) Dibenzo(a,h)anthracene	26.669	278	31381	0.361	ng/ul	99
29) Benzo(g,h,i)perylene	27.451	276	32844	0.351	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
Data File : BN024678.D
Acq On : 31 Mar 2023 14:27
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4255

Quant Time: Apr 01 01:01:41 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Apr 01 01:00:20 2023
Response via : Initial Calibration

