

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
 Data File : BN024679.D
 Acq On : 31 Mar 2023 15:03
 Operator : CG/JU
 Sample : PB151732BL
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK732

Quant Time: Apr 01 01:01:52 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Apr 01 01:00:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.000	152	9361	0.400	ng/u1	0.00
4) Naphthalene-d8	10.812	136	28283	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.638	164	16165	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.384	188	33721	0.400	ng/u1	0.00
17) Chrysene-d12	21.571	240	25099	0.400	ng/u1	0.00
23) Perylene-d12	24.038	264	24677	0.400	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.368	96	61827	6.435	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.396	152	13210	0.378	ng/u1	0.00
18) Fluoranthene-d10	19.408	212	27053	0.440	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
Data File : BN024679.D
Acq On : 31 Mar 2023 15:03
Operator : CG/JU
Sample : PB151732BL
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK732

Quant Time: Apr 01 01:01:52 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Apr 01 01:00:20 2023
Response via : Initial Calibration

