

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
 Data File : BN024709.D
 Acq On : 01 Apr 2023 09:56
 Operator : CG/JU
 Sample : PB151702BS
 Misc :
 ALS Vial : 64 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS702

Quant Time: Apr 03 01:36:29 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Apr 01 04:05:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.996	152	8695	0.400	ng/ul	0.00
4) Naphthalene-d8	10.812	136	25627	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.633	164	13726	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.380	188	29469	0.400	ng/ul	0.00
17) Chrysene-d12	21.565	240	22151	0.400	ng/ul	0.00
23) Perylene-d12	24.033	264	17471	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.372	96	7026	0.787	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.396	152	13665	0.432	ng/ul	0.00
18) Fluoranthene-d10	19.404	212	26449	0.488	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.406	88	18645	1.986	ng/ul	94
5) Naphthalene	10.861	128	28650	0.455	ng/ul	100
7) 2-Methylnaphthalene	12.467	142	17723	0.452	ng/ul	100
8) 1-Methylnaphthalene	12.687	142	18639	0.455	ng/ul	99
10) Acenaphthylene	14.356	152	22237	0.458	ng/ul	100
11) Acenaphthene	14.698	153	19359	0.460	ng/ul	99
12) Fluorene	15.679	166	21659	0.451	ng/ul	99
14) Pentachlorophenol	17.029	266	4288	0.673	ng/ul	99
15) Phenanthrene	17.422	178	35537	0.440	ng/ul	100
16) Anthracene	17.515	178	29340	0.435	ng/ul	100
19) Fluoranthene	19.436	202	38484	0.493	ng/ul	99
20) Pyrene	19.799	202	38694	0.487	ng/ul	98
21) Benzo(a)anthracene	21.548	228	32414	0.487	ng/ul	99
22) Chrysene	21.600	228	34907	0.477	ng/ul	99
24) Benzo(b)fluoranthene	23.275	252	32393	0.446	ng/ul	100
25) Benzo(k)fluoranthene	23.322	252	30797	0.441	ng/ul	99
26) Benzo(a)pyrene	23.919	252	25847	0.453	ng/ul	98
27) Indeno(1,2,3-cd)pyrene	26.603	276	30275	0.445	ng/ul#	96
28) Dibenzo(a,h)anthracene	26.619	278	23727	0.436	ng/ul	99
29) Benzo(g,h,i)perylene	27.394	276	25447	0.435	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
Data File : BN024709.D
Acq On : 01 Apr 2023 09:56
Operator : CG/JU
Sample : PB151702BS
Misc :
ALS Vial : 64 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS702

Quant Time: Apr 03 01:36:29 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Apr 01 04:05:52 2023
Response via : Initial Calibration

