

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
 Data File : BN024659.D
 Acq On : 31 Mar 2023 01:02
 Operator : CG/JU
 Sample : 02067-02
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BFZ79

Quant Time: Mar 31 02:14:11 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 31 02:10:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.000	152	9033	0.400	ng/ul	0.00
4) Naphthalene-d8	10.812	136	27320	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.638	164	16061	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.384	188	33725	0.400	ng/ul	0.00
17) Chrysene-d12	21.574	240	25916	0.400	ng/ul	0.00
23) Perylene-d12	24.062	264	23329	0.400	ng/ul	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.372	96	53495	5.770	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.396	152	16066	0.477	ng/ul	0.00
18) Fluoranthene-d10	19.408	212	40202	0.634	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.410	88	1994	0.204	ng/ul#	81


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
Data File : BN024659.D
Acq On : 31 Mar 2023 01:02
Operator : CG/JU
Sample : 02067-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BFZ79


Quant Time: Mar 31 02:14:11 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 31 02:10:21 2023
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.204 ng/ul
 RT: 3.410 min Scan# 30
 Delta R.T. -0.000 min
 Lab File: BN024659.D
 Acq: 31 Mar 2023 01:02

Instrument :
 BNA_N
 ClientSampleId :
 BFZ79



Tgt Ion:	88	Resp:	1994
Ion	Ratio	Lower	Upper
88	100		
43	73.2	37.2	55.8#
58	71.6	61.5	92.3

