

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
 Data File : BN024665.D
 Acq On : 31 Mar 2023 05:15
 Operator : CG/JU
 Sample : PB151691BL
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK691

Quant Time: Mar 31 06:09:40 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 31 02:10:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.000	152	9956	0.400	ng/u1	0.00
4) Naphthalene-d8	10.812	136	30293	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.638	164	17416	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.384	188	35822	0.400	ng/u1	0.00
17) Chrysene-d12	21.571	240	26031	0.400	ng/u1	0.00
23) Perylene-d12	24.059	264	21784	0.400	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.372	96	65289	6.389	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.396	152	14232	0.381	ng/u1	0.00
18) Fluoranthene-d10	19.408	212	28047	0.440	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
Data File : BN024665.D
Acq On : 31 Mar 2023 05:15
Operator : CG/JU
Sample : PB151691BL
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK691

Quant Time: Mar 31 06:09:40 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 31 02:10:21 2023
Response via : Initial Calibration

