

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
 Data File : BN024668.D
 Acq On : 31 Mar 2023 07:03
 Operator : CG/JU
 Sample : 02041-02
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 CB956

Quant Time: Mar 31 07:38:13 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 31 02:10:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.000	152	8398	0.400	ng/u1	0.00
4) Naphthalene-d8	10.812	136	25584	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.638	164	15242	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.384	188	33072	0.400	ng/u1	0.00
17) Chrysene-d12	21.574	240	26507	0.400	ng/u1	0.00
23) Perylene-d12	24.059	264	24700	0.400	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.372	96	44510	5.164	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.396	152	14391	0.456	ng/u1	0.00
18) Fluoranthene-d10	19.408	212	41459	0.639	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
Data File : BN024668.D
Acq On : 31 Mar 2023 07:03
Operator : CG/JU
Sample : 02041-02
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
CB956

Quant Time: Mar 31 07:38:13 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 31 02:10:21 2023
Response via : Initial Calibration

