

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
 Data File : BN024672.D
 Acq On : 31 Mar 2023 10:11
 Operator : CG/JU
 Sample : 02041-02
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 CB956

Quant Time: Apr 01 00:45:05 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 31 02:10:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.000	152	9341	0.400	ng/u1	0.00
4) Naphthalene-d8	10.812	136	28310	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.638	164	16400	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.379	188	35368	0.400	ng/u1	0.00
17) Chrysene-d12	21.565	240	26835	0.400	ng/u1	0.00
23) Perylene-d12	24.035	264	25692	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.372	96	48735	5.083	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.396	152	11836	0.339	ng/u1	0.00
18) Fluoranthene-d10	19.408	212	32160	0.490	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
Data File : BN024672.D
Acq On : 31 Mar 2023 10:11
Operator : CG/JU
Sample : 02041-02
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
CB956

Quant Time: Apr 01 00:45:05 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 31 02:10:21 2023
Response via : Initial Calibration

