

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
 Data File : BN024682.D
 Acq On : 31 Mar 2023 16:53
 Operator : CG/JU
 Sample : 02086-06
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BFZ81

Quant Time: Apr 01 01:02:11 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Apr 01 01:00:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.000	152	8205	0.400	ng/ul	0.00
4) Naphthalene-d8	10.812	136	25107	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.638	164	14509	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.384	188	30657	0.400	ng/ul	0.00
17) Chrysene-d12	21.571	240	24210	0.400	ng/ul	0.00
23) Perylene-d12	24.053	264	23341	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.368	96	39465	4.686	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.396	152	11209	0.362	ng/ul	0.00
18) Fluoranthene-d10	19.408	212	27281	0.460	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	9644	1.088	ng/ul	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033023\
Data File : BN024682.D
Acq On : 31 Mar 2023 16:53
Operator : CG/JU
Sample : 02086-06
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BFZ81


Quant Time: Apr 01 01:02:11 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN033023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Apr 01 01:00:20 2023
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 1.088 ng/ul
 RT: 3.405 min Scan# 29
 Delta R.T. -0.004 min
 Lab File: BN024682.D
 Acq: 31 Mar 2023 16:53

Instrument :
 BNA_N
 ClientSampleId :
 BFZ81



Tgt Ion:	88	Resp:	9644
Ion	Ratio	Lower	Upper
88	100		
43	47.1	37.2	55.8
58	82.5	61.5	92.3

