

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033024\
 Data File : BN030532.D
 Acq On : 29 Mar 2024 15:57
 Operator : MA/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Mar 29 23:48:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN033024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 29 23:46:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.704	152	5675	0.400	ng	0.00
7) Naphthalene-d8	10.485	136	16731	0.400	ng	0.00
13) Acenaphthene-d10	14.337	164	9041	0.400	ng	0.00
19) Phenanthrene-d10	17.090	188	20666	0.400	ng	0.00
29) Chrysene-d12	21.285	240	18481	0.400	ng	0.00
35) Perylene-d12	23.527	264	18721	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.291	112	1171	0.095	ng	0.00
5) Phenol-d6	6.873	99	1447	0.091	ng	0.00
8) Nitrobenzene-d5	8.851	82	1104	0.094	ng	0.00
11) 2-Methylnaphthalene-d10	12.085	152	2422	0.095	ng	0.00
14) 2,4,6-Tribromophenol	15.836	330	296	0.087	ng	0.00
15) 2-Fluorobiphenyl	12.969	172	3712	0.106	ng	0.00
27) Fluoranthene-d10	19.126	212	5074	0.091	ng	0.00
31) Terphenyl-d14	19.734	244	4336	0.102	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.240	88	740	0.108	ng	97
3) n-Nitrosodimethylamine	3.551	42	602	0.091	ng	# 95
6) bis(2-Chloroethyl)ether	7.140	93	1493	0.095	ng	100
9) Naphthalene	10.538	128	4508	0.097	ng	96
10) Hexachlorobutadiene	10.827	225	938	0.099	ng	# 98
12) 2-Methylnaphthalene	12.156	142	2908	0.096	ng	96
16) Acenaphthylene	14.060	152	3686	0.090	ng	99
17) Acenaphthene	14.402	154	2679	0.094	ng	99
18) Fluorene	15.396	166	3430	0.093	ng	100
21) 4-Bromophenyl-phenylether	16.295	248	1191	0.095	ng	# 69
22) Hexachlorobenzene	16.395	284	1466	0.098	ng	99
23) Atrazine	16.556	200	819	0.090	ng	# 90
25) Phenanthrene	17.127	178	5785	0.094	ng	99
26) Anthracene	17.226	178	4546	0.088	ng	99
28) Fluoranthene	19.154	202	6728	0.092	ng	99
30) Pyrene	19.521	202	7008	0.097	ng	99
32) Benzo(a)anthracene	21.267	228	6296	0.096	ng	98
33) Chrysene	21.321	228	7284	0.104	ng	100
34) Bis(2-ethylhexyl)phtha...	21.213	149	2628	0.093	ng	99
36) Indeno(1,2,3-cd)pyrene	25.796	276	7967	0.096	ng	100
37) Benzo(b)fluoranthene	22.852	252	7168	0.094	ng	# 82
38) Benzo(k)fluoranthene	22.898	252	7159	0.094	ng	# 82
39) Benzo(a)pyrene	23.428	252	5770	0.095	ng	# 76
40) Dibenzo(a,h)anthracene	25.816	278	6209	0.094	ng	# 89
41) Benzo(g,h,i)perylene	26.489	276	6920	0.098	ng	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033024\
Data File : BN030532.D
Acq On : 29 Mar 2024 15:57
Operator : MA/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: Mar 29 23:48:35 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN033024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 29 23:46:26 2024
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.704 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN030532.D
 Acq: 29 Mar 2024 15:57




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:152 Resp: 5675
 Ion Ratio Lower Upper
 152 100
 150 156.7 124.2 186.4
 115 61.6 48.6 72.8



#2
 1,4-Dioxane
 Concen: 0.108 ng
 RT: 3.240 min Scan# 36
 Delta R.T. -0.000 min
 Lab File: BN030532.D
 Acq: 29 Mar 2024 15:57

Tgt Ion: 88 Resp: 740
 Ion Ratio Lower Upper
 88 100
 43 29.9 21.9 32.9
 58 64.5 53.4 80.2









#3
n-Nitrosodimethylamine
Concen: 0.091 ng
RT: 3.551 min Scan# 79
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 42 Resp: 602
Ion Ratio Lower Upper
42 100
74 129.9 100.9 151.3
44 18.1 5.8 8.6#



#4
2-Fluorophenol
Concen: 0.095 ng
RT: 5.291 min Scan# 320
Delta R.T. -0.007 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57

Tgt Ion: 112 Resp: 1171
Ion Ratio Lower Upper
112 100
64 53.5 43.2 64.8
63 27.8 22.5 33.7









#5
Phenol-d6
Concen: 0.091 ng
RT: 6.873 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 99 Resp: 1447
Ion Ratio Lower Upper
99 100
42 22.5 15.8 23.6
71 32.6 25.9 38.9



#6
bis(2-Chloroethyl)ether
Concen: 0.095 ng
RT: 7.140 min Scan# 576
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57

Tgt Ion: 93 Resp: 1493
Ion Ratio Lower Upper
93 100
63 75.4 60.5 90.7
95 34.9 27.5 41.3









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.485 min Scan# 942
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	136	Resp:	16731
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.1	13.7
54	8.2	6.4	9.6
68	6.9	5.4	8.0



#8
Nitrobenzene-d5
Concen: 0.094 ng
RT: 8.851 min Scan# 789
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57

Tgt Ion:	82	Resp:	1104
Ion	Ratio	Lower	Upper
82	100		
128	46.3	32.3	48.5
54	56.6	42.7	64.1









#9
Naphthalene
Concen: 0.097 ng
RT: 10.538 min Scan# 947
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:128 Resp: 4508
Ion Ratio Lower Upper
128 100
129 13.0 9.0 13.6
127 14.5 10.5 15.7



#10
Hexachlorobutadiene
Concen: 0.099 ng
RT: 10.827 min Scan# 974
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57

Tgt Ion:225 Resp: 938
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 66.1 51.4 77.2









#11
2-Methylnaphthalene-d10
Concen: 0.095 ng
RT: 12.085 min Scan# 1189
Delta R.T. 0.004 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 2422
Ion Ratio Lower Upper
152 100
151 20.9 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.096 ng
RT: 12.156 min Scan# 1209
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57

Tgt Ion:142 Resp: 2908
Ion Ratio Lower Upper
142 100
141 89.6 69.8 104.8
115 39.7 28.3 42.5









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.337 min Scan# 1541
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:164 Resp: 9041
Ion Ratio Lower Upper
164 100
162 103.8 83.5 125.3
160 52.2 42.5 63.7



#14
2,4,6-Tribromophenol
Concen: 0.087 ng
RT: 15.836 min Scan# 1677
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57

Tgt Ion:330 Resp: 296
Ion Ratio Lower Upper
330 100
332 95.3 77.0 115.6
141 43.9 35.9 53.9









#15
2-Fluorobiphenyl
Concen: 0.106 ng
RT: 12.969 min Scan# 1413
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:172 Resp: 3712
Ion Ratio Lower Upper
172 100
171 35.3 27.1 40.7
170 24.1 18.3 27.5



#16
Acenaphthylene
Concen: 0.090 ng
RT: 14.060 min Scan# 1515
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57

Tgt Ion:152 Resp: 3686
Ion Ratio Lower Upper
152 100
151 19.2 15.6 23.4
153 13.0 10.3 15.5





#17

Acenaphthene

Concen: 0.094 ng

RT: 14.402 min Scan# 1547

Delta R.T. -0.000 min

Lab File: BN030532.D

Acq: 29 Mar 2024 15:57

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:154 Resp: 2679

Ion Ratio Lower Upper

154 100

153 114.9 91.0 136.6

152 58.8 47.8 71.6



#18

Fluorene

Concen: 0.093 ng

RT: 15.396 min Scan# 1640

Delta R.T. -0.000 min

Lab File: BN030532.D

Acq: 29 Mar 2024 15:57




Tgt Ion:166 Resp: 3430


Ion Ratio Lower Upper

166 100

165 97.6 77.9 116.9

167 13.2 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.090 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:188 Resp: 20666
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 9.8 7.4 11.0



#21
4-Bromophenyl-phenylether
Concen: 0.095 ng
RT: 16.295 min Scan# 1714
Delta R.T. 0.012 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57

Tgt Ion:248 Resp: 1191
Ion Ratio Lower Upper
248 100
250 104.7 72.2 108.2
141 45.5 71.1 106.7#









#22
Hexachlorobenzene
Concen: 0.098 ng
RT: 16.395 min Scan# 1722
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:284 Resp: 1466
Ion Ratio Lower Upper
284 100
142 39.3 31.8 47.6
249 33.4 25.8 38.6



#23
Atrazine
Concen: 0.090 ng
RT: 16.556 min Scan# 1735
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57

Tgt Ion:200 Resp: 819
Ion Ratio Lower Upper
200 100
173 38.1 25.0 37.6#
215 52.3 37.4 56.2









#25
Phenanthrene
Concen: 0.094 ng
RT: 17.127 min Scan# 1781
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:178 Resp: 5785
Ion Ratio Lower Upper
178 100
176 19.4 15.4 23.2
179 16.1 12.4 18.6



#26
Anthracene
Concen: 0.088 ng
RT: 17.226 min Scan# 1789
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57

Tgt Ion:178 Resp: 4546
Ion Ratio Lower Upper
178 100
176 18.6 14.8 22.2
179 15.8 12.2 18.4





#27

Fluoranthene-d10

Concen: 0.091 ng

RT: 19.126 min Scan# 2024

Delta R.T. -0.000 min

Lab File: BN030532.D

Acq: 29 Mar 2024 15:57

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:212 Resp: 5074

Ion Ratio Lower Upper

212 100

106 15.0 11.8 17.6

104 8.6 6.6 9.8



#28

Fluoranthene

Concen: 0.092 ng

RT: 19.154 min Scan# 2030

Delta R.T. -0.000 min

Lab File: BN030532.D

Acq: 29 Mar 2024 15:57




Tgt Ion:202 Resp: 6728


Ion Ratio Lower Upper

202 100

101 11.8 9.4 14.0

203 17.4 13.7 20.5









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.285 min Scan# 2460
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:240 Resp: 18481
Ion Ratio Lower Upper
240 100
120 13.0 10.6 15.8
236 27.8 22.1 33.1



#30
Pyrene
Concen: 0.097 ng
RT: 19.521 min Scan# 2109
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57

Tgt Ion:202 Resp: 7008
Ion Ratio Lower Upper
202 100
200 20.7 16.6 24.8
203 18.5 14.2 21.4









#31
 Terphenyl-d14
 Concen: 0.102 ng
 RT: 19.734 min Scan# 2154
 Delta R.T. 0.004 min
 Lab File: BN030532.D
 Acq: 29 Mar 2024 15:57




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:244 Resp: 4336
 Ion Ratio Lower Upper
 244 100
 212 9.4 6.9 10.3
 122 15.0 12.7 19.1



#32
 Benzo(a)anthracene
 Concen: 0.096 ng
 RT: 21.267 min Scan# 2458
 Delta R.T. -0.000 min
 Lab File: BN030532.D
 Acq: 29 Mar 2024 15:57

Tgt Ion:228 Resp: 6296
 Ion Ratio Lower Upper
 228 100
 226 28.4 21.8 32.8
 229 20.3 15.4 23.2









#33
Chrysene
Concen: 0.104 ng
RT: 21.321 min Scan# 2464
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:228 Resp: 7284
Ion Ratio Lower Upper
228 100
226 30.5 24.3 36.5
229 20.0 15.8 23.6



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.093 ng
RT: 21.213 min Scan# 2452
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57

Tgt Ion:149 Resp: 2628
Ion Ratio Lower Upper
149 100
167 27.3 21.4 32.0
279 3.3 2.3 3.5









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.527 min Scan# 2969
 Delta R.T. -0.000 min
 Lab File: BN030532.D
 Acq: 29 Mar 2024 15:57




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:	264	Resp:	18721
Ion Ratio	Lower	Upper	
264	100		
260	25.4	20.6	30.8
265	53.5	41.8	62.8



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.096 ng
 RT: 25.796 min Scan# 3745
 Delta R.T. -0.003 min
 Lab File: BN030532.D
 Acq: 29 Mar 2024 15:57


Tgt Ion:	276	Resp:	7967
Ion Ratio	Lower	Upper	
276	100		
138	28.9	23.0	34.4
277	25.1	20.1	30.1







#37
Benzo(b)fluoranthene
Concen: 0.094 ng
RT: 22.852 min Scan# 21
Delta R.T. -0.003 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1







Tgt Ion:252 Resp: 7168
Ion Ratio Lower Upper
252 100
253 30.4 19.2 28.8#
125 25.1 11.7 17.5#



#38
Benzo(k)fluoranthene
Concen: 0.094 ng
RT: 22.898 min Scan# 2754
Delta R.T. -0.000 min
Lab File: BN030532.D
Acq: 29 Mar 2024 15:57

Tgt Ion:252 Resp: 7159
Ion Ratio Lower Upper
252 100
253 29.8 18.9 28.3#
125 25.2 11.5 17.3#





#39

Benzo(a)pyrene

Concen: 0.095 ng

RT: 23.428 min Scan# 2935

Delta R.T. -0.003 min

Lab File: BN030532.D

Acq: 29 Mar 2024 15:57

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:252 Resp: 5770

Ion Ratio Lower Upper

252 100

253 33.8 20.0 30.0#

125 32.9 14.3 21.5#



#40

Dibenzo(a,h)anthracene

Concen: 0.094 ng

RT: 25.816 min Scan# 3752

Delta R.T. -0.000 min

Lab File: BN030532.D

Acq: 29 Mar 2024 15:57




Tgt Ion:278 Resp: 6209


Ion Ratio Lower Upper

278 100

139 23.7 16.2 24.2

279 32.7 20.6 30.8#





#41
 Benzo(g,h,i)perylene
 Concen: 0.098 ng
 RT: 26.489 min Scan# 3982
 Delta R.T. 0.003 min
 Lab File: BN030532.D
 Acq: 29 Mar 2024 15:57

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Tgt Ion:	276	Resp:	6920
Ion	Ratio	Lower	Upper
276	100		
277	27.7	19.8	29.6
138	29.8	20.6	31.0

