

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033121\
 Data File : BN014156.D
 Acq On : 01 Apr 2021 00:52
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Apr 01 01:25:42 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 31 12:15:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.692	152	7438	0.40	ng	0.00
7) Naphthalene-d8	10.467	136	27756	0.40	ng	0.00
13) Acenaphthene-d10	14.321	164	15738	0.40	ng	0.00
19) Phenanthrene-d10	17.067	188	32826	0.40	ng	0.00
29) Chrysene-d12	21.250	240	28693	0.40	ng	# 0.00
36) Perylene-d12	23.465	264	25640	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.284	112	7724	0.40	ng	0.00
5) Phenol-d6	6.855	99	9763	0.40	ng	0.00
8) Nitrobenzene-d5	8.832	82	6967	0.37	ng	0.00
11) 2-Methylnaphthalene-d10	12.057	152	17767	0.42	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	1870	0.37	ng	0.01
15) 2-Fluorobiphenyl	12.938	172	23325	0.39	ng	0.00
27) Fluoranthene-d10	19.099	212	35995	0.42	ng	0.00
31) Terphenyl-d14	19.700	244	28609	0.43	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.231	88	3964	0.41	ng	99
3) n-Nitrosodimethylamine	3.561	42	2598	0.35	ng	91
6) bis(2-Chloroethyl)ether	7.120	93	8029	0.42	ng	99
9) Naphthalene	10.518	128	28323	0.41	ng	99
10) Hexachlorobutadiene	10.809	225	5183	0.41	ng	# 100
12) 2-Methylnaphthalene	12.129	142	19000	0.42	ng	99
16) Acenaphthylene	14.042	152	22246	0.37	ng	100
17) Acenaphthene	14.384	154	17357	0.40	ng	100
18) Fluorene	15.374	166	21837	0.41	ng	100
20) 4,6-Dinitro-2-methylph...	15.450	198	1368	0.33	ng	# 83
21) 4-Bromophenyl-phenylether	16.264	248	6819	0.40	ng	93
22) Hexachlorobenzene	16.374	284	7578	0.40	ng	99
23) Atrazine	16.532	200	4224	0.37	ng	99
24) Pentachlorophenol	16.715	266	1909	0.33	ng	98
25) Phenanthrene	17.104	178	36007	0.41	ng	100
26) Anthracene	17.189	178	29025	0.39	ng	100
28) Fluoranthene	19.129	202	38963	0.42	ng	100
30) Pyrene	19.491	202	39012	0.42	ng	100
32) Benzo(a)anthracene	21.229	228	32491	0.41	ng	98
33) Chrysene	21.282	228	37232	0.42	ng	99
34) Bis(2-ethylhexyl)phtha...	21.165	149	10517	0.43	ng	100
35) Indeno(1,2,3-cd)pyrene	25.687	276	37884	0.42	ng	99
37) Benzo(b)fluoranthene	22.798	252	37890	0.42	ng	97
38) Benzo(k)fluoranthene	22.842	252	36915	0.42	ng	97
39) Benzo(a)pyrene	23.366	252	32796	0.42	ng	# 94
40) Dibenzo(a,h)anthracene	25.704	278	31792	0.40	ng	98
41) Benzo(g,h,i)perylene	26.368	276	33108	0.39	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033121\
Data File : BN014156.D
Acq On : 01 Apr 2021 00:52
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Quant Time: Apr 01 01:25:42 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 31 12:15:38 2021
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.692 min Scan# 610
 Delta R.T. 0.008 min
 Lab File: BN014156.D
 Acq: 01 Apr 2021 00:52


Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4





Tgt Ion:152 Resp: 7438
 Ion Ratio Lower Upper
 152 100
 150 153.0 121.2 181.8
 115 58.6 46.2 69.2




#2
 1,4-Dioxane
 Concen: 0.41 ng
 RT: 3.231 min Scan# 56
 Delta R.T. -0.008 min
 Lab File: BN014156.D
 Acq: 01 Apr 2021 00:52



Tgt Ion: 88 Resp: 3964
 Ion Ratio Lower Upper
 88 100
 58 79.6 63.1 94.7
 43 32.9 26.2 39.2









#3
n-Nitrosodimethylamine
Concen: 0.35 ng
RT: 3.561 min Scan# 97
Delta R.T. -0.008 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	42	Resp:	2598
Ion	Ratio	Lower	Upper
42	100		
74	108.2	94.6	141.8
44	4.8	4.2	6.2



#4
2-Fluorophenol
Concen: 0.40 ng
RT: 5.284 min Scan# 311
Delta R.T. -0.000 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52

Tgt Ion:	112	Resp:	7724
Ion	Ratio	Lower	Upper
112	100		
64	60.6	48.3	72.5
63	31.7	25.4	38.0









#5
Phenol-d6
Concen: 0.40 ng
RT: 6.855 min Scan# 506
Delta R.T. -0.000 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 9763
Ion Ratio Lower Upper
99 100
42 22.7 18.2 27.4
71 33.7 26.7 40.1



#6
bis(2-Chloroethyl)ether
Concen: 0.42 ng
RT: 7.120 min Scan# 539
Delta R.T. -0.000 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52

Tgt Ion: 93 Resp: 8029
Ion Ratio Lower Upper
93 100
63 82.3 65.1 97.7
95 33.3 26.6 39.8









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.467 min Scan# 857
Delta R.T. -0.000 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	27756
Ion Ratio	Lower	Upper	
136	100		
137	11.0	9.1	13.7
54	5.9	4.3	6.5
68	4.7	3.4	5.2



#8
Nitrobenzene-d5
Concen: 0.37 ng
RT: 8.832 min Scan# 728
Delta R.T. -0.000 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52

Tgt Ion:	82	Resp:	6967
Ion Ratio	Lower	Upper	
82	100		
128	39.1	33.5	50.3
54	52.6	40.3	60.5









#9
Naphthalene
Concen: 0.41 ng
RT: 10.518 min Scan# 861
Delta R.T. 0.012 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:128 Resp: 28323
Ion Ratio Lower Upper
128 100
129 11.2 9.2 13.8
127 12.8 10.8 16.2



#10
Hexachlorobutadiene
Concen: 0.41 ng
RT: 10.809 min Scan# 884
Delta R.T. 0.012 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52

Tgt Ion:225 Resp: 5183
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 51.6 77.4









#11
2-Methylnaphthalene-d10
Concen: 0.42 ng
RT: 12.057 min Scan# 1063
Delta R.T. 0.004 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 17767
Ion Ratio Lower Upper
152 100
151 20.0 16.1 24.1



#12
2-Methylnaphthalene
Concen: 0.42 ng
RT: 12.129 min Scan# 1079
Delta R.T. -0.000 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52

Tgt Ion:142 Resp: 19000
Ion Ratio Lower Upper
142 100
141 89.6 71.1 106.7
115 36.1 29.0 43.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.321 min Scan# 1307
Delta R.T. -0.000 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 15738
Ion Ratio Lower Upper
164 100
162 99.6 81.4 122.2
160 44.6 37.5 56.3



#14
2,4,6-Tribromophenol
Concen: 0.37 ng
RT: 15.818 min Scan# 1425
Delta R.T. 0.013 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52

Tgt Ion:330 Resp: 1870
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 37.8 31.2 46.8









#15
2-Fluorobiphenyl
Concen: 0.39 ng
RT: 12.938 min Scan# 1198
Delta R.T. -0.000 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172	Resp:	23325
Ion Ratio	Lower	Upper
172	100	
171	34.7	26.9 40.3
170	23.5	17.8 26.8



#16
Acenaphthylene
Concen: 0.37 ng
RT: 14.042 min Scan# 1285
Delta R.T. 0.013 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52

Tgt Ion:152	Resp:	22246
Ion Ratio	Lower	Upper
152	100	
151	19.4	15.4 23.2
153	13.1	10.3 15.5









#17
Acenaphthene
Concen: 0.40 ng
RT: 14.384 min Scan# 1312
Delta R.T. -0.000 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 17357
Ion Ratio Lower Upper
154 100
153 111.3 89.0 133.4
152 54.4 43.1 64.7



#18
Fluorene
Concen: 0.41 ng
RT: 15.374 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52

Tgt Ion:166 Resp: 21837
Ion Ratio Lower Upper
166 100
165 97.3 77.9 116.9
167 13.6 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.067 min Scan# 1527
Delta R.T. 0.000 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 32826
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.5 5.2 7.8



#20
4,6-Dinitro-2-methylphenol
Concen: 0.33 ng
RT: 15.450 min Scan# 1396
Delta R.T. -0.000 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52

Tgt Ion:198 Resp: 1368
Ion Ratio Lower Upper
198 100
51 45.5 47.8 71.6#
105 35.1 35.4 53.2#









#21
4-Bromophenyl-phenylether
Concen: 0.40 ng
RT: 16.264 min Scan# 1461
Delta R.T. 0.000 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 6819
Ion Ratio Lower Upper
248 100
250 100.3 77.0 115.6
141 40.9 39.7 59.5



#22
Hexachlorobenzene
Concen: 0.40 ng
RT: 16.374 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52

Tgt Ion:284 Resp: 7578
Ion Ratio Lower Upper
284 100
142 35.8 29.2 43.8
249 30.8 24.7 37.1









#23
Atrazine
Concen: 0.37 ng
RT: 16.532 min Scan# 1483
Delta R.T. 0.000 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:200 Resp: 4224
Ion Ratio Lower Upper
200 100
173 24.8 20.9 31.3
215 49.0 39.5 59.3



#24
Pentachlorophenol
Concen: 0.33 ng
RT: 16.715 min Scan# 1498
Delta R.T. 0.000 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52

Tgt Ion:266 Resp: 1909
Ion Ratio Lower Upper
266 100
264 62.4 50.2 75.2
268 63.4 52.5 78.7









#25
Phenanthrene
Concen: 0.41 ng
RT: 17.104 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 36007
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 15.3 12.4 18.6



#26
Anthracene
Concen: 0.39 ng
RT: 17.189 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52

Tgt Ion:178 Resp: 29025
Ion Ratio Lower Upper
178 100
176 18.4 14.7 22.1
179 15.4 12.3 18.5









#27
Fluoranthene-d10
Concen: 0.42 ng
RT: 19.099 min Scan# 1823
Delta R.T. 0.003 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 35995
Ion Ratio Lower Upper
212 100
106 12.6 10.1 15.1
104 7.1 5.5 8.3



#28
Fluoranthene
Concen: 0.42 ng
RT: 19.129 min Scan# 1829
Delta R.T. 0.003 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52

Tgt Ion:202 Resp: 38963
Ion Ratio Lower Upper
202 100
101 10.9 8.8 13.2
203 17.3 13.8 20.6









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.250 min Scan# 2123
Delta R.T. 0.003 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 28693
Ion Ratio Lower Upper
240 100
120 7.2 8.4 12.6#
236 25.6 21.0 31.6



#30
Pyrene
Concen: 0.42 ng
RT: 19.491 min Scan# 1902
Delta R.T. 0.003 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52


Tgt Ion:202 Resp: 39012
Ion Ratio Lower Upper
202 100
200 20.2 16.2 24.4
203 17.4 13.9 20.9








#31
Terphenyl-d14
Concen: 0.43 ng
RT: 19.700 min Scan# 1944
Delta R.T. 0.003 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:244 Resp: 28609
Ion Ratio Lower Upper
244 100
212 7.5 6.0 9.0
122 10.5 8.6 13.0



#32
Benzo(a)anthracene
Concen: 0.41 ng
RT: 21.229 min Scan# 2121
Delta R.T. -0.008 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52

Tgt Ion:228 Resp: 32491
Ion Ratio Lower Upper
228 100
226 27.2 20.6 31.0
229 19.4 16.2 24.2





#33

Chrysene

Concen: 0.42 ng

RT: 21.282 min Scan# 2126

Delta R.T. -0.008 min

Lab File: BN014156.D

Acq: 01 Apr 2021 00:52

Tgt Ion:228 Resp: 37232

Ion Ratio Lower Upper

228 100

226 29.6 24.4 36.6





229 19.7 15.5 23.3

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.43 ng

RT: 21.165 min Scan# 2115

Delta R.T. -0.008 min

Lab File: BN014156.D

Acq: 01 Apr 2021 00:52




Tgt Ion:149 Resp: 10517


Ion Ratio Lower Upper

149 100

167 26.8 21.4 32.2

279 4.0 3.5 5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.42 ng
RT: 25.687 min Scan# 3190
Delta R.T. -0.001 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:276 Resp: 37884
Ion Ratio Lower Upper
276 100
138 22.9 19.4 29.2
277 24.8 19.8 29.6



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.465 min Scan# 2541
Delta R.T. -0.001 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52

Tgt Ion:264 Resp: 25640
Ion Ratio Lower Upper
264 100
260 24.3 19.8 29.8
265 31.7 30.7 46.1









#37
Benzo(b)fluoranthene
Concen: 0.42 ng
RT: 22.798 min Scan# 2346
Delta R.T. -0.004 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 37890
Ion Ratio Lower Upper
252 100
253 24.7 20.6 30.8
125 14.0 9.4 14.2



#38
Benzo(k)fluoranthene
Concen: 0.42 ng
RT: 22.842 min Scan# 2359
Delta R.T. -0.001 min
Lab File: BN014156.D
Acq: 01 Apr 2021 00:52

Tgt Ion:252 Resp: 36915
Ion Ratio Lower Upper
252 100
253 24.5 20.6 30.8
125 10.5 9.5 14.3





#39

Benzo(a)pyrene

Concen: 0.42 ng

RT: 23.366 min Scan# 2512

Delta R.T. -0.001 min

Lab File: BN014156.D

Acq: 01 Apr 2021 00:52





Tgt Ion:252	Resp:	32796
Ion Ratio	Lower	Upper
252	100	
253	26.5	21.9 32.9
125	19.6	10.8 16.2#

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4



#40

Dibenzo(a,h)anthracene

Concen: 0.40 ng




RT: 25.704 min Scan# 3195


Delta R.T. 0.003 min

Lab File: BN014156.D

Acq: 01 Apr 2021 00:52

Tgt Ion:278	Resp:	31792
Ion Ratio	Lower	Upper
278	100	
139	16.2	13.4 20.0
279	25.6	21.3 31.9





#41
 Benzo(g,h,i)perylene
 Concen: 0.39 ng
 RT: 26.368 min Scan# 3389
 Delta R.T. -0.001 min
 Lab File: BN014156.D
 Acq: 01 Apr 2021 00:52

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion:	276	Resp:	33108
Ion Ratio	Lower	Upper	
276	100		
277	24.2	20.1	30.1
138	20.7	17.0	25.4

