




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033121\
Data File : BN014138.D
Acq On : 31 Mar 2021 11:02
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 31 11:48:08 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 26 11:50:50 2021
Response via : Initial Calibration



TIC: BN014138.D\data.ms

(77) Benzidine




19.316min (+ 0.018) 0.00 ng

response 2779815

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.45
183.00	13.20	11.61
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033121\
Data File : BN014138.D
Acq On : 31 Mar 2021 11:02
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 31 11:48:08 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 26 11:50:50 2021
Response via : Initial Calibration



TIC: BN014138.D\data.ms

(70) Pentachlorophenol (C)

16.716min (+ 0.012) 22567.84 ng

response 494659

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	61.52
264.00	61.60	57.63
0.00	0.00	0.00